

NEIGRIHMS

North Eastern Indira Gandhi Regional Institute of Health and Medical Sciences, Shillong
(An Autonomous Institute, Ministry of Health and Family Welfare, Government of India)
Director's Block, GPO Post Bag No.92, Mawdiangdiang, Shillong 793 018, Meghalaya

Store & Procurement:
Tele Fax: (0364) 2538032
F. NEIGR/S&P/04/2014

Email: storeneigrihms@gmail.com
Website: neigrihms.nic.in
Dated: 09/03/2015

To,
Director,
M/s S.M. Scientific Instruments Pvt. Ltd.,
'Jindal House' A -24, Anand Vihar, Delhi -1100921
Tel: 011 -22146522 /22166039; Dated: 011 -42427738
E-mail: info@smscientific.com

Sub: Matter relating to online open tender for processing of Cold Storage Chamber for keeping dead bodies, Anthropometric Set and Hot Plate for department of Forensic Medicine, vide Tender Enquiry No: NEIGR/S&P/OT/E -16A/2014-15; Dated: 26/08/2014.

Ref: Letter no: SMS/NEIGR/SP/OT-16A/2014/-15TEN-1603/RK; dated: 28/02/2015

Sir,

1. I am to refer to your communication pertaining to the opening of online price bid of the Equipments /Instruments for the department of Forensic Medicine.
2. The communication with regard to price bid opening under (eprocure.nic.in) eprocurement mode has been perused and as per the existing BOQ format of eprocurement platform which was designed by National Informatics Center (NIC), one BOQ (Price bid format) can be loaded in the CPPP Portal and that prices of stores which may be technically non -responsive are also seen. It may be noted that the technical evaluation report uploaded prior to opening of price bid indicates your offer in respect of Cold Storage 2 - Body Chamber for keeping dead bodies to be TECHNICALLY NON -RESPONSIVE.
3. As per the technical evaluation report, your offer for the Cold Storage 2 -body Chamber for keeping dead bodies has been found technically non -responsive by the technical committee based on the following points:-
 - Point no 17: (Special Loading Trolley) - Specialized loading trolley not offered as per operational requirement and not visible /indicated in the brochure submitted.
 - Point no 18: (CFC Free Compressor) - Model does not indicate twin compressor as per international Standard (one as standby) and also in the referral purchase order enclosed.
 - The brochure /compliance statement uploaded is perceived to be 'cut and paste' of the tendered specification whereas the same model supplied elsewhere, as per bidding document, does not offer twin compressor as per international Standard (one as standby).
4. In respect of the qualification criteria, as per Section - IX (Qualification Criteria) of the bidding document, the details of which reads as follows:-
 - The Tenderer must be a Manufacturer or its authorized Agent.
 - The Manufacturer should have supplied and installed in last three years from the date of Tender Opening, at least 100% of the quoted quantity of the similar



equipment meeting major specification parameters which is functioning satisfactorily. The foreign Manufacturer satisfying the above criteria should also have supplied and installed in last three years from the date of Tender Opening, at least 100% of quoted quantity of similar model which is functioning satisfactorily any where outside the country of manufacture.

- The Tenderers quoting as authorized representative of the manufacturer shall have three years of experience in the related field and should obtain documents from principals/manufacturer fulfilling the requirements above, taking full responsibility of technical support, service and organizational support.
5. It may also be noted that the Institute is in the process of awarding of contract for the Digital Hot Plate in which M/s S.M. Scientific Instruments Pvt. Ltd., New Delhi is the only techno-commercially compliant bidder, since this store has been tendered for more than two occasions. However it is clarified that since the BOQ (price bid) format is on a single sheet, therefore the software of the Central Public Procurement Portal reflects the cost/price of Non-Responsive items/stores also, which is beyond the control of the Institute.
 6. The Cold Storage Chambers were also procured by AIIMS, Jodhpur and PGIMER, Chandigarh vide Order No: Admin/General/263/2013-AIIMS.JDH/2688; dated: 16/05/2014 and No.3PG/OG2/2013/Forensic Science -689/260; dated: 02/06/2013 respectively, on similar cost and terms & conditions, so there is no instance of financial loss to the Institute.
 7. The opening of price bid has been done, as per existing norms, after uploading of technical evaluation report, the report /communication has been sent to your registered email, auto-generated from the Central Public Procurement Portal.
 8. Opening of e-tenders has been processed as per existing government procedures and that such accusation of being unprofessional, unjustified is hereby unacceptable and uncalled for. All details of the technical evaluation report are available on the portal and processing shall be done as per norms.
 9. This communication shall be uploaded on the respective portal, along with supporting documents, for perusal and transparency.

Yours faithfully,


Stores & Procurement Officer,
NEIGRIHMS, Shillong

Copy to;

1. Deputy Director (Admn.), NEIGRIHMS, Shillong
2. Medical Superintendent, NEIGRIHMS, Shillong
3. Incharge, department of Forensic Medicine, NEIGRIHMS, Shillong
4. M/s Areng Medical Supplier, Meghalaya

The technical committee deliberated technical aspects and specification of the equipment by the following company.

Tender no NEIGR/SSP/OT-16A/14-15 for Dept of Forensic Medicine.

BIDDERS

SL.NO.	ITEM & SPECIFICATION	M/s Aaregh Medical Suppliers	M/s S.M.Scientific Instruments	M/s Surgicchem	M/s Rim Pharmaceuticals & Infrastructure	M/s Asomi Enterprise
1	Cold Storage 2-body Chamber for Keeping Dead Bodies 1. Corrosion free interior and exterior. 2. Audio visual alarm for high and low temperature. 3. Designed for long storage of cadaverous. 4. PUF insulation on all sides. 5. Special design ensuring best hygiene with washing & draining facility. 6. Reliable 7. Special loading trolley. 8. Energy efficient and sturdy construction. 9. Light weight. 10. Digital temperature indication. 11. Low maintenance. 12. Microprocessor based / PLC temperature control. 13. Double walled cooling units. 14. Outer body of the mortuary chamber is constructed out of thick SS sheets. The inner chamber made of heavy gauge stainless steel sheet of SS-304 grade. The 100mm gap between the walls filled high grade poly urethane insulation, which ensures maximum thermal efficiency. 15. The doors connected by very sturdy chrome plate hinges and fitted with hard chrome plated lubricated latches for opening of the door. 16. The doors made of galvanized steel sheets, lined with stainless steel for extra protection and long life.	Make: Motech Manufacturing INC USA Model:1096-R100	Make: s S.M.Scientific Instruments, India Model: SMI 157	Make: Evermed Italy Model:MMC 2.2	Make: P.L.Tandon & Co. Model: Not indicated	Make: Funeraria GmbH, Germany Model: No. 80947/44
		Completed	Completed	Completed	Completed	Completed
		Specialize loading trolley indicated in the offer.	Specialize loading trolley not offered as per operational requirement and not visible/indicated in the brochure submitted.	Specialize loading trolley not offered as per operational requirement and not visible/indicated in the brochure submitted.	Specialize loading trolley not offered as per operational requirement and not visible/indicated in the brochure submitted.	Specialize loading trolley indicated in the offer.
		Completed	Completed	Completed	Completed	Completed
		Completed	Completed	Completed	Completed	Completed
		Completed	Completed	Completed	Completed	Completed
		Completed	Completed	Completed	Completed	Completed
		Completed	Completed	Completed	Completed	Completed
		Completed	Completed	Completed	Completed	Completed
		Completed	Completed	Completed	Completed	Completed
		Completed	Completed	Completed	Completed	Completed
		Completed	Completed	Completed	Completed	Completed
		Completed	Completed	Completed	Completed	Completed
		Completed	Completed	Completed	Completed	Completed
		Completed	Completed	Completed	Completed	Completed

A. Shastri
A. Shastri
Member

R. K. Roy
Dr. A. Danda Roy
Member

R. K. Roy
Dr. R. Halang
Member

S. H. Dey
Sri Harendra Dey
Member Secretary

17. All the doors fitted with high quality neoprene rubber gaskets for airtight fittings with very sturdy casters.	Complied	Complied	Complied	Complied	Complied	Complied
18. CFC free compressors, conforming to latest international standards and guidelines. Twin compressors of which one is standby.	Yes of International Standard	Model does not indicate twin compressor as per International standard, (one as standby) and also in Referral Purchase Order enclosed.	Complied	Complied	Model does not indicate twin compressor as per International standard, (one as standby)	Yes of International Standard
19. Vapor proof lamp inside.	Complied	Complied	Complied	Complied	Complied	Complied
20. Temperature range -2 to 4 deg C with temp failure alarm.	Complied	Complied	Complied	Complied	Complied	Complied
21. Suitable Voltage automatic stabilizer O/P	Complied	Complied	Complied	Complied	Complied	Complied
22. +/-10% I/P 150 - 280V/0.5s	Complied	Complied	Complied	Complied	Complied	Complied
23. To be installed at each site as per the site conditions.	Complied	Complied	Complied	Complied	Complied	Complied
24. The unit should be 2 ther mortuary cold room for 2 corpse	Complied	Complied	Complied	Complied	Complied	Complied
25. CE /FDA Certified	CE Certified	CE & GMP Certified	Complied	Complied	Complied	CE Certified
FINAL REMARKS	Acceptable, system offered from reputed manufacturer	Not acceptable. Specialize loading trolley not offered as per operational requirement and not visible/indicated in the brochure submitted. Model does not indicate twin compressor as per International standard, (one as standby) and also in Referral Purchase Order enclosed.	Complied	Not acceptable. Specialize loading trolley not offered as per operational requirement and not visible/indicated in the brochure submitted.	Not acceptable. Specialize loading trolley not offered as per operational requirement and not visible/indicated in the brochure submitted. Model does not indicate twin compressor as per International standard, (one as standby)	Acceptable, system offered from reputed manufacturer

A. Bhattacharya
 Dr. A. Bhattacharya
 Chairman

Dr. A. D. Karmay
 Dr. A. D. Karmay
 Member

Dr. R. H. Singh
 Dr. R. H. Singh
 Member

Shri Harindra Dey
 Shri Harindra Dey
 Member Secretary