

North Eastern Indira Gandhi Regional Institute of Health and Medical Sciences
(An Autonomous Institute, Ministry of Health and Family Welfare, Government of India)
Director's Block, Mawdiangdiang, Shillong 793 018 (Meghalaya)

Tender Corrigendum No: NEIGR/S&P/01/08/2017 -2018

Dated: 08/08/2017

E -TENDER CORRIGENDUM /ADDENDUM/EXTENSION

- A. Reference Enquiry No.: -NEIGR/S&P/OT/E -14/2017 -18, dated: 10/07/2017 in respect of Open e-Tender for processing Multipara monitors-10nos on buyback basis with five years warranty and five years CMC, for Institute.

The following amendment /addendum are considered in the technical specification of Multipara monitors:-

Heading: technical Specification:-

- Point No: 1 -may be read as "**Minimum 19 inches or more multi colored TFT-LCD touch screen Display**".
- Point No: 3 -may be read as "**Facility to monitor and display- ECG, Respiration, NIBP,SPO2, temp, IBP- 2 nos, CO2 with capnography (Optional), Gastrictonometry (optional). Cardiac output (optional), NMT (optional), BIS/Entropy (optional), EEG (optional)**".
- Point No: 19.b -may be read as "**To provide suitable facility for sending and receiving Dicom Compatible or equivalent radiological images like ultrasound, X-ray etc to and from the HIS/RIS through central monitoring station**".

Heading: System configuration accessories, spares and consumables:-

- Point No: 4 -may be read as "**IBP:-Includes 4 nos. Per monitor of reusable pressure transducer along with 2 nos per monitor IBP cables with bracket holders and 10 nos disposable domes per monitor**".
- Point no: 7-may be read as "**Cardiac output (Optional): should be by thermo dilution method with accessories**".

Heading: Power Supply:-

- Point no: 7-may be read as "**Deleted**".

In view of minor mismatch/error in the CMC column of BOQ/Price bid; bidders are conveyed to offer CMC cost from the 6th to 10th year exclusive of GST, as percentage of FOB cost of the first listed store (Section -XI C) , in the techno commercial bids, for further evaluation.

The last date and time of submission of e -bids has been extended upto **14:00 hrs of 24/08/2017**

All other terms and conditions remains the same.

- B. Reference Tender Enquiry No: NEIGR/S&P/OT/E -18/2017-18 for processing of PCA Pump -10 nos, PRP Centrifuge -01 no, Chest Compression Device and Difficulty Airway Management Simulator -01 no, with 5 years warranty and thereafter 5 years CMC, for department of Anaesthesia.

The following amendment /addendum are considered against the technical specification of **PCA Pump**:-

Heading: Specification:-

Point No: 1 -may be read as "**Should have built in safety features, like different lock levels (atleast 3), upstream, downstream occlusion sensor, and cassette /Reservoir detection**".

Point No: 3 -may be read as "**Should have the flexibility to power the pump up by rechargeable batteries with minimum battery backup of 6-7hours with provision to display residual battery life in hours and minutes and AC adaptor**".

Point No: 4 -may be read as "**Should have IPX rating (Moisture protection)>= 3**".

Heading: Administration Set:-

Point No: 1 -may be read as **"Used with reservoir /cassette"**.

Point No: 3 -may be read as **"Integrated 0.2-0.25 micron filter"**.

Point No: 4 -may be read as **"Fluid path materials made of non- PVC material"**.

In view of minor mismatch/error in the CMC column of BOQ/Price bid; bidders are conveyed to offer CMC cost from the 6th to 10th year exclusive of GST, as percentage of FOB cost of the first listed store (Section -XI C) , in the techno commercial bids, for further evaluation.

The last date and time of submission of e -bids has been extended upto **14:00 hrs of 24/08/2017**

All other terms and conditions remains the same.

- C. Reference Tender Notice No: NEIGR/S&P/02/07/2017-2018; dated: 12/07/2017, vide Tender Enquiry No: NEIGR/S&P/OT/E -17/2017-18 for processing of ICU Beds -08 nos with 5 years warranty and thereafter 5 years CMC, for the ICU's of the Institute

The following amendment /addendum are hereby being considered against the technical specification of **ICU Beds:-**

Heading: Specification:-

Point No: 4 -may be read as **"Should have embedded type control having backlighting facility for patient and care giver, Caregiver or nursing control should be lockable as per choice"**.

Point No: 5 -may be read as **"Bed should have two piece split safety side railing with embedded type controls"**.

Point No: 10 -may be read as **"It should have auto CPR on the embedded control for emergency"**.

Point No: 27 -may be read as **"The model which has to quote should have FDA approval. The bidder has to submit FDA approval Certificate copy i.e. 510 (k) approval copies in the technical bid. "**

In view of minor error in the CMC column of BOQ/Price bid; bidders are conveyed to offer CMC cost from the 6th to 10th year exclusive of GST, as percentage of FOB cost of the first listed store (Section -XI C) , in the techno commercial bids, for further evaluation.

The last date and time of submission of e-bids has been extended upto **14:00 hrs of 24/08/2017**

All other terms and conditions remains the same.

- D Reference Tender Enquiry No: NEIGR/S&P/OT/E -30 V,W.X.Y & Z/2017-18 for processing Chemicals, Reagents, Glassware, Instruments, Surgicals, Contrast, etc for the Institute, for a period of one year, extendable upto 6 months, or till the finalization of the next tender, whichever is later.

The last date and time of submission of e-bids has been extended upto **14:00 hrs of 24/08/2017**

All other terms and conditions remains the same.

- E. Reference Tender Enquiry No.:- NEIGR/S&P/OT/E -19/2017 -18 ,dated: 10/07/2017 for Open e-Tender for processing of Cavitation Ultrasonic surgical aspirator(CUSA) with five year Warranty & Five year CMC, for the Institute.

The following amendment /addendum are considered against the technical specification of CUSA:-

Heading: Specification:-

Point No: 12 -may be read as **"The system should have two working frequency hand pieces/probes i.e. 22-25 KHz & 33-36 KHz for different type of tissues."**

Point No: 13 –may be read as “Minimum Consumables for 100 cases (considering 80 open procedures & 20 lap procedures) including tubing set, Micro tips Tubing Set, Micro tips, Macro tips, laparoscopic tips and any system specific consumables as required for the system should be provided along with the system and prices of all the consumables has to be quoted as percentage at section XI (B) price schedule of the technical bid for processing of future requirements.

NOTE:- The manufacturer must submit the declaration mentioning that the Tips & Tubings are autoclavable which can be used for 100 procedures in the technical bid or else bidder has to provide 100 tips & 100 nos Tubing set & system specific consumables if any for 100 cases .”

Point No: 16 a - may be read as “The open headpieces of each working frequency supplied along with the system should support an integrated cautery device operation for simultaneous coagulation /through the same hand piece. This should be made possible through our existing cautery system or if not possible, may provide a compatible cautery system for this function.

Point No: 16 b - may be read as “At least two autoclavable containers from OEM should be provided along with the system for safe storage & sterilization of hand pieces”.

In view of minor error in the CMC column of BOQ/Price bid; bidders are conveyed to offer CMC cost from the 6th to 10th year exclusive of GST, as percentage of FOB cost of the first listed store (Section –XI C) , in the techno commercial bids, for further evaluation.

The last date and time of submission of e-bids has been extended upto **14:00 hrs of 24/08/2017**

All other terms and conditions remains the same.

- F Reference Tender Enquiry No.: -NEIGR/S&P/OT/E -25/2017 -18, dated: 10/07/2017 in respect of Open e-Tender for processing of automated blood gas analyzer, on cost per test basis, on rate contract for a period of two years, or till the finalization of the next tender, whichever later, for the institute.

The following amendment /addendum are considered in the technical specification of Blood Gas Analyzer:-

Heading: Specification:-

Point No:17 –may be read as “**Analyzer should have automatic onboard/internal QC for maintenance free operations and in view to obtain qualitative result .”**

In view of minor mismatch/error in the CMC column of BOQ/Price bid; bidders are conveyed to offer CMC cost from the 6th to 10th year exclusive of GST, as percentage of FOB cost of the first listed store (Section –XI C) , in the techno commercial bids, for further evaluation.

The last date and time of submission of e-bids for Tender Enquiry No.: -NEIGR/S&P/OT/E -25/2017 -18 **14:00 hrs of 07/09/2017**

All other terms and conditions remains the same.

For further details regarding amendment, addendum, extension and downloading of documents, please visit our website: website: www.eprocure.gov.in; Tenders can also be downloaded from the Central Public Procurement Portal website: www.eprocure.gov.in; Tel/Fax: 0364-2538032.

Stores & Procurement Officer,
For and on behalf of Director, NEIGRIHMS