

E -TENDER CORRIGENDUM

Reference Tender Notice No: NEIGR/S&P/02/07/2017-2018; dated: 12/07/2017, vide Tender Enquiry No: NEIGR/S&P/OT/E -18/2017-18 for processing of PCA Pump -10 nos, PRP Centrifuge -01 no, Chest Compression Device and Difficulty Airway Management Simulator -01 no, with 5 years warranty and thereafter 5 years CMC, for department of Anaesthesia.

The following amendment /addendum are hereby being considered against the technical specification of **PCA Pump**:-

Heading: Specification:-

Point No: 1 –may be read as “**Should have built in safety features, like different lock levels (atleast 3), upstream, downstream occlusion sensor, and cassette /Reservoir detection**”.

Point No: 3 –may be read as “**Should have the flexibility to power the pump up by rechargeable batteries with minimum battery backup of 6-7hours with provision to display residual battery life in hours and minutes and AC adaptor**”.

Point No: 3 –may be read as “**Should have IPX rating (Moisture protection)>= 3**”.

Heading: Administration Set:-

Point No: 1 –may be read as “**Used with reservoir /cassette**”.

Point No: 3 –may be read as “**Integrated 0.2-0.25 micron filter**”.

Point No: 4 –may be read as “**Fluid path materials made of non- PVC material**”.

All other terms and conditions remains the same.

For further details regarding amendment, addendum, extension and downloading of documents, please visit our website: www.neigrihms.nic.in; Tenders can also be downloaded from the Central Public Procurement Portal website: www.eprocure.gov.in; Tel/Fax: 0364-2538032.

Sd/
Stores & Procurement Officer,
For and on behalf of Director, NEIGRIHMS