

North Eastern Indira Gandhi Regional Institute of Health and Medical Sciences
(An Autonomous Institute, Ministry of Health and Family Welfare, Government of India)
Director's Block, Mawdiangdiang, Shillong 793 018 (Meghalaya)
Tender Corrigendum No: NEIGR/S&P/07/09/2017-18 Dated: 13/09/2017

E -TENDER CORRIGENDUM

Reference e-Tender Enquiry No.:- NEIGR/S&P/OT/E -38/2017 -18, dt.31/08/2017 for Open e-Tender for processing of Processing of Processing of OT Light, Skin Calipers and Height Measuring Machine for the Department of General Surgery with five years warranty and five years CMC for the Institute.

The technical specification Committee perused all details and indicated the following, in view of requirement of the Institute. In view of Institute requirement and wider participation, the following amendment /addendum are hereby being considered against the technical specification of Mobile OT light:-

Specification:-

- Point No:1 may be read as "To provide pure white cool light variable light intensity of 80,000 lux or more for minor OT's critical care and emergency areas."
- Point No: 3 may be read as "Extremely flat, compact dome with two handles."
- Point No: 5 may be read as "Light head made of durable polycarbonate and having one sterilizable handle."
- Point No:12 may be read as "Electrical commutators to be present to provide continuous rotation of light head and rotatable by 250-300 deg."
- Point No: 8 may be read as "Light pattern size to be fixed and should be min 170-270mm."
- Point No: 11 may be read as "Color temperature between 3800k-5500k."
- Point No: 12 may be read as "Color rendering index (CRI): 90-95 %"

All other terms and conditions remains the same.

For further details regarding amendment, addendum, extension and downloading of documents, please visit procurement portal: www.eprocure.gov.in Tel/Fax: 0364-2538032; E-mail: storeneigrihms@gmail.com.

Sd/-
Stores & Procurement Officer,
For and on behalf of Director, NEIGRIHMS