

North Eastern Indira Gandhi Regional Institute of Health and Medical Sciences
(An Autonomous Institute, Ministry of Health and Family Welfare, Government of India)
Director's Block, Mawdiangdiang, Shillong 793 018 (Meghalaya)

Tender Corrigendum No: NEIGR/S&P/06/09/2017-18

Dated: 13/09/2017

E -TENDER CORRIGENDUM

Reference e-Tender Enquiry No.:- NEIGR/S&P/OT/E -41/2017 -18, dt.21/08/2017 for Open e-Tender for processing of Processing of Steam Sterilizer -01 nos against buyback of old existing Rapid autoclave - 02nos (Make-Yorco,)for the Institute with five years warranty and five years CMC, for Institute.

Pre bid Conference was held on 8/9/2017, which was attended to by the prospective bidders. The users and technical specification Committee perused all details and indicated the following, in view of requirement of the Institute. In view of Institute requirement and wider participation, the following amendment /addendum are hereby being considered against the technical specification of Steam Sterilizer:-

Heading: b) **Construction:-**

Specification:-

- Point No:5 may be read as "Steam Generator: The sterilizer should have inbuilt steam generator of adequate capacity. Inbuilt steam generator, it should be mounted under the sterilizer chamber and should be made of 316 quality stainless steel. The steam generator should have chloride free mineral wool/mineral glasswool of thickness 25 mm to 50 mm insulation with SS 316. It should have a built in thermostat, pressure safety valve & water level glass gauge inspection device visible from service area. The heating element should be of sufficient capacity to make the sterilization process faster with maximum cycle time of 45-50 mins in pre vacuum. It should also have the automatic blow down valve & degassing system for feeding water to steam generator .Above 35 KW Built in Steam Generator."

Specification:-

- Point Noq:"In case of suppliers offering standalone steam generator they should provide alternatives for ensuring clean steam (as per International Standards).
 - i. With standalone generator
 - ii. For preheating the sterilizer with steam from a central boiler having adequate stand by supply
"may be read as" Deleted".

All other terms and conditions remains the same.

For further details regarding amendment, addendum, extension and downloading of documents, please visit e-procurement portal: www.eprocure.gov.in Tel/Fax: 0364-2538032; E-mail: storeneigrihms@gmail.com.

Sd/-

Stores & Procurement Officer,
For and on behalf of Director, NEIGRIHMS