

e -TENDER AMENDMENT

Reference Tender Notice No: NEIGR/S&P/03/05/2018-2019; Dated: 17.05.2018 pertaining to Tender Enquiry No: NEIGR/S&P/OT/E-02/2018 -19 for processing of HDR Brachytherapy Unit (C -60 Source) with complete turnkey works, with 5 years warranty and thereafter 5 years CMC, for department of Radiation Oncology.

1. **Point No. 1.** General Specifications: Clause No. 10: The sentence "UPS, Color CCTV and two-way patient communication system must be included with the offer" should be read as "**Online UPS, Colour CCTV and two-way patient communication system must be included with the offer**".
2. **Point no 8,** Treatment Planning System Under Sub Point A: **Hardware:** The Sentence "24" color TFT Flat screen" Should be Read as "**High Resolution, 24" colour TFT Flat screen**"
3. **Point no 8,** Treatment Planning System Under Sub Point C: Networking - **Connectivity:** The sentence "The TPS should be connected to the entire HDR Brachytherapy Unit, C-Arm & Ultra Sound Unit, CT-Simulator/CT-Scan and MRI machine in NEIGRIHMS" should be read as "**The TPS should be connected to the entire HDR Brachytherapy Unit, C-Arm & Ultra Sound Unit, CT-Simulator/CT-Scan and MRI machine and External Beam Radiotherapy TPS available in NEIGRIHMS and all necessary licenses should be provided and the responsibility lies with bidder.**
4. **Point no 4. Applicators - All the applicators for brachytherapy (Intracavitary, Interstitial etc.) should be reusable multiple times and the method of sterilisation for repeat use should be mentioned clearly for each.**

All other terms and conditions remains the same.

For further details regarding amendment, addendum, extension and downloading of documents, please visit website: www.eprocure.gov.in /www.neigrihms.gov.in; Tel/Fax: 0364-2538032; E-mail: storeneigrihms@gmail.com.

Sd/-
Stores & Procurement Officer,
For and on behalf of Director, NEIGRIHMS