

North Eastern Indira Gandhi Regional Institute of Health and Medical Sciences
(An Autonomous Institute, Ministry of Health and Family Welfare, Government of India)
Director's Block, Mawdiangdiang, Shillong 793 018 (Meghalaya)

Tender Corrigendum No: NEIGR/S&P/

Dated: 04/07/2018

E -TENDER CORRIGENDUM

Reference Tender Enquiry No.:- **NEIGR / S&P / OT / E – 05 / 2018 – 19 Dated 08.06.18.** ONLINE OPEN TENDER FOR PROCESSING OF HIGH END DIGITAL ANESTHESIA WORKSTATION (2 NOS) WITH 5 YEARS WARRANTY AND THEREAFTER 5 YEAR CMC AND Tender Enquiry No.:- **NEIGR / S&P / OT / E – 11 / 2018 – 19 Dated 08.06.2018** ONLINE OPEN TENDER FOR PROCESSING OF ANESTHESIA WORKSTATION (2 NOS) AGAINST BUY BACK OF 3 NOS OF OLD/OBSOLETE EXISTING & UNSERVICEABLE VENTILATORS ASSOCIATED WITH ANESTHESIA WORKSTATION (MAKE: PENLON; MODEL: AV 900) FOR DIFFERENT OT'S OF THE INSTITUTE

The following amendment /addendum are hereby being considered against the technical specification of For ANESTHESIA WORKSTATION:-

Technical Specifications:

1. Heading (2) Flow Management : Point No. 9. "Total Agent consumption in ml should be measured and displayed on the screen with a log of at least last 1 case."

May be read as

"Total Agent consumption in ml should be measured and displayed on the screen."

2. Heading – (3) Vaporizers : Point No. 6. "'Low agent level' Alarm for all Anesthetic agent".

May be read as

"Low agent level / Prismatic Level Indicator Alarm for all Anesthetic agent".

3. Heading – (6) Anesthesia Monitoring : Point No. 18. "Recorder for printing the up to 4 waveforms and alphanumeric data, and trends etc."

May be read as

"Thermal Recorder for printing the up to 3 waveforms and alphanumeric data, and trends etc."

4.Heading – (6) Anaesthesia Monitoring : Point No. 19. "Should be supplied with Main stream ETCO2 Module with sensor cable".

May be read as

"Should be supplied with ETCO2 Module with accessories".



"Should be supplied with ETCO2 Module with accessories".

5.Heading – (8) Scope of Supply : Point No. 17. "Anesthesia Monitor, 15–inch color touch–screen with 8 waveforms".

May be read as

"Anesthesia Monitor, 19–inch color touch–screen with minimum 10 waveforms".

6.Heading – (8) Scope of Supply : Point No. 30. "Isoflurane Electronic Vaporizer".

May be read as

"Isoflurane Vaporizer".

7. Heading – (8) Scope of Supply : Point No. 30. " Sevoflurane Electronic Vaporizer".

May be read as

"Sevoflurane Vaporizer".

8. Heading – (8) Scope of Supply : Point No. 30. " Desflurane Electronic Vaporizer".

May be read as

"Desflurane Vaporizer".

9.Heading – (8) Scope of Supply : Point No. 33. "Main Stream ETCO2 Module – 1 No. with Sensor Cable and Adopter – 2 Nos."

May be read as

"ETCO2 Module – 1 No. with accessories – 20 Nos".

All other terms and conditions remains the same.

For further details regarding amendment, addendum, extension and downloading of documents, please visit website: www.eprocure.gov.in /www.neigrihms.gov.in; Tel/Fax: 0364-2538032; E-mail: storeneigrihms@gmail.com.

Sd/-

Stores & Procurement Officer,
For and on behalf of Director, NEIGRIHMS

