

Tender Corrigendum No: NEIGR/S&P/05/2018 -19

Dated: 11.07.2018

E -TENDER CORRIGENDUM

Reference Tender Enquiry No.:- NEIGR/S&P/OT-E -04/2018 -19, Dt. 22/05/2018 Online Open Tender for processing of Vascular Set, Major Basic Set and Basic Neck Set with 5 years warranty, for department of Surgical Oncology.

The following amendment /addendum are hereby being considered against the technical specification of For **Vascular Set, Major Basic Set Basic Neck Dissection Set:-**

Technical Specifications:

1. Point No. 58 for Vascular Set & ADD the specification for Major Basic Set Basic Neck Dissection Set .May be read as

The container has to supply along with each instrument set must be of following specification:-

- a. The Instrument and Container should be of the same parent company.
- b. The sterilization container and accessories should be of the same parent company.
- c. The Sterilization containers should meet the international standards and approved for steam sterilization procedures to EN 285: 2008 and validated acc to ISO 17665 Part1:2006.
- d. The Sterilization Containers should offer superior filtration efficiency of 99.99997%.
- e. It should have a visual indicator for sterile and non sterile which can clearly visible from the outside labels/self destructive label indicator without opening the container if the container was opened once. The labels (if not reusable) have to provide free up cost.
- f. It should have reusable microbial barriers instead of disposable filters for reducing the recurring cost. The microbial barriers should be easy to remove and clean.
- g. It should have lateral flow ducts at the top for flow of air.
- h. The instruments should remain sterile in the container and the container should be capable of being brought into the Operation Room without any essential packaging.
- i. It should also consist of tray and silicon matt (for micro instruments).
- j. Container should opening mechanism in order to prevent accidental opening.
- k. Adequate size & no of containers has to be providing for each sets.

All other terms and conditions remains the same.

For further details regarding amendment, addendum, extension and downloading of documents, please visit website: www.eprocure.gov.in /www.neigrihms.gov.in; Tel/Fax: 0364-2538032; E-mail: storeneigrihms@gmail.com.

Sd/-
Stores & Procurement Officer,
For and on behalf of Director, NEIGRIHMS