

e-TENDER AMENDMENT

Reference Tender Notice No: NEIGR/S&P/01/03/2017-2018; Dated: 05.03.2018 pertaining to Tender Enquiry No: NEIGR/S&P/OT/E-75/2017 -18 for processing of Fully Automated Capillary Electrophoresis System, with 5 years warranty and thereafter 5 years CMC, for department of Biochemistry.

The following amendment /addendum are hereby being considered against the technical specification of Fully Automated Capillary Electrophoresis System:-

Point: 3:-May be read as "Diameter of capillary should be between 25-40µm to avoid any clog during sample run to increase efficiency of system and avoid loss of proteins."

Point: 7:-May be read as "System should have 2-6 buffer containers with each kits for ease of daily operations."

Point: 10:-May be read as "Deleted."

Point: 11:-May be read as "System should have temperature controlled reagent positions for on board assay flexibility for addition of antisera like IgD and IgE."

Point: 12:-May be read as "Deleted."

Additional Technical Specifications:

Monthly 600 samples (Serum protein-,urineprotein-,immunodisplacement (IgG,IgA,IgM,Kappa,Lambda) - Serum protein high resolution - ,CDT- ,Hemoglobin IEF -) will be taken into consideration for a period of 10 years for price evaluation for a period of 10 years, so the bidder has to mention the operational cost per test for each test mentioned in the technical specification as % of FOB price, in Schedule XI of the technical bid.

All other terms and conditions remains the same.

For further details regarding amendment, addendum, extension and downloading of documents, please visit our website: www.neigrhms.nic.in; Tenders can also be downloaded from the Central Public Procurement Portal website: www.eprocure.gov.in; Tel/Fax: 0364-2538032.

Sd/-
Stores & Procurement Officer,
For and on behalf of Director, NEIGRIHMS