

North Eastern Indira Gandhi Regional Institute of Health and Medical Sciences  
(An Autonomous Institute, Ministry of Health and Family Welfare, Government of India)

Director's Block, Mawdiangdiang, Shillong 793 018 (Meghalaya)

Tender Amendment No: NEIGR/S&P/01/10/2018 -2019

Dated: 10.09.2018

**e-TENDER AMENDMENT**

Reference Tender Notice No: NEIGR/S&P/03/05/2018-2019; Dated: 17.05.2018 pertaining to Tender Enquiry No: NEIGR/S&P/OT/E-40/2018 -19 for processing of 05 nos of High End Suction Machine, with 5 years warranty and thereafter 5 years CMC, for various Wards, OT's, ICU's of the Institute against buy back of old existing obsolete 25 nos of Suction Machine (Make: Simex, Germany; Model: DC 20SU/DC30SU; Year: 2009)

The following amendment /addendum are hereby being considered against the technical specification of the above system:-

1. Point No. 6:- " Should have Noise level: 48dB (A) @ 1 m (acc. to ISO 7779).".

*May be read as*

"Should have Noise level: less than 70dB (A) @ 1 m (acc. to ISO 7779)."

2. Heading 3 Technical Specifications: Point No. 3.1. "Should have Air flow rate of pump: 35 ± 3 l/min"

*May be read as*

"Should have Air flow rate of pump: 40 ± 5 l/min"

3. Point No. 11. "It should have two autoclavable PSU (Polysulfone) jars of 3 liters with a change switch from one jar to another."

*May be read as*

"It should have two autoclavable PSU (Polysulfone) jars of 2 liters or more with a change switch from one jar to another."

4. Point No. 16. "Should be a complete non corrosive body and mounted on castors for easy mobility. System should be easy mountable and dismountable from the trolley. (The suction system and trolley should be from ONE manufacturer)."

*May be read as*

"Should be a complete non corrosive body and mounted on castors for easy mobility. (The suction system and trolley should be from ONE manufacturer)."

All other terms and conditions remains the same.

For further details regarding amendment, addendum, extension and downloading of documents, please visit website: [www.eprocure.gov.in](http://www.eprocure.gov.in) /[www.neigrihms.gov.in](http://www.neigrihms.gov.in); Tel/Fax: 0364-2538032; E-mail: [storeneigrihms@gmail.com](mailto:storeneigrihms@gmail.com).

Sd/-

Stores & Procurement Officer,  
For and on behalf of Director, NEIGRIHMS