

North Eastern Indira Gandhi Regional Institute of Health and Medical Sciences  
(An Autonomous Institute, Ministry of Health and Family Welfare, Government of India)  
Director's Block, Mawdiangdiang, Shillong 793 018 (Meghalaya)

Tender Amendment No: NEIGR/S&P/o8/o8/2019 -20

Dated: 13.08.2019

**e -TENDER AMENDMENT**

Reference Tender Enquiry No.:- NEIGR / S&P / OT / E -19/2019-20; dated 03/06/2019 for processing of OT Table against buy back of old /obsolete existing & unserviceable OT Table (Make: Takauchi; Year: 2007) for Operation Theatre.

The following amendment /addendum are hereby being considered against the technical specification:-

Point No.	Technical Specification	May be Read as
2.1	High quality suitable for General Surgery surgery Width 3.5 feet, length 6- 6.5 ft for the support of legs and breakable on all table positioning Mattress should be x-ray translucent. Should have all the accessories Electro- Hydraulic O.T. Table	High quality suitable for Orthopedic Surgery Width 1.5-2 feet, length 6- 6.5 ft for the support of legs and breakable on all table positioning Mattress should be x-ray translucent. Should have all the accessories Electro- Hydraulic O.T. Table
3.1	Operation Table should be electro- hydraulically operated with facilities for wireless remote control along with backup control on the table column side(Max movement functions must available) and have the following specifications.	Operation Table should be electro- hydraulically operated with control through wireless remote along with backup control on the table column side(Max movement functions must available) and have the following specifications. Wireless Remote must be supplied as standard supply.
3.1(b)	All table positioning i.e. Height, Back section, lateral tilt, Trendelenburg and Reverse – Trendelenburg, Upward Movement should be achieved by control on the table column side as well as by remote control. It should have a button for the homeposition.	All table positioning i.e. Height, Back section, lateral tilt, Trendelenburg and Reverse – Trendelenburg, Upward Movement should be achieved by control on the table column side as well as by remote control. It should have a button for the homeposition. It must have a beach chair position for different arthroscopy procedures.
3.1(e)	The casing on the frame and center supporting column should be made up of Hygienic stainless steel of high grade rustproof quality.(epoxy coated steel)	The casing on the frame and center supporting column should be made up of Hygienic stainless steel of high grade 316L rust proof quality.
3.1(f)	“Height 650-700 mm to 1100-1150mm or better, Trendelenburg / reverse Trendelenburg = -30 to +30 degree or better, back section = -45 to + 90 degree, side tilt = +/- 25 degree or more, motorized leg section = - 90 to 80 degree, detachable. Head rest = -45 to + 45 degree, manually, detachable. Facility perineal unit.	“Height 600-650 mm to 1050-1150mm or better, Trendelenburg / reverse Trendelenburg = -30 to +30 degree or better, back section = -45 to + 90 degree, side tilt = +/- 25 degree or more, motorized leg section = - 90 to + 35 degree, detachable. Head rest = -30 to + 45 degree, manually, detachable. Facility perineal unit.
3.1(h)	Remote control and additional controls on table base for emergency.	Remote control and additional controls on table base/column for emergency.
3.1(j)	The safe working load capacity of table should not be less than 400kg Idle/home position and 250kg or more in all positions in reverse mode	The safe working load capacity of table should not be less than 400kg Idle/home position and 230kg or more in all positions in reverse mode
3.1(i)	The table top has a motorized longitudinal slide facility of minimum 300 mm for better access & maximum scan area of Carm	The table top has a motorized longitudinal slide facility of minimum 400 mm for better access & maximum scan area of Carm.
4(g)	Gel Pads For positioning of patients	Gel Pads For all positioning of patients
4(m)	Additional Point	<b>Original Mayfield frame along attachment set must be provided along with the system. Compatible frame/set will not be considered.</b>

All other terms and conditions remains the same.

For further details regarding amendment, addendum, extension and downloading of documents, please visit website: [www.eprocure.gov.in](http://www.eprocure.gov.in) /[www.neigrihms.gov.in](http://www.neigrihms.gov.in); Tel/Fax: 0364-2538032; E-mail: [storeneigrihms@gmail.com](mailto:storeneigrihms@gmail.com).

Sd/-  
Stores & Procurement Officer,  
For and on behalf of Director, NEIGRIHMS