

North Eastern Indira Gandhi Regional Institute of Health and Medical Sciences
(An Autonomous Institute, Ministry of Health and Family Welfare, Government of India)
Director's Block, Mawdiangdiang, Shillong 793 018 (Meghalaya)

Tender Amendment No: NEIGR/S&P/07/08/2019 -20

Dated: 13.08.2019

e -TENDER AMENDMENT

Reference Tender Enquiry No.:- NEIGR / S&P / OT / E –34/2019-20; dated 03/06/2019 for processing of one unit of Steam Sterilizer against buy back of old /obsolete existing & unserviceable Central Sterile Stores Department (CSSD) equipments (Make: Yorco Sales Pvt. Ltd.).

The following amendment /addendum are hereby being considered against the technical specification:-

Point No.	Technical Specification	May be Read as
	Horizontal steam Sterilizer (Capacity 1000 liters): Qty. 1 Nos.	Horizontal steam Sterilizer : Qty. 1 Nos.
1	The Sterilizer should have the autoclaving chamber capacity of 900- 1000 liters.	The Sterilizer should have the autoclaving chamber capacity of 850- 1000 liters.
14	The front and rear both door seal should be durable, Non-lubricated Steam/Air activated door gasket made of silicon rubber.	The front and rear both door seal should be durable, Lubricated /Non-lubricated Steam/Air activated door gasket made of silicon rubber.
16	The Sterilizer should be supplied with Automatic and Manual opening in case of automatic mechanism failure of vertical sliding door. Also manual operation of autoclave in case of electrical failure to complete the cycle(Air operated manual valve)	The Sterilizer should be supplied with Automatic and Manual opening in case of automatic mechanism failure of vertical / Horizontal sliding door. Also manual operation of autoclave in case of electrical failure to complete the cycle(Air operated manual valve)
26	Steam generator should be made of 316 grade stainless steel material. The steam generator should have chloride free mineral wool or mineral glass wool thickness of 25 mm to 50 mm insulation sheet.	Steam generator should be made of 316 grade stainless steel material. The steam generator should have chloride free mineral wool or mineral glass wool / Melamine thickness of 25 mm to 50 mm insulation sheet.
27	The inbuilt Steam generator should have a built in thermostat, pressure safety valve, water level glass gauge.	The inbuilt/standalone Steam generator should have a built in thermostat, pressure safety valve, water level glass gauge/digital display.
69	The sterilizer must able to run with normal Tap water with TDS below 100 ppm.	The bidder should make necessary arrangement with respect to filtration and water pressure for the supplied tap water as per there system requirement.
70	The supplier should quote the RO plant of 250 ltrs per hour optionally as percentage value in the schedule XI of technical bid	The bidder should make necessary arrangement with respect to filtration and water pressure for the supplied tap water as per there system requirement.
6		Heating Elements, NRV and other related spares must be covered under warranty & CMC period.

All other terms and conditions remains the same.

For further details regarding amendment, addendum, extension and downloading of documents, please visit website: www.eprocure.gov.in /www.neigrihms.gov.in; Tel/Fax: 0364-2538032; E-mail: storeneigrihms@gmail.com.

Sd/-
Stores & Procurement Officer,
For and on behalf of Director, NEIGRIHMS