

North Eastern Indira Gandhi Regional Institute of Health and Medical Sciences
(An Autonomous Institute, Ministry of Health and Family Welfare, Government of India)
Director's Block, Mawdiangdiang, Shillong 793 018 (Meghalaya)

Tender Amendment No: NEIGR/S&P/09/08/2019 -20

Dated: 13.08.2019

e -TENDER AMENDMENT

Reference Tender Enquiry No.:- NEIGR / S&P / OT / E -19/2019-20; dated 03/06/2019 for processing of ENT Drill & accessories against buy back of 02 units of old /obsolete existing & unserviceable Microtome Drill (Make: Eagle Medical System; Year: 2009) for ENT department.

The following amendment /addendum are hereby being considered against the technical specification:-

Point No.	Technical Specification	May be Read as
1.1(3)	Should be able to adjust the irrigation levels and bur speed with the touch screen control.	Should be able to adjust the irrigation levels and bur speed with the touch screen /touch button control.
1.1(5)	Should have large Touch screen monitor	Should have large Touch screen monitor/Touch Button with display.
1.1(6)	The various parameters should be able to adjust either from touch screen panel or from the multifunction foot switch	The various parameters should be able to adjust either from touch screen/touch button panel or from the multifunction foot switch
1.1(7)	Should have multifunction ergonomically designed foot control with light emission for easy identification	Should have multifunction ergonomically designed foot control with light emission /clear marking for easy identification
1.1(10)	Stapes drill should have the speed of 12000 RPM	Deleted
1.1(11)	Mastoid drill should have the speed up to 60000 RPM. Brushless micro motor with inbuilt auto cooling system.	Mastoid drill should have the speed range 75000-80000 RPM. Brushless micro motor with inbuilt auto cooling system.
1.1(12)	Should have the provision to mount the console on various sizes of IV poles	Should have the provision to mount the console on various sizes of IV poles/can be placed on a flat surface.
1.1(1)	Should have High Speed (up to 80k) and ample torque	Should have High Speed 75k to 80 k and ample torque
1.1(5)	Should have wide variety of burs in sizes, lengths and materials as well as reusable line	Should have wide variety/universal of burs in sizes, lengths

All other terms and conditions remains the same.

For further details regarding amendment, addendum, extension and downloading of documents, please visit website: www.eprocure.gov.in /www.neigrihms.gov.in; Tel/Fax: 0364-2538032; E-mail: storeneigrihms@gmail.com.

Sd/-
Stores & Procurement Officer,
For and on behalf of Director, NEIGRIHMS