

North Eastern Indira Gandhi Regional Institute of Health and Medical Sciences
(An Autonomous Institute, Ministry of Health and Family Welfare, Government of India)
Director's Block, Mawdiangdiang, Shillong 793 018 (Meghalaya)

Tender Amendment No: NEIGR/S&P/07/06/2019 -20

Dated: 24.07.2019

e -TENDER AMENDMENT

Reference Tender Enquiry No.:- NEIGR / S&P / OT / E -10/2019-20 for processing of Thromboelastography System, Advanced Cardiac Output Monitor and DVT Prophylaxis System, with 5 years warranty and thereafter 5 years CMC, for department of Anesthesiology.

The following amendment /addendum are hereby being considered against the technical specification of Thromboelastography System:-

Point no1:-Should have two Independent measuring channels which can run simultaneously run two tests

May be read as

“Should have two / four independent measuring channels to run the test”

Point no 2:-Cup drive should be line synchronized, with synchronous motor.

May be read as

“Cup/Pin drive should be line synchronized, with synchronous motor”

Point no 5:-Measuring Technique should be on Shear elasticity of a coagulating sample, determined by motion of cup.

May be read as

“Measuring Technique should be on Shear elasticity of a coagulating sample, determined by motion of cup/pin”

Point no 10:-Low molecular weight heparin management.

May be read as

“Unfractionated heparin management”.

Point no:-Should have the facility of Platelet Mapping Assay to assess percentage inhibition of antiplatelet drugs such as clopidogrel, aspirin etc. in patients who have received platelet inhibiting drug.

May be read as

“Should have the facility of Platelet Mapping Assay(Optional) to assess percentage inhibition of antiplatelet drugs such as clopidogrel, aspirin etc. in patients who have received platelet inhibiting drug.”

Additional Point:-

6. Please quote the all the consumables & reagents required (including control & calibrator)for the following tests along with quantity as the percentage in the schedule xi of the technical bid which will be considered for evaluation of the price .

Sl. No.	Test Name	Test Per year	
1	Extrinsic Pathway activation	100	
2	Intrinsic Pathway activation	100	
3	Differentiation between Platelet and Fibrinogen Deficiency	100	
4	Check the Heparin- Protamine dose monitoring	100	
5	Identification of Factor deficiency	100	
6	Checking and confirming for hyper fibrinolysis	100	
7	Platelet activation with arachidonic acid.	100	
8	Platelet activation with adenosine Diphosphate.	100	
9	Platelet activation with thrombin activating peptide	100	

Note: The last date and time of submission has been extended upto 1400 hours of 19.07.2019.

All other terms and conditions remains the same.

For further details regarding amendment, addendum, extension and downloading of documents, please visit website: www.eprocure.gov.in /www.neigrihms.gov.in; Tel/Fax: 0364-2538032; E-mail: storeneigrihms@gmail.com.

Sd/-
Stores & Procurement Officer,
For and on behalf of Director, NEIGRIHMS