

**e -TENDER AMENDMENT**

Reference Tender Enquiry No.:- NEIGR / S&P / OT / E -13/2019-20; dated 03/06/2019 for processing of Calibration and Testing Equipments, with 5 years warranty and thereafter 5 years CMC, for Biomedical Engineering Unit.

The following amendment /addendum are hereby being considered against the technical specification of Vital Sign Monitor:-

1. Point No: 1:-LCD Touch –Screen color Display.

*'May be read as'*

“LCD Display.”

2. ECG RATE 30 BPM, 60 BPM, 80 BPM, 90 BPM, 120 BPM, 150 BPM, 180 BPM, 210 BPM, 240 BPM, 270 BPM, 300 BPM, and 320 BPM. Preset and monitor testing sequence hypertensive condition is at 40 BPM

*'May be read as'*

“ECG Rate 30-320 BPM.Preset and monitor testing sequence hypertensive condition is at 40 BPM”

**Additional Points:-**

3. OEM must have Authorized service center in India & Supplier must able give demonstration of the system prior to finalize the tender for all the listed equipment.

4. Yearly calibration & validation (along with traceability certificate ) of these equipment throughout the warranty & CMC period is the responsibility of the supplier/OEM/Manufacturer & stand by machine has to be provided prior to taking the systems for calibration every year .

Note: The last date and time of submission has been extended upto 1400 hours of 19.07.2019.

All other terms and conditions remains the same.

For further details regarding amendment, addendum, extension and downloading of documents, please visit website: [www.eprocure.gov.in](http://www.eprocure.gov.in) /[www.neigrihms.gov.in](http://www.neigrihms.gov.in); Tel/Fax: 0364-2538032; E-mail: [storeneigrihms@gmail.com](mailto:storeneigrihms@gmail.com).

Sd/-  
Stores & Procurement Officer,  
For and on behalf of Director, NEIGRIHMS