

North Eastern Indira Gandhi Regional Institute of Health and Medical Sciences
(An Autonomous Institute, Ministry of Health and Family Welfare, Government of India)
Director's Block, Mawdiangdiang, Shillong 793 018 (Meghalaya)

Tender Amendment No: NEIGR/S&P/06/06/2019 -20

Dated: 17.06.2019

e -TENDER AMENDMENT

Reference Tender Enquiry No.:- NEIGR / S&P / OT / E -02/2019-20; dated 03/06/2019 for processing of Automated Coagulation Analyzer against buy back of old existing obsolete and unserviceable Coagulation Analyzer (Make: Diamed, Switzerland; Model: Diamed CD-X Haemostasis) and 2 units of Fully Automated Electrolyte ABG Machine (Make: Roche Diagnostic; Model: Cobas b 221) and with fixed Reagents /Consumables on rate contract basis for a period of ten years, for department of Pathology.

The following amendment /addendum are hereby being considered against the tender mentioned above:-

Point no: 3 :-Principle based on change in viscosity by electromagnetic clot detection system with steel ball oscillation or multi wave length scanning and sample liquid-sensing technology.

'May be read as'

“Principle based on change in viscosity by electromagnetic clot detection system with steel ball oscillation or multi wave length scanning or Photo Optical and sample liquid-sensing technology with availability of clot curve/ waveform analysis with auto validation program.”

Point no 10: Minimum 96 sample positions with all STAT facility should be provided.

'May be read as'

“Minimum 50 sample position with continuous loading facility with compulsory STAT Facility should be provided.”

Point no 11: Refrigerated reagent positions of a minimum of 30 all at 15c should be available.

'May be read as'

“Refrigerated Reagent position of a minimum 40 all at 10-12C should be available.”

Point no 22: Minimum test menu available should include PT, APTT, Fibrinogen, TT,LA, ALL Factors,ATIII,Heparin, PC, PS, PLG, AP, APCR, DDI, FDP, FM, VWF

'May be read as'

“Minimum test menu available should include PT, APTT, Fibrinogen, TT, LA, ALL Factors,ATIII,Heparin, PC, PS, PLG, AP, APCR, DDI, and VWF”

Point no 25: *May be read as* “Deleted”.

Point no 30: *May be read as* “Startup Kit for at least 200 tests (All the consumables, reagents, control, calibrator, cleaning solution, Coolant, and other related accessories for 200 no's of PT& APTT test) should be provided along with the machine during the supply of the system”.

Additional Points:

- Test Throughput of the system should be minimum 150 Test /Hour for PT and minimum 110 Test / Hour for PT /APTT simultaneously.
- Counter Guarantee acknowledgement duly seal& signed the manufacturer/OEM for “Supply of reagents ,consumable reagents and other related accessories as quoted & provision of service throughout the warranty & CMC period “ must be attached in the tender .
- The system must be HL7 Compliant & have open architecture protocol to connect through Hospital HIS/LIS.
- Warranty and Maintenance:
 - a. Warranty for 5 years followed by CMC for 5 years including Spares & service.
 - b. Mandatory PM with unlimited breakdown calls has to be attended by the bidder/ manufacturer throughout the warranty & CMC period at site i.e. NEIGRIHMS, Shillong.
 - c. Duly Signed Mandatory PM & service Reports has to be submitted periodically, falling which necessary action will be initiated as per term& condition of the tender.

Documentation :

- 1 User / Technical / Maintenance manuals to be supplied in English.
- 2 Log book with instructions for daily, weekly, monthly and quarterly maintenance checklist. The job description of the hospital technician and company service engineer should be clearly spelt out.
- 3 Cost of spare parts, consumables & accessories which are not covered under warranty & CMC period has to quote in schedule XI as percentage value in the Technical Bid, or else will be consider to be cover throughout the warranty & CMC period.
- 4 Calibration and routine Preventive Maintenance Support as per manufacturer documentation in service / technical manual has to be done throughout the warranty & CMC period

- 5 Compliance report to be submitted in a tabulated and point wise manner clearly mentioning the page / Para number of original catalogue / data sheet and the offer details has to submit in the technical bid. Any point, if not substantiated with authenticated catalogue / manual, will not be considered.
- 6 Certificate of inspection and quality control indicating the S / N for all non-consumable items with date at the time of installation.

Note: The last date and time of submission has been extended upto 1400 hours of 19.07.2019.

All other terms and conditions remains the same.

For further details regarding amendment, addendum, extension and downloading of documents, please visit website: www.eprocure.gov.in / www.neigrihms.gov.in; Tel/Fax: 0364-2538032; E-mail: storeneigrihms@gmail.com.

Sd/-
Stores & Procurement Officer,
For and on behalf of Director, NEIGRIHMS