

**e -TENDER AMENDMENT**

Reference Tender Enquiry No: NEIGR/S&P/OT/E -14/2018 -19 for processing of 3.0 Tesla MR Scanner, with 5 years warranty, thereafter 5 year CMC and with complete turnkey works, for department of Radiodiagnosis

The following amendment /addendum are hereby being considered against the tender mentioned above:-

Point no: 4 ( RF Transmitter, Receiver, Coils) - RF Transmit

Latest RF transmit system (like Multi-Transmit /Multi Drive / True shape Parallel Transmit, etc) with at least two independent output source / Port should be offered, to improve RF uniformity and signal homogeneity and to reduce patient induced in homogeneities.

*'May be read as'*

Latest RF transmits system to improve RF uniformity and signal homogeneity and to reduce patient induced in homogeneities.

Note: The last date and time of submission has been extended upto 1400 hours of 07.06.2019.

All other terms and conditions remains the same.

For further details regarding amendment, addendum, extension and downloading of documents, please visit website: [www.eprocure.gov.in](http://www.eprocure.gov.in) /[www.neigrihms.gov.in](http://www.neigrihms.gov.in); Tel/Fax: 0364-2538032; E-mail: [storeneigrihms@gmail.com](mailto:storeneigrihms@gmail.com).

Sd/-  
Stores & Procurement Officer,  
For and on behalf of Director, NEIGRIHMS