

North Eastern Indira Gandhi Regional Institute of Health and Medical Sciences  
(An Autonomous Institute, Ministry of Health and Family Welfare, Government of India)  
Director's Block, Mawdiangdiang, Shillong 793 018 (Meghalaya)

Tender Amendment No: NEIGR/S&P/12/08/2019 -20

Dated: 30.08.2019

**e -TENDER AMENDMENT**

Reference Tender Enquiry No.:- NEIGR / S&P / OT / E -35/2019-20; dated 01/07/2019 for for processing of 03 units of ACT Analyzer against buy back of 03 units of old /obsolete existing & unserviceable ACT Analyzer (Make: Helena Laboratories; Year: 2008).

The following amendment /addendum are hereby being considered against the technical specification:-

Point No.	Technical Specification	May be read as
2.	It should be compact & portable for bed-side testing.	It should be compact & portable for bed-side testing with rechargeable batteries for battery backup.
3.	It should have inbuilt mechanism to heat the cartridge.	It should have inbuilt mechanism to heat the cartridge if required.
6	It should be capable of displaying two reports at one time.	It should be capable of displaying reports during test & able to store minimum 100 patient data
7	Hundred cartridges for each test to be supplied with machine.	Deleted
9	Measurement range 0-1500 sec.	Measurement range 0-1000 sec or better
14	Dual well testing method.	Double type of clot detection one of mechanical type & Optical type
15	Accepts actalyte ACT tubes with celite, glass bead activator, MAX ACT tubes with blended activator; all international technidyne Hemochron tubes.	Accepts ACT tubes / cartridges like Celite / Kaolin for different types of tests.
19	Desirable: Rate of Actual Clot Formation (CR, Clot Rate: Thrombin Activity, Low Molecular Weight Heparin Management).	Desirable: Rate of Actual Clot Formation (CR, Clot Rate: Thrombin Activity, Low Molecular Weight Heparin Management,Low Molecular Weight Heparin Management).
23	Cartridge /Tubes/Consumables required for 200 tests must be supplied along with the system .	Cartridge /Tubes/Consumables required for 300 tests must be supplied along with the system & Individual Cartridge /Tubes/Consumables required must be quoted in percentage of the FOB value in the Schedule XI of the technical bid. <u>Cost of the Cartridge /Tubes/Consumables required 500tests per year for ten years will be calculated for the price evaluation.</u>

All other terms and conditions remains the same.

For further details regarding amendment, addendum, extension and downloading of documents, please visit website: [www.eprocure.gov.in](http://www.eprocure.gov.in) /[www.neigrihms.gov.in](http://www.neigrihms.gov.in); Tel/Fax: 0364-2538032; E-mail: [storeneigrihms@gmail.com](mailto:storeneigrihms@gmail.com).

Sd/-  
Stores & Procurement Officer,  
For and on behalf of Director, NEIGRIHMS