

North Eastern Indira Gandhi Regional Institute of Health and Medical Sciences
(An Autonomous Institute, Ministry of Health and Family Welfare, Government of India)
Director's Block, Mawdiangdiang, Shillong 793 018 (Meghalaya)

Tender Amendment No: NEIGR/S&P/14/01/2019 -20

Dated: 27.01.2020

e -TENDER CORRIGENDUM

Reference Tender Enquiry No.:- NEIGR /S&P /OT /E –09/ 2019-20 for processing of Intra Aortic Balloon Pump (IABP) Machine against buy back of old existing obsolete and unserviceable IABP Machine (Make: Datascope Corp., USA; Model: Datascope System 98; Year: 2007), with 5 years warranty and thereafter 5 years CMC, for department of CTVS.

The following corrigendum is hereby being considered against the technical specification of the above system:-

Sl. No.	Technical Specification	May be read as
3.1	Drive system: Stepper motor driven bellows	Drive system: Stepper motor driven bellows/reciprocating compressor based.
3.10	System should be approved for use on Adult patients of different sizes as well as Pediatrics patients.	System should be approved for use on Adult patients of different sizes.
3.17	Should have software which allows the user to monitor the IABP from any remote location via a modem	Deleted
3.19	Should have capability to connect on the Hospital network	Should be HL7 /HIPPA compliant
4.iii	Intra-Aortic Balloon Catheter for Adults, Size: 30 cc Qty: 2 Nos.	Intra-Aortic Balloon Catheter for Adults, Size: 40 cc Qty: 2 Nos.
6.3	Cost of spare parts, consumables (Battery Ballon catheters etc.) and accessories (ECG Cable, Hylyum Cylinder, Pressure transducer cable, Pressure transduser, Printer paper etc ..) which are not covered under warranty & CMC period has to quote in schedule XI as percentage value in the Technical Bid, or else will be consider to be cover throughout the warranty & CMC period.	Cost of spare parts, consumables (Battery ,Ballon catheters etc.) and accessories(ECG Cable, Helium Cylinder, Pressure transducer cable, Pressure transduser, Printer paper etc ..) which are not covered under warranty & CMC period has to quote in schedule XI as percentage value in the Technical Bid which will be fixed for entire contract period or else will be consider to be cover throughout the warranty & CMC period/bid will be treated as technically non-compliant .

All other terms and conditions remains the same.

For further details regarding amendment, addendum, extension and downloading of documents, please visit website: www.eprocure.gov.in /www.neigrihms.gov.in; Tel/Fax: 0364-2538032; E-mail: storeneigrihms@gmail.com

Sd/-
Stores & Procurement Officer,
For and on behalf of Director, NEIGRIHMS