

Bid Document

Bid Details	
Bid End Date/Time	07-08-2020 19:00:00
Bid Opening Date/Time	07-08-2020 19:30:00
Bid Life Cycle (From Publish Date)	90 (Days)
Bid Offer Validity (From End Date)	60 (Days)
Ministry/State Name	Ministry Of Health And Family Welfare
Department Name	Department Of Health And Family Welfare
Organisation Name	North Eastern Indira Gandhi Regional Institute Of Health And Medical Sciences (neigrihms)
Office Name	Neigrihms, Shillong
Total Quantity	1
Item Category	BOD Incubator
Bidder Turnover (Last 3 Years)	1 Lakh (s)
OEM Average Turnover (Last 3 Years)	1 Lakh (s)
Experience Criteria	3 Year (s)
MSE Exemption for Years of Experience and Turnover	No
Startup Exemption for Turnover	Yes
Document required from seller	Experience Criteria,Past Performance,Bidder Turnover,Certificate (Requested in ATC),OEM Authorization Certificate,OEM Annual Turnover *In case any bidder is seeking exemption from Experience / Turnover Criteria, the supporting documents to prove his eligibility for exemption must be uploaded for evaluation by the buyer
Past Performance	30 %
Bid to RA enabled	No
Inspection Required	No
Estimated Bid Value	100000

EMD Detail

Required	No
----------	----

ePBG Detail

Required	No
----------	----

Splitting

Bid splitting not applied.

1. Experience Criteria: In respect of the filter applied for experience criteria, the Bidder or its OEM {themselves or

through reseller(s)} should have regularly, manufactured and supplied same or similar Category Products to any Central / State Govt Organization / PSU / Public Listed Company for number of years as indicated in the bid document before the bid opening date. Copies of relevant contracts to be submitted along with bid in support of having supplied some quantity during each of the year. In case of bunch bids, the category of primary product having highest value should meet this criterion.

2. OEM Turn Over Criteria: The minimum average annual financial turnover of the OEM of the offered product during the last three years, ending on 31st March of the previous financial year, should be as indicated in the bid document. Documentary evidence in the form of certified Audited Balance Sheets of relevant periods or a certificate from the Chartered Accountant / Cost Accountant indicating the turnover details for the relevant period shall be uploaded with the bid. In case the date of constitution / incorporation of the OEM is less than 3 year old, the average turnover in respect of the completed financial years after the date of constitution shall be taken into account for this criteria.

3. Past Performance: The Bidder or its OEM {themselves or through re-seller(s)} should have supplied same or similar Category Products for 30% of bid quantity, in at least one of the last three years before the bid opening date to any Central / State Govt Organization / PSU / Public Listed Company. Copies of relevant contracts (proving supply of cumulative order quantity in any one year) to be submitted along with bid in support of quantity supplied in the relevant year. In case of bunch bids, the category related to primary product having highest bid value should meet this criterion.

BOD Incubator (1 pieces)

Technical Specifications

[* As per GeM Category Specification](#)

Specification	Specification Name	Values	Bid Requirement (Allowed Values)
ITEM	Item	BOD Incubator	*
DESIGN	Features available in BOD Incubator	Toughened glass window,Air circulation fan,Removable tray,Temperature sensor,Door hinges,Door handle with lock & key,PID temperature controller,Analog ampere meter,Pilot lamp,Safety thermostat,On/Off MCB,Refrigeration system,Power cord,Caster wheels,Solid door,Heater,Temperature High/Low Safety alarm	*
	Optional Features available in BOD Incubator	Digital event timer	*
	No of Trays	3	2, 3, 4, 5
	Air Circulation	Axial fan	Motorized Blower, Axial fan
MATERIAL	Exterior body material	Powder Coated GI sheet	CR Steel Sheet, Powder Coated GI sheet, Stainless Steel 304
	Interior body material	Stainless Steel	*
	Insulation material	PUF	PUF, Polystyrene Foam, Glass wool
	Outer door material	Powder Coated GI sheet	Stainless Steel 304, Mild Steel, Powder Coated GI sheet
	Inner door material Material of removable tray	Acrylic Sheet	Toughed Glass, Acrylic Sheet

DIMENSION	Chamber Capacity(in Itr)	150-199	*
	Overall length(in mm) Overall Width(in mm) Overall Height(in mm) Thickness of sheet (in MM) Thickness of Toughed Glass/Acrylic Sheet (in MM) Diameter of Stainless Steel wire in mesh (in MM)	505	*
	Overall Width(in mm)	800	*
	Overall Height(in mm)	450	*
	Thickness of sheet (in MM)	0.7	0.7, 0.8, 0.9, 1.0, 1.1, 1.2
	Thickness of Toughed Glass/Acrylic Sheet (in MM)	7	7, 8, 9, 10, 11, 12, 13, 14, 15
	Diameter of Stainless Steel wire in mesh (in MM)	NA	3, 4, 5, 6, 7, NA
	Wire Length(in mtr)	1.5	*
POWER	Temperature Range	5°C to 60°C	5°C to 60°C, 10°C to 70°C, 15°C to 80°C, 20°C to 90°C
	Temperature Accuracy	± 0.5°C	± 0.5°C, ± 1.0°C, ± 2.0°C
	Temperature Uniformity	± 1°C	± 1°C, ± 2°C, ± 5°C
	Power supply	230 Volts 50 Hz	*
WARRANTY	Warranty(in year)	1	1, 2, 3, 4 Or higher

* Specifications highlighted in bold are the Golden Parameters.

* Bidders may note that In respect of non-golden Parameters, the specifications 'Values' chosen by Buyer will generally be preferred over 'Bid requirement (allowed Values) by the Buyer.

Additional Specification Documents

Consignees/Reporting Officer and Quantity

S.No.	Consignee/Reporting Officer	Address	Quantity	Delivery Days
1	Khrawkumar Jithod Katrai	793018,P.O. NEIGRIHMS, Mawdiangdiang, Shillong	1	30

Bid Specific Additional Terms and Conditions

- 1.Experience Certificate for the supply of the same to any Govt/ PSU/ any renowned private organisation along with Supply/ Purchase Order.
- 2.If the agency is registered under MSME or NSIC, then EMD exemption certificate needs to be enclosed.
- 3.Make in india specific authorisation certificate needs to be enclosed.
- 4.Bidder's offer is liable to be rejected if they don't upload any of the certificates / documents sought in the Bid document, ATC and Corrigendum if any.
- 5.ISO 9001: The bidder must have ISO 9001 certification.
- 6.Material Test Certificate Should Be Sent Along with The Supply. The Material Will Be Checked by Buyer's Lab & the Results of the Lab will be the Sole Criteria for Acceptance of the Item.
- 7.The bidder is required to upload, along with the bid, all relevant certificates such as BIS licence, type test certificate, approval certificates and other certificates as prescribed in the Product Specification given in the bid document.
- 8.Successful bidder will have to ensure that adequate number of dedicated technical service personals / engineers are designated / deployed for attending to the Service Request in a time bound manner and for ensuring Timely Servicing / rectification of defects during warranty period, as per Service level agreement indicated in the relevant clause of the bid.

9. IMPORTED PRODUCTS: In case of imported products, OEM or Authorized Seller of OEM should have a registered office in India to provide after sales service support in India. The certificate to this effect should be submitted.

[This Bid is also governed by the General Terms and Conditions](#)

---Thank You---