

Bid Document

Bid Details	
Bid End Date/Time	07-08-2020 19:00:00
Bid Opening Date/Time	07-08-2020 19:30:00
Bid Life Cycle (From Publish Date)	90 (Days)
Bid Offer Validity (From End Date)	60 (Days)
Ministry/State Name	Ministry Of Health And Family Welfare
Department Name	Department Of Health And Family Welfare
Organisation Name	North Eastern Indira Gandhi Regional Institute Of Health And Medical Sciences (neigrihms)
Office Name	Neigrihms, Shillong
Total Quantity	1
Item Category	Deep freezer
Bidder Turnover (Last 3 Years)	2 Lakh (s)
OEM Average Turnover (Last 3 Years)	2 Lakh (s)
Experience Criteria	3 Year (s)
MSE Exemption for Years of Experience and Turnover	No
Startup Exemption for Turnover	Yes
Document required from seller	Experience Criteria,Past Performance,Bidder Turnover,Certificate (Requested in ATC),OEM Authorization Certificate,OEM Annual Turnover *In case any bidder is seeking exemption from Experience / Turnover Criteria, the supporting documents to prove his eligibility for exemption must be uploaded for evaluation by the buyer
Past Performance	30 %
Bid to RA enabled	No
Inspection Required	No
Estimated Bid Value	300000

EMD Detail

Required	No
----------	----

ePBG Detail

Required	No
----------	----

Splitting

Bid splitting not applied.

1. Experience Criteria: In respect of the filter applied for experience criteria, the Bidder or its OEM {themselves or

through reseller(s)} should have regularly, manufactured and supplied same or similar Category Products to any Central / State Govt Organization / PSU / Public Listed Company for number of years as indicated in the bid document before the bid opening date. Copies of relevant contracts to be submitted along with bid in support of having supplied some quantity during each of the year. In case of bunch bids, the category of primary product having highest value should meet this criterion.

2. OEM Turn Over Criteria: The minimum average annual financial turnover of the OEM of the offered product during the last three years, ending on 31st March of the previous financial year, should be as indicated in the bid document. Documentary evidence in the form of certified Audited Balance Sheets of relevant periods or a certificate from the Chartered Accountant / Cost Accountant indicating the turnover details for the relevant period shall be uploaded with the bid. In case the date of constitution / incorporation of the OEM is less than 3 year old, the average turnover in respect of the completed financial years after the date of constitution shall be taken into account for this criteria.

3. Past Performance: The Bidder or its OEM {themselves or through re-seller(s)} should have supplied same or similar Category Products for 30% of bid quantity, in at least one of the last three years before the bid opening date to any Central / State Govt Organization / PSU / Public Listed Company. Copies of relevant contracts (proving supply of cumulative order quantity in any one year) to be submitted along with bid in support of quantity supplied in the relevant year. In case of bunch bids, the category related to primary product having highest bid value should meet this criterion.

Deep Freezer (1 pieces)

Technical Specifications

[* As per GeM Category Specification](#)

Specification	Specification Name	Values	Bid Requirement (Allowed Values)
Standards	Conformity to Standards	CE (with 4 digit notified number)	CE (with 4 digit notified number)
	Certification date and number	-	*
	Conformity to electrical safety	IEC 61010-1, IEC 61010-2-40 or IS 13450	*
Performance Parameters	Purpose	For storage of various biological products including ATCC cultures, enzymes, vaccines, chemicals or material testing components for a longer period of time	*
	Capacity in liters	250-300	250-300
	Type of cabinet	Vertical	Vertical
	Ground Clearance in mm	100 mm	*
	Cooling type	Direct cooling	*
	Freezer body	Galvanised steel body with tough powder coated exterior finish	*
	Castors	Heavy duty lockable	Heavy duty lockable
	Lockable outer doors and lids	Yes	*
	Material of internal cabinet	SS	*
	Freezer should use CFC-free, HCFC-free non flammable refrigerants	Yes	*
	Refrigeration system must be energy efficient	Yes	*
	Pull down time in hrs	5.2 hrs	*

Operating Temperature in degree celcius	minus 10 degree to minus 40	*	
Increment variation setting	1° C	*	
Number of corrosion resistant racks	3	3	
Control	Fully programmable microprosessor controlled with membrane keypad and eye level control panel	*	
Number of inner stroage compartments with insulated doors and	3	3	
Digital temperature display	Yes	*	
Should have battery back up for the display and security lock for the display	Yes	*	
Alarms	Audible and visible alarm system for unwanted temperature rise, sudden power failure , system failure and battery low	*	
Noise level in dB	47	*	
Energy consumption rate (@KWh/day)	1.5	*	
Stabilizer should be capable to run any voltage between 190V - 270 V	Yes	*	
Warranty of Stabilizer in Years	5 years	*	
Warranty of Freezer in Years (from the date of installation)	2	2	
Additional Parameters	Freezer must have washable condenser filter indicator which should keep fins free of dust to maintain peak cooling efficiency	Yes	*
	One 5KVA servo stabilizer to be supplied with freezer	Yes	*
	CO₂ cylinder should be supplied with freezer for backup	No	No
	Tempearture Chart recorder to be provided	Yes	Yes
Reports	Manufacturer should have ISO certification	Yes	*
	ISO certification date and number	-	*
	Copies of all certifications and reports to be provided to buyer on demand at time of supplies	Yes	*

* Specifications highlighted in bold are the Golden Parameters.

* Bidders may note that In respect of non-golden Parameters, the specifications 'Values' chosen by Buyer will generally be preferred over 'Bid requirement (allowed Values) by the Buyer.

Additional Specification Documents

Consignees/Reporting Officer and Quantity

S.No.	Consignee/Reporting Officer	Address	Quantity	Delivery Days
1	Khrawkupar Jithod Katrai	793018,P.O. NEIGRIHMS, Mawdiangdiang, Shillong	1	30

Bid Specific Additional Terms and Conditions

- 1.Experience Certificate for the supply of the same to any Govt/ PSU/ any renowned private organisation along with Supply/ Purchase Order.
- 2.If the agency is registered under MSME or NSIC, then EMD exemption certificate needs to be enclosed.
- 3.Make in india specific authorisation certificate needs to be enclosed.
- 4.IMPORTED PRODUCTS: In case of imported products, OEM or Authorized Seller of OEM should have a registered office in India to provide after sales service support in India. The certificate to this effect should be submitted.
- 5.Bidder's offer is liable to be rejected if they don't upload any of the certificates / documents sought in the Bid document, ATC and Corrigendum if any.
- 6.ISO 9001: The bidder must have ISO 9001 certification.
- 7.Material Test Certificate Should Be Sent Along with The Supply. The Material Will Be Checked by Buyer's Lab & the Results of the Lab will be the Sole Criteria for Acceptance of the Item.
- 8.The bidder is required to upload, along with the bid, all relevant certificates such as BIS licence, type test certificate, approval certificates and other certificates as prescribed in the Product Specification given in the bid document.
- 9.Successful bidder will have to ensure that adequate number of dedicated technical service personals / engineers are designated / deployed for attending to the Service Request in a time bound manner and for ensuring Timely Servicing / rectification of defects during warranty period, as per Service level agreement indicated in the relevant clause of the bid.

[This Bid is also governed by the General Terms and Conditions](#)

---Thank You---