

Critical Care Block Furniture on Complete Turnkey Basis under PM-ABHIM			
GEM / 2024 / B / 4791171, Dated 16/03//2024			
Reference queries from M/S arti enterprises Dated 12/03/2024 ,M/S Facio Marketing Solution vide emails dated 10/05/2024 andM/S Biosis Medical Services vide email dated30/04/2024 and 29/05/2024 ,M/S RGS Enterprise vide email dated 24/05/2024,vide letter dated 02/05/2024 & M/S Elk Drugs Distributor vide email dated 20/05/2024,Bharat Surgical Impex vide letter dated 15/05/2024 Follwing amendment/Modification are placed			
S. No.	Specification mentioned in the GeM bid document	Accepted corrigendum	Remarks
Item no 1.MOTORIZED ICU BED 4 SECTION WITH MATTRESS			
	Should have angle indicator for back & TR/RTR/digital display	Should have Digital Angle Indicator for back & TR/RTR in digital display	Clarified on digital display
	Patient weight load of minimum 150 kg	Patient weight load of minimum 180 kg	For safety
	Shall meet IEC-60601-1-2:2001(Or Equivalent BIS) General Requirements of Safety for Electromagnetic Compatibility. Or should comply with 89/366/EEC; EMC-directive.	No change	As per Govt Of India ,Dept of Pharmaceuticals public Notice no 31026/83/2021/MD dated 1/02/2022 and MoHFW OM No 35017/2/2010- Policy dated 31/08/2023
Hospital Fowler Bed(2 section)with mattress			
7	Overall Size 210 l x 90 w x 60-8	Overall Size 210 l x 90 w x 60-8	For wider participation
10	Bed mounted on 125 mm diameter two brakes/Individual lock Bed mounted on 125 mm d	Bed mounted on 125 mm diameter two brakes/Individual lock unlock castors	For patient safety

S. No.	Specification mentioned in the GeM bid document	Accepted corrigendum	Remarks
	Bed must have Manually operated 3 crank system which can be folds away underneath the bed for various positions like Head section, Leg section & back section	Bed must have Manually operated 2 crank system which can be folds away underneath the bed for various positions like backrest and thigh (knee break)adjustment .	For wider participation
	C-Four section mattress	Single section Mattress	For wider participation
	ed must have Safe working load 200 Kg & Max. Patient weight 160 kg	ed must have Safe working load 150 Kg	For wider participation
Hospital Fowler Bed(4 section) with mattress			
3	Backrest, knee rest, TB/RTB and hi-lo positions maneuvered by separate.	Backrest, knee rest, TB/RTB and hi-lo adjustment must be avaialbale .	For wider participation
7	Overall Size 210 l x 90 w x 60-80Hcms	Overall Size 210 l x 90 w x 60-8	For wider participation
10	Bed mounted on 125 mm d	Bed mounted on 125 mm diameter two brakes/Individual lock unlock castors	For patient safety
	Bed must have Manually operated 3 crank system which can be folds away underneath the bed for various positions like Head section, Leg section & back section	Bed must have Manually operated 3/4 crank system which can be folds away underneath the bed for various positions like Backrest, knee rest, TB/RTB and hi-lo adjustment.	For wider participation
	Manually operated crank allows adjusting backrest (to 72-80degree), leg section and foot section	Manually operated crank allows adjusting backrest (0 to 32 degree), leg section and foot section	For wider participation

S. No.	Specification mentioned in the GeM bid document	Accepted corrigendum	Remarks
	ed must have Safe working load 200 Kg & Max. Patient weight 160 kg	ed must have Safe working load 150 Kg	For wider participation
Crash Cart			
2	Material of Crash cart Console - Steel with high density polymer finish	MS with powder coated	For strudyness
3	Number of Baskets in over bridge (Nos) - 6	Number of Baskets in over bridge (Nos) - 6 nos Handout Bins	Clarified on basket type
23		Number of Utility Trays in over bridge / Handout Bins-6 nos	Clarified on bin type
24	Shall have a over bridge of ABS thermoplastic with baskets and Utility Trays for keeping consumables	It have a over bridgemade from Mild Steel (MS) with powder coated and Utility Trays for keeping consumables	For strudyness
26	Height of over bridge from the trolley (mm) - 1000	Height of over bridge from the trolley (mm) - 600 mm	For wider participation
Bed side locker			
ix	CPR board - 1 nos must be supplied along with each bed	Deleted	Error in the specification
x	Optional Accessory: Traction pulley attachment for trauma cases.	Deleted	Error in the specification
Item no 80. LDR Bed (Labour Delivery From Bed Electric(LDR Bed)			
3	It should have control devise for making height and back adjustments.[Manual as well as remote control].	It should have control devise for making height and back adjustments.[Manual CPR Lever Should be present in Bed].	For wider participation

S. No.	Specification mentioned in the GeM bid document	Accepted corrigendum	Remarks
6	It should have three sectional mattress and seat section should have large perineal cut.	Should have push grip handel & have a removavable Plastic ABS bowl at perinal part of the table to collect Fluid	For wider participation
7	Mattress should be made of high resilience high quality P.U foam with stain resistant cover and zip.	Mattress should be made of of 150mm high resilience high quality P.U foam with stain resistant cover and zip.	Clarified on the thickness of the mattress
15	It should be able to give trendelenburg, reverse trendelburg and 60 - 70 degree sitting position both mechanically and electronically.	15. It should be able to give trendelenburg, and 60 - 70 degree sitting position both mechanically and electronically.	For wider participation
	Shall meet IEC-60601-1-2:2001(Or Equivalent BIS) General Requirements of Safety for Electromagnetic Compatibility. Or should comply with 89/366/EEC; EMC-directive.	No change	As per Govt Of India ,Dept of Pharmaceuticals public Notice no 31026/83/2021/MD dated 1/02/2022 and MoHFW OM No 35017/2/2010- Policy dated 31/08/2023

List of Furniture for CCB-NEIGRIHMS

	Tendered item with quantity	Amended item with quantity	
	Item no 87 General Furniture & ancillary furniture AS mentioned in the ATC	Deleted	Only Hospital/medical furniture included for wider participation .Separate bidding with detailed boq for general/Office furniture

5% deviations is acceptable in dimension mentioned in the specifications for all the furniture

Successful Bidder have to obtain respective trading license from KHADC ,Meghalaya