

E -TENDER AMENDMENT

Reference Tender Enquiry No: NEIGR/S&P/OT/E-05A/2024-25 for processing of Processing of Implants /Pacemaker /Accessories, etc, on case to case consignment basis, on rate contract for a period of two years, extendable upto 6 months or till the finalization of the next tender, whichever is later, for department of Cardiology.

The following amendment /addendum are hereby considered against the above tender.

Sl. No.	Category	Description of Stores /Items	Unit /Pack	Manufacturer /Company
1.	Leadless Pacemaker (Sl. No. 6.01 of BOQ)	May be read as 'US FDA /CE/DCGI/BIS approved · Volume < 0.9 cc · Weight of pacemaker < 2.0 grams · Atraumatic fixation of pacemaker to minimize perforation risk. · MRI conditional in 1.5 or 3.0 T environment. Automatically adjust sensitivity to maintain adequate sensing margins. · Must have ventricular capture management and with an active fixation mechanism that can be retracted at the time of replacement'	Each	
2.	Conduction System Pacing (Sl. No. 22 of BOQ)	May be read as 'Pre-shaped catheter for Conduction System Pacing compatible with lumen less lead. <ul style="list-style-type: none"> • Catheter should be able to deliver all Active Pacing fixation, stylet driven extendable helix leads with diameter of 6Fr or below. • Multiple options of shape availability for various anatomies. • Catheter should have integrated Hemostasis valve along with facility of contrast injection port. • Atraumatic radiopaque tip to facilitate imaging under fluoroscopy. • Catheter should have outer diameter of 9fr or less. • Should be supplied with compatible 0 .035" guidewire and slitter. • Helix Locking Tool (HLT) along with the catheter provides alternative to the clip – on tool used for retracting /extending of helix' 	Each	
3.	Manufacturer has to submit; <ol style="list-style-type: none"> 1. Performance Certificate from any Central Government-reputed Institution. 2. Proof of regular supply (<i>invoice copies</i>) for the last 3 years in an institution of national repute. 			

***The last date and time for submission has been extended upto 1400 hours of 11.11.2024

All other terms and conditions remains the same.