


BB/CT/G- 383/2025 - 26, dt: 25/6/2025



Sanction Order

Sanction No: 511687735515239
Sanction Date: 27-May-2025

Sanction of the competent authority is hereby conveyed for incurring an expenditure of amount as under towards the cost of Purchase order/Contract placed on the Seller for Supply of Goods/Services as per the contract for making payment to the Seller subject to deduction of TDS as applicable:

Organisation Details		Buyer Details	
Type:	Central Autonomous	Name:	Harendra Dey
Ministry:	Ministry of Health and Family Welfare	Designation:	Stores and Procurement Officer
Department:	Department of Health and Family Welfare	Email ID:	harendra.dey@nic.in
Organisation Name:	North Eastern Indira Gandhi Regional Institute of Health and Medical Sciences (NEIGRIHMS)	GSTIN:	N
Office Name:	Neigrihms, Shillong	Address:	P.O. NEIGRIHMS, Mawdiangdiang, Shillong Khasi Hills East MEGHALAYA - 793018

Financial Approval Detail	
C-760/2025 - 26, dt: 16/5/2025	
Designation of official providing Administration approval:	Director- agenda No: 27/EC-7 dt. 21st March 2025, file no STOPRO-CTVS/1/2024-Stores, Compulsory KHADC trading license as per tender conditions at NEIGRIHMS Shillong
IFD Concurrence / Competent Authority (HOD / Head of Office) Approval Required?	YES
Budget availability	YES
Designation of official providing Financial approval:	DDA & DFA, C- 760/25-26 Dt. 16.5.2025, Tripartite Agreement for provision of service as per T&C with OEM to be submitted with in 21days
Designation Function/Budget head of Account:	NA
IFD/Competent Authority Dtd No:	C760
IFD/Competent Authority Dtd Date:	2025-05-16
Financial Year:	NA
DDO:	NA
PD Code:	NA
Grant No:	NA

Minimum Reserve Declared Value
~ RS: 10,300,43.

Heart Lung machine 1 set (Make: Sano, Model: APS-1 as per order No: HSCC/PUR/NEIGR/2006/HL mach/33, dt: 27/4/2007

Seller Details	
Company Name:	SAIMED INNOVATION
Email ID:	shishir@saimed.in
Address:	SAIMED INNOVATION Kolkata WEST BENGAL - 700013

Product Details						
#	Item Description	Model	Ordered Quantity	Unit	Price per Unit inclusive of all Duties and Taxes (in INR)	Total Price (inclusive of all Duties and Taxes (in INR))
1	HEART LUNG MACHINE AND HEATER COOLER MACHINE	HEART LUNG MACHINE AND HEATER COOLER MACHINE	1	pieces	24800000.0	24800000.0
Total Order Value (in INR)						24800000.0

Consignee Details						
S.No	Consignee	Item	Lot No.	Quantity	Delivery Start After	Delivery To Be Completed By
1	Khawkar Jithod Katia con18.neigrihms.mil@nic.in P.O. NEIGRIHMS, Mawdiangdiang, Shillong KHASI HILLS EAST MEGHALAYA - 793018	HEART LUNG MACHINE AND HEATER COOLER MACHINE	-	1	27-May-2025	25-Aug-2025

Terms & Conditions	
1. This issues under the Financial Power Rules, 1978 as amended from time to time or as per applicable delegation of financial power rules as approved and amended time to time by the competent authority of the buyer organization.	
Note: This system generated file. No signature is required. Print out of this document is not valid for payment/ transaction purpose.	

Handwritten signatures and stamps at the bottom of the document.

B/B/IT/G-383/2025-26, dt: 25/6/2025

अनुबंध | Contract



अनुबंध क्रमांक | Contract No: GEMC-511687735515239

अनुबंध तिथि | Generated Date : 27-May-2025

बोली/आपूर्ति/पीपी संख्या | Bid/RA/PBP No.: GEM/2024/B/5400916

संगठन विवरण Orgar		sation Details		खरीदार विवरण Buyer Details		STOPRO-CTVS/1-2024-141	
प्रकार Type :		Central Autonomous		पद Designation :		Stores and Procurement Officer	
मंत्रालय Ministry :		Ministry of Health and Family Welfare		संपर्क नंबर Contact No. :		0364-2539032-213	
विभाग Department :		Department of Health and Family Welfare		ईमेल आईडी Email ID :		harendra.dey@nic.in	
संगठन का नाम Organisation Name :		North Eastern Indira Gandhi Regional Institute of Health and Medical Sciences (NEIGRIHMS)		जीएसटीआईएन GSTIN :		N	
कार्यालय क्षेत्र Office Zone :		neighrims, Shillong		पता Address :		P.O. NEIGRIHMS, Mawdlangdlang, Shillong, KHASI HILLS EAST, MEGHALAYA-793018, India	
वित्तीय स्वीकृति विवरण 		Financial Approval Detail		भुगतान प्राधिकरण विवरण 		Paying Authority Details	
आईएफडी सहमति IFD Yes		Concurrence :		Role:		PAO	
प्रशासनिक अनुमोदन का पदनाम		Direct agenda No: 27/EC-7 dt. 21st March 2025. file no STOPRO-C		भुगतान का तरीका		Offline	
Designation of TVS/1		1024-Stores, Compulsory KHADC trading license as per tender		पद Designation :		Thwet Star Syngkon	
Administrative conditions at NEIGRIHMS Shillong		Approval:		ईमेल आईडी Email ID :		thwet.syngkon@neighrims.gov.in	
वित्तीय अनुमोदन का पदनाम		DDA, DFA, C-760/25-26 Dt. 16.5.2025, Tripartited Agreement for pr		जीएसटीआईएन GSTIN :		-	
Designation of Financial Approval :		of service as per T&C with OEM to be submitted with In 21da		पता Address:		P.O. NEIGRIHMS, Mawdlangdlang, Shillong, KHASI HILLS EAST, MEGHALAYA-793018, India	
C 760/2025-26, dt: 16/5/2025				Minimum declared value - ₹: 10,300,43.			
विक्रेता विवरण Seller		etails					
जेम विक्रेता आईडी GeM Sell ID :		0EAA200001263196					
कंपनी का नाम Company Name :		SAIMED INNOVATION					
संपर्क नंबर Contact No. :		09830119232					
ईमेल आईडी Email ID :		shishir@saimed.in					
पता Address :		412, KAMALALAYA CENTRE, 156A LENIN SARANI, DHARAMTOLLA, Kolkata, WEST BENGAL-700013, -					
एमएसएमई पंजीकरण संख्या ME Registration number :		UDYAM-WB-10-0001774					
जीएसटीआईएन GSTIN:		19ACDFS0128G1ZB (R)					
खरीदार द्वारा मूल्यांकित एमआईई		स्थिति MII Status as					
evaluated by buyer :		Not Verified					
खरीदार द्वारा सत्यापित एमएसएमई		स्थिति MSME Status as					
verified by buyer :		Not Verified					
एमएसएमई सामाजिक श्रेणी MSE Social Category :		General					
एमएसएमई लिंग श्रेणी MSE Gender :		Male					
* जिसके नाम के पक्ष		GST/TAX इनवॉइस पेश किया जाएगा GST / Tax invoice to be raised in the name of - Buyer					
वितरण निर्देश Delivery Instructions :		NA					
उत्पाद विवरण Product		Details					
#	आइटम विवरण Item Description	आइटम विवरण Ordered Quantity	इकाई Unit	इकाई मूल्य (INR) Unit Price (INR)	कर विभाजन (INR) Tax Bifurcation (INR)	मूल्य (INR में सभी शुल्क और कर सहित) Price (Inclusive of all Duties and Taxes in INR)	
1	उत्पाद का नाम Product Name : HEART LUNG MACHINE AND HEATER COOLER MACHINE ब्रांड Brand : LIVANOVA ब्रांड प्रकार Brand Type : A कैटलॉग की स्थिति Catalog Status: NA केसे बेचा जा रहा है Selling Status: NA श्रेणी का नाम और चतुर्थांश Category Name & Quadrant : NA (-) मॉडल Model: HEART LUNG MACHINE AND HEATER COOLER MACHINE एचएसएन कोड HSN Code: HSN not specified by seller	1	pieces	24,800,000	NA	24,800,000	
कुल ऑर्डर मूल्य Total Order Value (in INR)						24,800,000	
वैकल्पिक विवरण Consignee Detail							
दिनांक के बाद							

क्र.सं. S.No	प्रेषित Consignee	वस्तु Item	लॉट नंबर Lot No.	मात्रा Quantity	डिलीवरी शुरू करना है Delivery Start After	तक करना है Delivery To Be Completed By
1	पद Designation :- ईमेल आईडी Email ID : con18.neigrhms.ml@gembuyer.in संपर्क Contact : 0364-2538044- जीएसटीआई GSTIN :- पता Address : P.O. NEIGRIHMS, Mawdlangdiang, Shillong, KHAIST HILL, S EAST, MEGHALAYA-793018, India	HEART LUNG MACHINE AND HEATER COOLER MACHINE	-	1	27-May-2025	25-Aug-2025
Product Specification for HEART LUNG MACHINE AND HEATER COOLER MACHINE						
विनिर्देश Specification		उप-विनिर्देश Sub-Spec		मूल्य Value		
Custom Specification		Custom Specification		Yes		
व्यापक रखरखाव के लिए शुल्क Comprehensive maintenance charges for HEART LUNG MACHINE AND HEATER COOLER MACHINE						
सेवा वर्ष Service Year		सीएमसी प्रतिशत CMC Percentage	कर विभाजन Tax Bifurcation			
CMC charges for 1 st year	after warranty period In % of cost of equipment	5%	NA			
CMC charges for 2 nd year	after warranty period In % of cost of equipment	5%				
CMC charges for 3 rd year	after warranty period In % of cost of equipment	5%				
CMC charges for 4 th year	after warranty period In % of cost of equipment	5%				
CMC charges for 5 th year	after warranty period In % of cost of equipment	5%				
CMC charges for 6 year	after warranty period In % of cost of equipment	5%				
CMC charges for 7 year	after warranty period In % of cost of equipment	5%				
CMC charges for 8 year	after warranty period In % of cost of equipment	5%				
विक्रेता विनिर्देश दस्तावेज Seller Specification Document:						
1. Specification Document		mkp.gem.gov.in/catalog_data/catalog_support_document/43/62/765/CatalogAttrs/SpecificationDocument/2024/10/21/2024_10_21_17_36_22_hl_2024-09-14-12-50-03_d13f5a7e_2024-10-21-17-36-23_6da0b4a5fe6504d766850f5900e32e1b.pdf				
खरीदार विनिर्देश दस्तावेज Buyer Specification Document:						
1. Specification Document		mkp.gem.gov.in/catalog_data/catalog_support_document/buyer_documents/52738/54/78/703/CatalogAttrs/SpecificationDocument/2024/9/14/hl_2024-09-14-12-50-03_d13f5a7e838a7845870e0d8a40cc3c6a.pdf				
सुद्धिपत्र Corrigendum						
1. तक बढ़ाया गया Extended Up to : 2024-11-09 11:00:00						
सीएमसी के लिए अतिरिक्त खंड Additional Clauses for CMC						
<p>1.CMC shall include preventive maintenance including calibration as per technical/ service /operational manual of the manufacturer, service charges and spares, after satisfactory completion of Warranty. During the CMC period commencing from date of the successful completion of warranty period, Service personnel shall visit each consignee site as recommended in the manufacturer's technical/ service /operational manual, at least once in six months or as per user requirement. Cost of consumables shall not be included in CMC.Further there will be 98% uptime warranty during CMC period on 24 (hrs) X 7 (days) X 365 (days) basis, with penalty, to extend CMC period by double the downtime period.</p> <p>2.CMC charges to be indicated as percentage of cost of equipment quoted for each year after the warranty period.</p> <p>3.GST shall be included in the CMC Charges quoted.</p> <p>4.Cost of CMC will be added for Ranking/Evaluation purpose with depreciation formula.A 10% discounting rate per year shall be applied on CMC Charges for price evaluation on Net Present Value.</p> <p>5.The payment of CMC will be made on quarterly basis after satisfactory completion of said period, duly certified by end user.</p> <p>6.While creating a bid or after the warranty period, buyers shall indicate whether CMC is required against Yes/No* options. If CMC Charges are included, an option for number of years for CMC required shall be available.Under this option up to 10 years can be chosen for CMC charges beyond warranty period.</p> <p>7.In case the bid has a provision for CMC, the warranty of the product will also be deemed to have been converted into Comprehensive warranty including preventive maintenance as per technical/ service /operational manual of the manufacturer, service charges and spares, during the Warranty Period also. Sellers are therefore advised to include the cost of Comprehensive Warranty including spares (excluding consumables) also in product Cost.</p> <p>8.The CMC functionality shall be available in bid only and no direct RA shall be applicable.In case of bid / R/A decrement rules shall be applicable on total price inclusive of CMC charges. Bunching of products shall not be available while creating bids with CMC charges.</p> <p>8.1.Buyer shall indicate number of years of warranty by selecting different options available in the field depending on warranty parameter applicable in category parameters for the equipment. No. of years of warranty indicated here shall supersede the warranty period indicated elsewhere in bid or product specifications. The Seller while participating in Bid/RA will get fields to indicate CMC charges as percentage depending on number of years of CMC selected by Buyer. The following shall be applicable, if Seller CMC selected:</p> <p>CMC charges for 1st year after warranty period- Percentage to be indicated- A1</p> <p>CMC charges for 2nd year after warranty period- Percentage to be indicated- A2</p> <p>CMC charges for 3rd year after warranty period - Percentage to be indicated- A3</p> <p>CMC charges for 4th year after warranty period - Percentage to be indicated- A4</p>						

Signature

Signature

Signature

Signature

CMC charges for 5th year after warranty period – Percentage to be indicated- A5
 Similarly, A6 to A10 are to be indicated for 6th to 10th year of CMC if applicable.
 8.2. The calculation of CMC Charges shall take into account the number of years of warranty and duration of CMC as specified while creating bid.
 8.3. In the price evaluation, the system shall provide function to calculate the cost of each equipment by formula indicated below including CMC and then show the inter-se-ranking of the bidders. The following are the variables
 (i) Number of years for which CMC required.
 (ii) Number of years of product warranty
 The formula for calculating total cost including CMC charges shall be as under:
 Total Cost for valuation =
 $C + C \times \left\{ \frac{A1}{100} \times (1.10^n) + \frac{A2}{100} \times (1.10^{n+1}) + \frac{A3}{100} \times (1.10^{n+2}) + \frac{A4}{100} \times (1.10^{n+3}) + \frac{A5}{100} \times (1.10^{n+4}) \right\}$ and so on
 C – Cost for equipment quoted and n shall be number of years of product warranty specified.
 If 2 year warranty specified, n shall be 2 and if 5 year warranty specified, n shall be 5. A1, A2, A3, A4 & A5 shall depend on how many years CMC selected. For 3 year CMC, only A1, A2 and A3 factors are to be taken into account and A4 and A5 will not be applicable.
 8.4. CMC charges offered for each subsequent year should be same or higher than preceding year.
 8.5. The CMC charges shall be offered within range of 3 to 10% of cost of equipment.

9. Since CMC charges are to be paid only later for each year during CMC period, applicable performance guarantee amount after placement of contract shall be based on the cost of equipment excluding the cost of CMC Charges.
 10. Performance bank guarantee applicable for CMC is to be submitted at start of the CMC and shall be applicable between 2.5% to 10% as specified in bid on total CMC Charges. The PBG submitted after award of contract shall be released only after new PBG for the CMC period is submitted and accepted by buyer/consignee after due verification. Bank guarantee for CMC is to remain valid till completion of CMC period plus one year. The bank guarantee for CMC shall be submitted to buyer directly. In case, seller fails to submit the PBG or does not provide services for the CMC contract after expiry of warranty period then PBG of equipment shall be forfeited.
 11. In case of splitting of order quantity, equipment cost and CMC charges offered by L1 bidder shall be matched by higher quoting eligible bidders on one-to-one basis. The equipment cost and CMC charges (year to year) shall be matched individually.
 12. The CMC Contract shall be an offline contract to be handled by buyer. The payment of CMC will be made on quarterly basis after satisfactory completion of said period, duly certified by end user and scope of CMC will be as per para 1 above.

ईपीबीजी विवरण | ePBG Detail

सहायक बैंक Advisory Bank :	Bank Of Baroda
ईपीबीजी प्रतिशत (%) ePBG Percentage(%) :	3.00
बोली लगाने वाले को बोली के नियम और शर्तों के अनुसार लागू ईपीबीजी प्रस्तुत करना होगा The bidder shall furnish ePBG as applicable as per bid's terms and conditions	

नियम और शर्तें | Terms and Conditions

1. General Terms and Conditions-

- 1.1 This contract is governed by the General Terms and Conditions, conditions stipulated to this Product/Service as provided in the Marketplace.
 1.2 This Contract between the Seller and the Buyer, is for the supply of the Goods and/ or Services, detailed in the schedule above, in accordance with the General Terms and Conditions (GTC) unless otherwise superseded by Goods / Services specific Special Terms and Conditions (STC) and/ or BID/Reverse Auction Additional Terms and Conditions (ATC), as applicable
 1.3 All GeM Sellers / Service Providers are mandated to ensure compliance with all the applicable laws / acts / rules including but not limited to all Labour Laws such as The Minimum Wages Act, 1948, The Payment of Wages Act, 1936, The Payment of Bonus Act, 1965, The Equal Remuneration Act, 1976, The Payment of Gratuity Act, 1972 etc. Any non-compliance will be treated as breach of contract and Buyer may take suitable actions as per GeM Contract.

2. Buyer Added Bid Specific Terms and Conditions-

- 2.1 Buyer Added Bid Specific ATC:
 Buyer Added text based on ATC clauses

File no:- NEIC R/S&P/C-06/2020-21

Approved in agenda C-5/53rd Procurement Committee

Scope of work & Document details

A	Following mandatory documents must be attached in the bid document as specified, failing which bid will be treated as " non-responsive"
1	Cost of spares, consumables and accessories not covered under warranty and CMC period shall be offered as percentage value of the system/Unit in the Technical Bid Additional Doc1 (Requested in ATC)"
2	Documents with regard to Details compliance statement to be attached At "Additional Doc 2(Requested in ATC)"
	Documents with regard to Original Literature, Pr

[Handwritten signature]

3	Product catalogue, technical datasheet from the firm/O.E.M with Highlighting as per the technical specification must attach At "Additional Doc 3(Requested In ATC)"
4	Documents with regard to list of Offering/Quoted Items mentioning make, model & quantity of each Items must be "Additional Doc 4(Requested In ATC)"
5	Component wise pricing of all equipment/turnkey /electrical/accessories/HVAC etc , must be submitted in the "Financial Document" . Not in technical Bid Any Detail price bid/Component wise pricing should not be attached in the technical bid ,failing which bid will be consider as "Techno Commercially Non Responsive "
B	Warranty / and Maintenance
1	Warranty / for 2 years followed by CMC for 8 years including Spares & service for all the Items supplied In this particular tender including third-party Items and turnkey works .
2	Mandatory 2 PMs / Year with unlimited breakdown calls has to be attended by the Bidder/manufacturer throughout the warranty & CMC period at site.i.e. NEIGRIHMS, SHILLONG.
3	Duly signed Mandatory PM reports has to be submitted periodically, failing which necessary action will be initiated as per term& condition of the tender.
3	E-bidder have to adhere to Government of India, Ministry of Finance, PPD division Public procurement order OM F.No.6/18/2019-PPD dated 23rd July ,2020 inserting Rule 144(XI)in GFR 2017 ,No 1 dated: 23/7 2020 and subsequent Orders No 2 & 3 or as amended from time to time , failing which the bids shall be treated as non-responsive.

Sl. No.	Description of Stores Make	Qty.	Minimum ed Value	Reserve Declared
1.	Heart lung machine 1 set (Make: Srans, Model: -APS-1) with five years comprehensive warranty vide order no HSCC /PUR/NEIGRIHMS/2006/HL mach/33,dt:27/04/2007	1 no.	Rs 10,300,43	
Note: - The price offered in the e-price bid /BOQ shall be considered as <u>inclusive of the buyback value.</u> - The buyback value of the system /assets should <u>also be indicated in the ATC-1 & financial Document as percentage of the offered FOB /Basic Cost</u> of the system /asset: in order to arrive at Book /Reserve value for the record of the Institute. - The buyback value should essentially cover the Minimum Reserve Declared Value. - The Institute reserve the <u>right to trade or not to trade</u> the old existing machines offered while purchasing a new one.				

[Signature]
22/11/20

Buyer Added Bid Specific Terms and Conditions

1. Generic

End User Certificate: Wherever Bidders are insisting for End User Certificate from the Buyer, same shall be provided in Buyer's standard format only.

2. Generic

Experience Criteria: The Bidder or its OEM (themselves or through reseller(s)) should have regularly, manufactured and supplied same or similar Category Products to any Central / State Govt Organization / PSU / Public Listed Company for 3 years before the bid opening date. Copies of relevant contracts to be submitted along with bid in support of having supplied some quantity during each of the year. In case of bunch bids, the primary product having highest value should meet this criterion.

3. Generic

IT equipment shall be IPv6 ready from day one.

4. Generic

Installation, Commissioning, Testing, Configuration, Training (As applicable as per scope of supply) is to be carried out by OEM / OEM Certified resource or OEM authorized Reseller.

5. Generic

Upload Manufacturer authorization: Wherever Authorized Distributors are submitting the bid, Manufacturers Authorization Form (MAF)/Certificate with OEM details such as name, designation, address, e-mail id and Phone No. required to be furnished along with the bid.

6. Generic

The successful bidder has to supply all essential accessories required for the successful installation and commissioning of the goods supplied. Besides standard accessories as per normal industry practice, following accessories must be part of supply and cost should be included in bid price: All the items and accessories as per Technical Specification.

7. Generic

The Buyer has an existing set up / inventory of similar products. The offered / supplied product must be compatible with existing system. The bidder has to ensure Compatibility of the supplied items or shall have to include in the supply the necessary hardware / software to make them compatible at no extra cost to the buyer. The details of items with which compatibility is required are as under: all the spares including UPS, P C, battery, Printer, Probes & upgradation of System Software & third-party Software

8. Scope of supply

Scope of supply (Bid price to include all cost components) : Supply Installation Testing Commissioning of Goods, Training of operators and providing Statutory Clearances required (If any)

9. Turnover

Handwritten signatures and initials:
- A signature that appears to be "John D." with a checkmark above it.
- A signature that appears to be "Rajendra".
- A signature that appears to be "S. K. Srinivasan" with the date "28/10/2023" written below it.

Bidder Turn Over Criteria: The minimum average annual financial turnover of the bidder during the last three years, ending on 31st March of the previous financial year, should be as indicated in the bid document. Documentary evidence in the form of certified Audited Balance Sheets of relevant periods or a certificate from the Chartered Accountant / Cost Accountant indicating the turnover details for the relevant period shall be uploaded with the bid. In case the date of constitution / incorporation of the bidder is less than 3 year old, the average turnover in respect of the completed financial years after the date of constitution shall be taken into account for this criteria.

10. Turnover

OEM Turn Over Criteria: The minimum average annual financial turnover of the OEM of the offered product during the last three years, ending on 31st March of the previous financial year, should be as indicated in the bid document. Documentary evidence in the form of certified Audited Balance Sheets of relevant periods or a certificate from the Chartered Accountant / Cost Accountant indicating the turnover details for the relevant period shall be uploaded with the bid. In case the date of constitution / incorporation of the OEM is less than 3 year old, the average turnover in respect of the completed financial years after the date of constitution shall be taken into account for this criteria. In case of bunch bids, the OEM of CATEGORY RELATED TO primary product having highest bid value should meet this criterion.

11. OEM

IMPORTED PRODUCTS: In case of imported products, OEM or Authorized Seller of OEM should have a registered office in India to provide after sales service support in India. The certificate to this effect should be submitted.

12. Purchase Preference (Centre)

As per DPIIT notification at the time of e-tender, bidding or solicitation the bids shall be required to indicate percentage of local content and provide self-certification (by Director / Company Secretary) and also give details of the location/s at which value addition is made". Since the bidder here is not the local supplier, the same was required to be obtained from the "Class-I local supplier /Class II local supplier"

Further the details of Calculations of local content areas under:

Question 1. How to calculate Local Content?

Answer Para 2 of the PPP-MII Order, 2017 (as amended on 16.09.2020) defines local content as

Local content' means the amount of value added in India which shall, unless otherwise prescribed by the Nodal Ministry, be the total value of the item procured (excluding net domestic indirect taxes) minus the value of imported content in the item (including all customs duties) as a proportion of the total value, in percent.

Mathematically,

Local content = (Sale price - Value of Imported content) * 100/ Sale price

Where, 'Sale price' means price excluding net domestic indirect taxes and "Value of Imported content" means price of imported content inclusive of all customs duties

Question 2. How to calculate Local Content in bids involving supply of multiple items from single bidder?

Answer In case of bids requiring supply of multiple items (say "X1", "X2" and "X3") by a single bidder, the local content in the bid shall be

Local content = ((Sale price of "X1" - Value of imported content in "X1") + (Sale price of "X2" - Value of imported content in "X2") + (Sale price of "X3" - Value of imported content in "X3")) * 100/ (Sale price of "X1" + Sale price of "X2" + Sale price of "X3")

27/08/20

Approved

27/08/20

13. Service & Support

Availability of Service Centres: Bidder/OEM must have a Functional Service Centre in the State of each Consignee's Location in case of carry-in warranty. (Not applicable in case of goods having on-site warranty). If service center is not already there at the time of bidding, successful bidder / OEM shall have to establish one within 30 days of award of contract. Payment shall be released only after submission of documentary evidence of having Functional Service Centre.

14. Service & Support

Dedicated /toll Free Telephone No. for Service Support : BIDDER/OEM must have Dedicated/toll Free Telephone No. for Service Support.

15. Service & Support

Escalation Matrix For Service Support : Bidder/OEM must provide Escalation Matrix of Telephone Numbers for Service Support.

16. Certificates

Bidder's offer is liable to be rejected if they don't upload any of the certificates / documents sought in the Bid document, ATC and Corrigendum if any.

17. Certificates

The bidder or the OEM of the offered products must have BIS/WHO-GMP/ CDSCO Indian certification or alternate certification as recognized by Government of India

18. Certificates

Material Test Certificate Should Be Sent Along with The Supply. The Material Will Be Checked by Buyer's Lab & the Results of the Lab will be the Sole Criteria for Acceptance of the Item.

19. Certificates

The bidder is required to upload, along with the bid, all relevant certificates such as BIS license, type test certificate, approval certificates and other certificates as prescribed in the Product Specification given in the bid document.

20. Certificates

To be eligible for award of contract, Bidder / OEM must possess following Certificates / Test Reports on the date of bid opening (to be uploaded with bid): All the quality & electrical safety certificates .

21. Warranty

Bidder / OEM has to give an undertaking that after expiry of warranty period, it will provide Comprehensive Maintenance Service for next 8 years for the offered products at the rate not more than 5% of contract price per annum. Buyer reserves the right to enter into a CMC agreement with the Successful Bidder / OEM after expiry of the Warranty period at above mentioned rate and the payment for the CMC charges would be made Biannually after rendering of the CMC Services of the relevant CMC period. Performance Security of the successful bidder shall be forfeited if it fails to accept the CMC contract when called upon by the buyer. CMC would include cost of all the spares including UPS, PC, battery, Printer, Probes & upgradation of Systems

AB

Hein

Bened

Walter

software & third party Software (Upload the undertaking). The original Performance Security of contract will be returned only after submission and verification of AMC Performance Security for 5% of total CMC value valid up to CMC period plus 2 months (if there is no other claim).

22. Warranty

Warranty period of the supplied products shall be 2 years from the date of final acceptance of goods or after completion of installation, commissioning & testing of goods (if included in the scope of supply), at consignee location. OEM Warranty certificates must be submitted by Successful Bidder at the time of delivery of Goods. The seller should guarantee the rectification of goods in case of any break down during the guarantee period. Seller should have well established Installation, Commissioning, Training, Troubleshooting and Maintenance Service group in INDIA for attending the after sales service. Details of Service Centres near consignee destinations are to be uploaded along with the bid.

23. Warranty

Over and above the normal Warranty terms as per GeM GTC, the successful bidder / OEM shall have to provide Comprehensive Warranty during the entire Standard warranty period as per contract. The comprehensive warranty shall be covering the following scope all the spares including UPS, PC, battery, Printer, Probes & upgradation of System Software & third party Software (Upload an undertaking with the bid confirming compliance by the bidder if Bidder is taking onus of this compliance. In case OEM is taking onus of this compliance, OEM undertaking is to be uploaded along with Bidder undertaking)

24. Warranty

Successful bidder will have to ensure that adequate number of dedicated technical service personals / engineers are designated / deployed for attending to the Service Request in a time bound manner and for ensuring Timely Servicing / rectification of defects during warranty period, as per Service level agreement indicated in the relevant clause of the bid.

25. Warranty


Timely servicing / rectification of defects during warranty period: After having been notified of the defects / service requirement during warranty period, Seller has to complete the required Service / Rectification within 3 days' time limit. If the Seller fails to complete service / rectification with defined time limit, a penalty of 0.5% of Unit Price of the product shall be charged as penalty for each week of delay from the seller. Seller can deposit the penalty with the Buyer directly else the Buyer shall have a right to recover all such penalty amount from the Performance Security (PBG). Cumulative Penalty cannot exceed more than 10% of the total contract value after which the Buyer shall have the right to get the service / rectification done from alternate sources at the risk and cost of the Seller besides forfeiture of PBG. Seller shall be liable to reimburse the cost of such service / rectification to the Buyer.

26. Past Project Experience

For fulfilling the experience criteria any one of the following documents may be considered as valid proof for meeting the experience criteria:

- a. Purchase Order copy along with Invoice(s) with self-certification by the bidder that supplies against the invoices have been executed.
- b. Execution certificate by client with order value.
- c. Any other document in support of order execution like Third Party Inspection release note, etc.

27. Forms of EMD and PBG



Bidder can also submit the EMD with Account Payee Demand Draft in favour of NEIGRIHMS EMD SECURITY DEPOSITS payable at MAWDIANGDIANG, SHILLONG-793018, MEGHALAYA. Bidder has to upload scanned copy / proof of the DD along with bid and has to ensure delivery of hardcopy to the Buyer within Bid End date & time / Bid Opening date & time.

28. Forms of EMD and PBG

Bidder can also submit the EMD with Fixed Deposit Receipt made out or pledged in the name of A/C (Name of the Buyer). The bank should certify on it that the deposit can be withdrawn only on the demand or with the sanction of the pledgee. For release of EMD, the FDR will be released in the favour of the bidder by the Buyer after making endorsement on the back of the FDR duly signed and stamped along with covering letter. Bidder has to upload scanned copy/ proof of the FDR along with bid and has to ensure delivery of hardcopy to the Buyer within Bid End date & time / Bid Opening date & time.

29. Forms of EMD and PBG

Bidders can also submit the EMD with Banker's Cheque in favour of NEIGRIHMS EMD SECURITY DEPOSITS payable at MAWDIANGDIANG, SHILLONG-793018, MEGHALAYA. Bidder has to upload scanned copy / proof of the BC along with bid and has to ensure delivery of hardcopy to the Buyer within Bid End date & time / Bid Opening date & time.

30. Forms of EMD and PBG

Bidders can also submit the EMD with Payment online through RTGS / Internet banking in Beneficiary name NEIGRIHMS EMD SECURITY DEPOSITS Account No. 302702 0000027 IFSC Code BARBOMAWDIA Bank Name BANK OF BARODA Branch address MAWDIANGDIANG, SHILLONG-793018, MEGHALAYA. Bidder to indicate bid number and name of bidding entity in the transaction details field at the time of online transfer. Bidder has to upload scanned copy / proof of the Online Payment Transfer along with bid.

31. Forms of EMD and PBG

Successful Bidder can submit the Performance Security in the form of Account Payee Demand Draft also (besides PBG which is allowed as per GeM GTC). DD should be made in favour of NEIGRIHMS EMD SECURITY DEPOSITS payable at MAWDIANGDIANG, SHILLONG-793018, MEGHALAYA. After award of contract, Successful Bidder can upload scanned copy of the DD in place of PBG and has to ensure delivery of hard copy to the original DD to the Buyer within 15 days of award of contract.

32. Forms of EMD and PBG

Successful Bidder can submit the Performance Security in the form of Fixed Deposit Receipt also (besides PBG which is allowed as per GeM GTC). FDR should be made out or pledged in the name of NEIGRIHMS EMD SECURITY DEPOSITS A/C (Name of the Seller). The bank should certify on it that the deposit can be withdrawn only on the demand or with the sanction of the pledgee. For release of Security Deposit, the FDR will be released in favour of bidder by the Buyer after making endorsement on the back of the FDR duly signed and stamped along with covering letter. Successful Bidder has to upload scanned copy of the FDR document in place of PBG and has to ensure delivery of hard copy of Original FDR to the Buyer within 15 days of award of contract.

33. Forms of EMD and PBG

Successful Bidder can submit the Performance Security in the form of Payment online through RTGS / Internet banking also (besides PBG which is allowed as per GeM GTC).

[Handwritten signatures and initials are present at the bottom of the page, including a large signature on the right and several initials on the left.]

On-line payment shall be in Beneficiary name NEIGRIHMS EMD SECURITY DEPOSITS Account No. 30270200000027 IFSC Code BARB0MAWDIA Bank Name BANK OF BARODA Branch Address MAWDIANGDIANG, SHILLONG-793018, MEGHALAYA. Successful Bidder to indicate Contract number and name of Seller entity in the transaction details field at the time of on-line transfer. Bidder has to upload scanned copy / proof of the Online Payment Transfer in place of PBG within 15 days of award of contract.

29. RELATIONSHIP CERTIFICATE In Bidder, a letter Head with detail Declaration must be submitted in the following format "It is certified that I/We, the undersigned, do (With Detail Name & details) /do not have relationship with any of the employees working at NEIGRIHMS. The above statement is true and is submitted against the Gem Tender Enquiry _____ Dated _____,

Date : _____

(Signature) Name of the Company/Firm Seal

30. In case of need of fulfilment of statutory requirement for receipt/Installation /Operation of stores /system such as AERB clearance /approval, PC-PNDT, Clearance from fire department, environmental /Site clearance etc, the delivery/Installation period shall commence from the date of obtaining such clearance.

31. In order to ensure provision of services (cmc), spares, consumables, reagents for the quoted system as per condition bidding and to ensure compliance as per the provisions of the Contract Acts as amended from time to time, a triparted agreement is required to be concluded in prior to Final Acceptance of the store/System.

(C) 3. Additional Terms and conditions & Scope of Work for CMC

Tenderer/ Vendors/contractor should note that the following terms and conditions will apply specifically in addition to the Rules and the Regulation as applicable to such provide services in the Government of India.

1. Comprehensive Annual Maintenance Contract must include Labour, spares & Preventive Maintenance of all the excluding of battery, Accessories/Consumables
2. The terms and conditions of the tender and the agreement executed will be binding on the vendor/contractor. This offer is being issued in accordance with the terms & conditions of NEIGRIHMS /Government of India and in the manner specified herein shall operate to create a specific contract between the vendor/contractor (with whom the contract referred to) on one part and NEIGRIHMS, Shillong, on the other part.
3. The required spares to be replace must be genuine and certified from the OEM.
4. Repairs to be undertaken should be within specified configuration and maintaining the integration on internal circuit of equipment, any deviation on configuration/ specification the repair will not be acceptable. After repairs, a certificate to the effect that the equipment is in working order and safe for patient care and non-hazardous for the handler's shall be submitted by the CMC holder.
5. Tenderer/Vendors/contractor is responsible to provide electrical and patient safety certificate after major repair of equipment which are used for direct patient care.
6. The system must be checked & calibrated after every spare changes and detail service report must be submitted to the user & BME.
7. 2 nos. of Periodic preventive maintenance is mandatory irrespective of unlimited service / breakdown calls.
8. Same Similar Standby system must be provided by the bidder if the system needs to send to workshop for any major repair.
9. Receipt of this offer may be acknowledged and a copy duly signed/stamped by the authorized signatory should be submitted before finalization of the agreement.
10. The Performance security shall be denominated in any one of the forms namely Account Payee Demand Draft or Fixed Deposit Receipt drawn from any nationalized bank in India or Bank Guarantee issued by a nationalized bank in India, pledge in favor of Deputy Director, NEIGRIHMS, and Shillong-793018 for an amount equivalent to 3% of the total cost of annual CMC. The validity of the Fixed Deposit receipt or Bank Guarantee will be upto 2 months beyond CMC period.
11. It may also be noted that there should be no negligence in providing services of any type, if any, complaint is received the contract will be terminated with immediate effect.
12. There will be 98% uptime warranty during CMC period on 24 (hrs) X 7 (days) X 365 (days) basis, with penalty, to extend CMC period by double the downtime period. The vendor shall ensure optimum uptime of the system during CMC period, failing which the initiate shall initiate stern action, as deemed fit.
13. During Comprehensive Maintenance Contract period, the supplier shall visit at each consignee's site for preventive maintenance including testing and calibration as per the manufacturer's service/ technical/ operational manual. The supplier shall visit each consignee site as recommended in the manufacturer's manual, but at least once in 6 months commencing from the date of the successful completion of warranty period for preventive maintenance of the goods.

Mr

Attn: Dr.

Beneal

refers


14. Processing of bill may be considered on yearly basis with satisfactory report from the user department. The AMC/CMC bills should be certified by the concerned Head of the Department/ In- Charge , BME and the respective DMS/MS.
15. Software updates should be provided free of cost during CMC. The first service call by the team of service engineers should be within 7 days of issue of this order.
16. Settlement of disputes - Director, NEIGRIHMS or his authorized representative shall be the final authority in all disputes and decision will be binding on all concerned.
17. All other terms & conditions are as per award of contract mentioned in pre-page.
18. Bidders are required to sign the CMC contract agreement within 15 (fifteen) days from the issue of the letter of award/supply order, failing which EMD/security deposit may be forfeited or Contract declared null and void.

पूजो / उपभोग्य सामग्रियों के लिए प्रस्तावित मूल्य | Price Offered for Spares / Consumables:

पूजो / उपभोग्य सामग्रियों के दस्तावेज लिंक के लिए प्रस्तावित मूल्य | Price Offered for Spares / Consumables Document link

नोट: यह सिस्टम जनरेट फाइल है। कोई हस्ताक्षर की आवश्यकता नहीं है। इस दस्तावेज का प्रिंट आउट भुगतान/लेनदेन उद्देश्य के लिए मान्य नहीं है।

Note: This is system generated file. No signature is required. Print out of this document is not valid for payment/ transaction purpose.



Technical Specifications Heart Lung Machine (Cardio Pulmonary Bypass Machine) and Accessories

Part A.

S. No	Heart Lung Machine
1.	The Machine should switch-on within 15 seconds. The unit should have at least 5 large pumps. These pumps can be used as an arterial, suction, vent and cardioplegia with separate power supply and control modules.
2.	Should have easy access connectors for interchanging the pump.
3.	The pumps should show accurate measurements and the machine should be capable of being used for performing all types of cardiac surgeries.
4.	Pumps should have direct drive system or belt drive system. Each Module should work on its own.
5.	Machine should show all warnings and alarms in a complete text message and have audible tones.
6.	Machine should be able to configure all thresholds, alarms, settings, and pump assignments in up to 12 different user profiles
7.	Should have a spill proof base.
8.	All the pumps should have individually controlled knobs allowing fine tuning and coarse adjustment of flows.
9.	All the pump heads should be rotatable to optimise the tubing length and contribute towards minimizing haemodilution.
10.	The unit should be supplied with a battery backup of minimum of 60 minutes. Switch over from main power to battery backup should be automatic and immediate. The battery unit should be built into the pump base, and it should immediately and automatically recharge with battery test information and time remaining feature. The machine should be able to start on battery.
11.	Should have hand crank facility as a safety feature. Pump should give audible sound if rotated in wrong set direction or should be unidirectional.
12.	The display monitor should have an option to show up minimum of 2 timers.
13.	Pressure sensor and level sensor should have 2 modes – Stop Mode & Control Mode/alert mode.
14.	The machine should be supplied with following minimum safety modules: i. One Pressure module with 2 pressure monitoring outlets. ii. One Bubble module with ability to connect to sensors. iii. One Flow Sensor to get the measured flow (should be optional) iv. One Level sensor. v. One Temperature module with ability to measure up to two different temperatures.
15.	The machine should have ability to deliver cardioplegia.
16.	The machine should be provided with mechanical gas blender.

Handwritten signature/initials

Handwritten signature/initials

Copy Babel

Handwritten signature/initials

Handwritten signature/initials

Handwritten signature/initials

7.	Optional feature: Inline Blood Parameter Monitoring to measure Venous and Arterial parameters with possibility to control from the main display of machine.
8.	Optional feature - Patient monitor with the machine which continuously displays and stores all parameters from Heart Lung Machine and can support up to 6 External Devices available within the operating theatre like anaesthesia monitor, NIRS monitors, blood gas analysers and ACT devices.
9.	Optional feature -Patient Monitor with an option to connect through HL7 to Hospital Electronic Medical Records.

Part B

S. N	Heater Cooler Machine
1	The unit shall be capable of supplying water of temperature 3 - 41 degree Celsius for patient circuit, 2-10 degree Celsius for Cold cardioplegia and 15-41 degree Celsius for warm cardioplegia circuit.
2.	The unit should have 3 independent tanks and 2 separate circuits, and these circuits should be able to control patient's temperature and also heating and cooling of cardioplegia and should work simultaneously. It should have an easy de-airing process and an easy drainage system.
3.	The accuracy should be at least 0.1 degree Celsius. Settings should be adjustable to approx. 0.1 degree Celsius.
4.	The heater cooler unit should also be compatible to get integrated into the heart lung machine.
5.	Warranty: Standard 5-years warranty plus 5 year CMC

UJ 742

1974

Any Babel

15/10/2024

8

AV

Heater cooler
already
fitted

TECHNICAL SPECIFICATIONS OF HEART LUNG MACHINE AND HEATER COOLER MACHINE-

A. HEART LUNG MACHINE

1. DESCRIPTION :

Heart Lung Machine is an apparatus through which blood is temporarily diverted, during heart surgery, to oxygenate it and pump it throughout the body, thus maintaining circulation until the heart and lungs are able to return to normal functioning.

2. TECHNICAL SPECIFICATIONS:

1. The system should work on 220V/50Hz, single phase AC.
2. The unit should be supplied with a Battery backup of minimum of 90 minutes. Switch over from main power to battery backup should be automatic and immediate. The battery unit should be built in to the pump base and it should be recharged automatically and immediately.
3. Two unidirectional hand cranks should be available as an additional critical safety feature.
4. The unit should have 3 single pumps and 1 twin pump (5 console pump) can be used as arterial, suction, vent and cardioplegia with separate power supply and control modules. Should have easy access connectors for interchanging the pump. The pumps should show accurate measurements and the machine should be used for performing paediatric and neonate surgery. The small pump should be able to incorporate tubing sizes 1/4" and 3/8".
5. The design of pump must be horse shoe race way design and the pumps should have direct drive system and maintenance free. All the Pumps should have pulsatile mode option built. Each module should work its own.
6. The base should provide at least 5 dedicated connection for the pumps, which can be mounted on the base or on the pole close to the surgical field to minimize tubing lengths and also the floor space. There should be option for selection of desired number of pumps. (Dual pump will be counted as only one pump). The base should be spill proof.
7. Each head should be controlled individually and rotatable in different direction with master – slave control.
8. Pump should display tube size, RPM rotation directions, speed flow rate, safety, status and error message. Machine should show all warnings and alarms in complete text message and different audible tones with different colour codings.
9. Facility of monitoring all patient parameters; pumps status, patient information entered or system setting should be available on the touch screen of the central control monitor. One touch should provide any menu on the screen.
10. The screen should automatically change if any alarm condition arises thus alerting the perfusionist to take immediate medical action. Important patient parameters should remain displayed.
11. Accuracy: Occlusion accuracy should be – 0.015mm, Occlusion rollers accuracy should be adequate.

Saimed Innovation

ISO 9001:2015

156A, Lenin Nagarani, # 412, Kamalalaya Centre, Kolkata - 700 013, India

Tel. : +91 33 1062 7462, Fax : +91 33 2215 6067, E-mail : acct@saimed.in / info@saimed.in

Website : www.saimed.in





12. Occlusion: Should have Thumb wheel locking Mechanism
13. Monitor: Pressure monitor (2 nos), Timers(3 nos), Temperature Monitor(4 nos) and all the monitors should be touch screen.
14. The central control monitor should be able to configure 10-14 different perfusion screens for different equipment setups and 1 perfusion protocol setups.
15. Pressure Sensor should have 2 modes – Stop Mode & Control Mode.
16. Cardioplegia module should have both Manual as well as Automatic operation.
17. Facility for setting Cardioplegia volume, dose delivery and time on the Central Control Monitor should be available
18. There should be option for master follower operation for multidose cardioplegia.
19. Should be provided with mechanical gas blender.
20. Should be provided with Level Sensor and air Bubble sensor/detector.
21. Bubble sensor should have different bubble detection thresholds including even micro-bubble detection function.
22. Level sensor should be with 2 modes - Normal & Control Mode.
23. Must have inbuilt Master UPS - shows all the details like Battery time, Load time & Remaining time. Should have BSA Calculation

3. ACCESSORIES

1. Level Sensor strips (100 nos)
2. Pressure transducer (100 nos)
3. Pressure transducer cable (4 nos)
4. Each unit should be supplied with standard accessories, 2 skin temperature probes and 2 air temperature probes
5. Stainless steel line clamps x 12 nos.
6. Stainless steel intracardiac suckers (2 Adult, 2 Pediatric)

B. HEATER COOLER MACHINE

1. DESCRIPTION:

The heater cooler machine is an essential accompaniment to the heart lung machine for cooling / heating up the core body temperature of the body which is required in different phases of cardiac surgery

2. TECHNICAL SPECIFICATIONS:

1. The unit should be capable of operating continuously in ambient temperature of 2 - 40.5 degree Celsius.
2. The unit should have 3 independent tanks and 3 separate circuits and these circuits should be able to control patient's temperature and also heating and cooling of cardioplegia and should work simultaneously.
3. The accuracy should be 0.1°C. Settings should be adjustable to 0.1°C

Saimed Innovation

ISO 9001:2015

156A, Lenin Sarani, # 412, Kamalajoy Centre, Kolkata - 700 013, India


Tel. : +91 33 262 7462, Fax : +91 33 2215 6067, E-mail : acc@saimeed.in / info@saimeed.in

Website : www.saimeed.in





4. The heater cooler should have automatic De-airing facility.
5. The heater cooler unit should also be compatible to get integrated into the heart lung machine.
6. 2 Adult and 2 Pediatric heater cooler water blankets should be supplied with the heater cooler machine.



*As per
celebrity
document
here
27/11/11*