

DEPARTMENT OF OTORHINOLARYNGOLOGY AND
HEAD & NECK SURGERY

No. NEIGR/ ENT/ 2025/151

Dated: Shillong, the 24th September 2025

To,

The Store & Procurement Officer,
NEIGRIHMS, Shillong.

Sub: Amendment in Technical Specification of VNG- ENT.

Sir,

With reference to the subject cited above, submitting herewith the amendment in the Technical Specification for VNG for the department of ENT.

Sl.No.	Tendered Technical Specification	May be read as
A	Software	
	Integrates with VNG, SUV, CCG,OVA, stabilometry	Integrates with VNG and VHIT software with provision for up gradation with SUV, CCG,OVA, Stabilometry (prices to be quoted separately)
	Static SVV, head vertical & 30 degree inclined, dynamic moving background test	Head vertical & 30 degree inclined, dynamic moving background test
B	Motorized Chair and Controller	Rotary Chair (Motorized Chair)
	Linear servo amplifier control	Linear servo amplifier control or equivalent
	Coin silver slip rings	Coin silver slip rings or equivalent
F.i	VHIT Multi with goggles (separate from VNG goggles)	VHIT Multi with / without goggles .
K	Pediatric (Optional)	Deleted
L	Computerized Stabilometry (posturography)	Deleted
M	CranioCorpography (CCG)	Deleted

Thanking you,

Yours sincerely,

Prof. Suvamoy Chakraborty, MS, MAMS, FICS, FIMSA
Professor & Head, Department of ENT
NEIGRIHMS

*This is discussed with
other faculty and audiologists
in the dept and a
cumulative decision
taken by all.*

[Signature]
24/09/2025
Prof. Suvamoy Chakraborty
Professor & HOD
प्रध्यापक व विभागाध्यक्ष/Professor & HOD
ईएनटी विभाग/Department of ENT
नेग्रिहम्स, शिलांग 18/NEIGRIHMS, SHILLONG