

बिड दस्तावेज़ / Bid Document

बिड विवरण/Bid Details	
बिड बंद होने की तारीख/समय /Bid End Date/Time	07-03-2026 16:00:00
बिड खुलने की तारीख/समय /Bid Opening Date/Time	07-03-2026 16:30:00
बिड पेशकश वैधता (बंद होने की तारीख से)/Bid Offer Validity (From End Date)	180 (Days)
मंत्रालय/राज्य का नाम/Ministry/State Name	Ministry Of Health And Family Welfare
विभाग का नाम/Department Name	Department Of Health And Family Welfare
संगठन का नाम/Organisation Name	North Eastern Indira Gandhi Regional Institute Of Health And Medical Sciences (neigrihms)
कार्यालय का नाम/Office Name	Neigrihms, Shillong
कुल मात्रा/Total Quantity	3
वस्तु श्रेणी /Item Category	Microneurography (MSNA, SSNA) , SSR Add-On Powerlab And Labchart Software , Portable Hemodynamic Recorder for Cardiovascular Lab(Portable Continous Non Invasive Haemodynamic M , Functional Near-Infrared Spectroscopy (FNIRS) System For Human Brain Imaging
GeMARPTS में खोजी गई स्ट्रिंग्स / Searched Strings used in GeMARPTS	Functional Near-Infrared Spectroscopy (FNIRS) System for Human Brain Imaging
GeMARPTS में खोजा गया परिणाम / Searched Result generated in GeMARPTS	Category not available on GeM for the text string uploaded by the buyer
अधिसूचना के लिए चयनित प्रासंगिक श्रेणियाँ / Relevant Categories selected for notification	Infrared Illuminator
बिडर का न्यूनतम औसत वार्षिक टर्नओवर (3 वर्षों का) /Minimum Average Annual Turnover of the bidder (For 3 Years)	31 Lakh (s)
मूल उपकरण निर्माता का औसत टर्नओवर (गत 3 वर्षों का)/OEM Average Turnover (Last 3 Years)	31 Lakh (s)
उन्हीं/समान सेवा के लिए अपेक्षित विगत अनुभव के वर्ष/Years of Past Experience Required for same/similar service	3 Year (s)
टर्नओवर के लिए एमएसई को छूट प्राप्त है / MSE Relaxation for Turnover	Yes Partial Turn over value - 23 (in lakhs)
टर्नओवर के लिए स्टार्टअप को छूट प्राप्त है / Startup Relaxation for Turnover	Yes Partial Turn over value - 23 (in lakhs)

बिड विवरण/Bid Details	
विक्रेता से मांगे गए दस्तावेज़/Document required from seller	Experience Criteria,Past Performance,Bidder Turnover,Certificate (Requested in ATC),OEM Authorization Certificate,OEM Annual Turnover,Additional Doc 1 (Requested in ATC),Additional Doc 2 (Requested in ATC),Additional Doc 3 (Requested in ATC),Additional Doc 4 (Requested in ATC),Compliance of BoQ specification and supporting document *In case any bidder is seeking exemption from Experience / Turnover Criteria, the supporting documents to prove his eligibility for exemption must be uploaded for evaluation by the buyer
क्या आप निविदाकारों द्वारा अपलोड किए गए दस्तावेज़ों को निविदा में भाग लेने वाले सभी निविदाकारों को दिखाना चाहते हैं? संदर्भ मेनू है/Do you want to show documents uploaded by bidders to all bidders participated in bid?	Yes (Documents submitted as part of a clarification or representation during the tender/bid process will also be displayed to other participated bidders after log in)
बिड लगाने की समय सीमा स्वतः नहीं बढ़ाने के लिए आवश्यक बिड की संख्या। / Minimum number of bids required to disable automatic bid extension	2
दिनों की संख्या, जिनके लिए बिड लगाने की समय-सीमा बढ़ाई जाएगी। / Number of days for which Bid would be auto-extended	10
ऑटो एक्सटेंशन अधिकतम कितनी बार किया जाना है। / Number of Auto Extension count	1
विगत प्रदर्शन /Past Performance	30 %
बिड से रिवर्स नीलामी सक्रिय किया/Bid to RA enabled	No
व्यापक रखरखाव शुल्क आवश्यक / Comprehensive Maintenance Charges Required	Yes
बिड का प्रकार/Type of Bid	Two Packet Bid
प्राथमिक उत्पाद श्रेणी/Primary product category	Functional Near-Infrared Spectroscopy (FNIRS) System For Human Brain Imaging
तकनीकी मूल्यांकन के दौरान तकनीकी स्पष्टीकरण हेतु अनुमत समय /Time allowed for Technical Clarifications during technical evaluation	7 Days
निरीक्षण आवश्यक (सूचीबद्ध निरीक्षण प्राधिकरण /जेम के साथ पूर्व पंजीकृत एजेंसियों द्वारा)/Inspection Required (By Empanelled Inspection Authority / Agencies pre-registered with GeM)	No
अनुमानित बिड मूल्य /Estimated Bid Value	10400000
मूल्यांकन पद्धति/Evaluation Method	Item wise evaluation
वित्तीय दस्तावेज की आवश्यकता है / Financial Document Required	Yes
मध्यस्थता खंड/Arbitration Clause	No

बिड विवरण/Bid Details	
सुलह खंड/Mediation Clause	No

ईएमडी विवरण/EMD Detail

एडवाइजरी बैंक/Advisory Bank	Bank Of Baroda
Schedule 1 ईएमडी राशि/EMD Amount (In INR)	24000
Schedule 2 ईएमडी राशि/EMD Amount (In INR)	112000
Schedule 3 ईएमडी राशि/EMD Amount (In INR)	72000

ईपीबीजी विवरण /ePBG Detail

एडवाइजरी बैंक/Advisory Bank	Bank Of Baroda
ईपीबीजी प्रतिशत (%) /ePBG Percentage(%)	3.00
ईपीबीजी की आवश्यक अवधि (माह) /Duration of ePBG required (Months).	26

(a). जेम की शर्तों के अनुसार ईएमडी छूट के इच्छुक बिडर को संबंधित कैटेगरी के लिए बिड के साथ वैध समर्थित दस्तावेज प्रस्तुत करने है। एमएसई कैटेगरी के अंतर्गत केवल वस्तुओं के लिए विनिर्माता तथा सेवाओं के लिए सेवा प्रदाता ईएमडी से छूट के पात्र हैं। व्यापारियों को इस नीति के दायरे से बाहर रखा गया है।/EMD EXEMPTION: The bidder seeking EMD exemption, must submit the valid supporting document for the relevant category as per GeM GTC with the bid. Under MSE category, only manufacturers for goods and Service Providers for Services are eligible for exemption from EMD. Traders are excluded from the purview of this Policy.

(b). The EMD Amount will be applicable for each schedule/group selected during Bid creation.

(c). ईएमडी और संपादन जमानत राशि, जहां यह लागू होती है, लाभार्थी के पक्ष में होनी चाहिए। / EMD & Performance security should be in favour of Beneficiary, wherever it is applicable.

लाभार्थी /Beneficiary :

Stores and Procurment Officer
Neigrihms, Shillong, Department of Health and Family Welfare, North Eastern Indira Gandhi Regional Institute of Health and Medical Sciences (NEIGRIHMS), Ministry of Health and Family Welfare
(Harendra Dey)

बोली विभाजन लागू नहीं किया गया/ Bid splitting not applied.

एमआईआई खरीद वरीयता / MII Purchase Preference

एमआईआई खरीद वरीयता / MII Purchase Preference	Yes
मेक इन इंडिया विक्रेताओं को खरीद में प्राथमिकता, यदि उनका मूल्य $L1 + X\%$ तक की सीमा में है / Purchase Preference to MII sellers available upto price within $L1 + X\%$	20

मेक इन इंडिया खरीद में प्राथमिकता के लिए बिड की मात्रा का अधिकतम प्रतिशत / Maximum Percentage of Bid quantity for MII purchase preference	50
सार्वजनिक खरीद (मेक-इन-इंडिया को प्राथमिकता) आदेश 2017 के अनुसार केवल क्लास 1/क्लास 2 के स्थानीय आपूर्तिकर्ताओं को ही भागीदारी की अनुमति है दिनांक 16.09.2020 (समय-समय पर संशोधित एवं लागू) / Allow participation only from Class 1/Class 2 local suppliers as per the Public procurement(Preference to Make-in-india) order 2017 date 16.09.2020(as amended and applicable time to time)	Yes, in compliance with the MII ORDER : DPIIT Order(as amended and applicable time to time)

एमएसई खरीद वरीयता/MSE Purchase Preference

एमएसई खरीद वरीयता/MSE Purchase Preference	Yes
सूक्ष्म और लघु उद्यम मूल उपकरण निर्माताओं को खरीद में प्राथमिकता, यदि उनका मूल्य $L1+X\%$ तक की सीमा में हो / Purchase Preference to MSE OEMs available upto price within $L1+X\%$	15
सूक्ष्म और लघु उद्यम को खरीद में प्राथमिकता के लिए बिड की मात्रा का अधिकतम प्रतिशत / Maximum Percentage of Bid quantity for MSE purchase preference	25

1. If the bidder is a Micro or Small Enterprise (MSE) as per latest orders issued by Ministry of MSME, the bidder shall be relaxed from the eligibility criteria of "Bidder Turnover" as defined above subject to meeting of quality and technical specifications. If the bidder itself is MSE OEM of the offered products, it would be relaxed from the "OEM Average Turnover" criteria also subject to meeting of quality and technical specifications. The bidder seeking Relaxation from Turnover, shall upload the supporting documents to prove his eligibility for Relaxation.
2. If the bidder is a DPIIT registered Startup, the bidder shall be relaxed from the the eligibility criteria of "Bidder Turnover" as defined above subject to their meeting of quality and technical specifications. If the bidder is DPIIT Registered OEM of the offered products, it would be relaxed from the "OEM Average Turnover" criteria also subject to meeting of quality and technical specifications. The bidder seeking Relaxation from Turnover shall upload the supporting documents to prove his eligibility for Relaxation.
3. The minimum average annual financial turnover of the bidder during the last three years, ending on 31st March of the previous financial year, should be as indicated above in the bid document. Documentary evidence in the form of certified Audited Balance Sheets of relevant periods or a certificate from the Chartered Accountant / Cost Accountant indicating the turnover details for the relevant period shall be uploaded with the bid. In case the date of constitution / incorporation of the bidder is less than 3-year-old, the average turnover in respect of the completed financial years after the date of constitution shall be taken into account for this criteria.
4. Experience Criteria: In respect of the filter applied for experience criteria, the Bidder or its OEM of the product offered in the bid {themselves or through reseller(s)} should have regularly, manufactured and supplied same or similar Category Products to any Central / State Govt Organization / PSU for number of Financial years as indicated above in the bid document before the bid opening date. Copies of relevant contracts and delivery acceptance certificates like CRAC to be submitted along with bid in support of having supplied some quantity during each of the Financial year. In case of bunch bids, the category of primary product having highest value should meet this criterion.
5. OEM Turn Over Criteria: The minimum average annual financial turnover of the OEM of the offered product during the last three years, ending on 31st March of the previous financial year, should be as indicated in the bid document. Documentary evidence in the form of certified Audited Balance Sheets of relevant periods or a certificate from the Chartered Accountant / Cost Accountant indicating the turnover details for the relevant period shall be uploaded with the bid. In case the date of constitution / incorporation of the OEM is less than 3 year old, the average turnover in respect of the completed financial years after the date of constitution shall be taken into

account for this criteria.

6. Preference to Make In India products (For bids < 200 Crore): Preference shall be given to Class 1 local supplier as defined in public procurement (Preference to Make in India), Order 2017 as amended from time to time and its subsequent Orders/Notifications issued by concerned Nodal Ministry for specific Goods/Products. The minimum local content to qualify as a Class 1 local supplier is denoted in the bid document. If the bidder wants to avail the Purchase preference, the bidder must upload a certificate from the OEM regarding the percentage of the local content and the details of locations at which the local value addition is made along with their bid, failing which no purchase preference shall be granted. In case the bid value is more than Rs 10 Crore, the declaration relating to percentage of local content shall be certified by the statutory auditor or cost auditor, if the OEM is a company and by a practicing cost accountant or a chartered accountant for OEMs other than companies as per the Public Procurement (preference to Make-in -India) order 2017 dated 04.06.2020. Only Class-I and Class-II Local suppliers as per MII order dated 4.6.2020 will be eligible to bid. Non - Local suppliers as per MII order dated 04.06.2020 are not eligible to participate. However, eligible micro and small enterprises will be allowed to participate. The buyers are advised to refer the OM No.F.1/4/2021-PPD dated 18.05.2023. [OM No.1 4 2021 PPD dated 18.05.2023](#) for compliance of Concurrent application of Public Procurement Policy for Micro and Small Enterprises Order, 2012 and Public Procurement (Preference to Make in India) Order, 2017.

7. Purchase preference will be given to MSEs having valid Udyam Registration and whose credentials are validated online through Udyam Registration portal as defined in Public Procurement Policy for Micro and Small Enterprises (MSEs) Order, 2012 dated 23.03.2012 issued by Ministry of Micro, Small and Medium Enterprises and its subsequent Orders/Notifications issued by concerned Ministry. If the bidder wants to avail themselves of the Purchase preference, the bidder must be the manufacturer / OEM of the offered product on GeM. Traders are excluded from the purview of Public Procurement Policy for Micro and Small Enterprises and hence resellers offering products manufactured by some other OEM are not eligible for any purchase preference. In respect of bid for Services, the bidder must be the Service provider of the offered Service. Relevant documentary evidence in this regard shall be uploaded along with the bid in respect of the offered product or service and Buyer will decide eligibility for purchase preference based on documentary evidence submitted, while evaluating the bid. If L-1 is not an MSE and MSE Seller (s) has / have quoted price within L-1+ 15% (Selected by Buyer) of margin of purchase preference /price band defined in relevant policy, such MSE Seller shall be given opportunity to match L-1 price and contract will be awarded for 25% (selected by Buyer) percentage of total quantity. The buyers are advised to refer the OM No. F.1/4/2021-PPD dated 18.05.2023 [OM No.1 4 2021 PPD dated 18.05.2023](#) for compliance of Concurrent application of Public Procurement Policy for Micro and Small Enterprises Order, 2012 and Public Procurement (Preference to Make in India) Order, 2017. Benefits of MSE will be allowed only if seller is validated on-line in GeM profile as well as validated and approved by Buyer after evaluation of documents submitted.

8. Estimated Bid Value indicated above is being declared solely for the purpose of guidance on EMD amount and for determining the Eligibility Criteria related to Turn Over, Past Performance and Project / Past Experience etc. This has no relevance or bearing on the price to be quoted by the bidders and is also not going to have any impact on bid participation. Also this is not going to be used as a criteria in determining reasonableness of quoted prices which would be determined by the buyer based on its own assessment of reasonableness and based on competitive prices received in Bid / RA process.

9. Past Performance: The Bidder or its OEM {themselves or through re-seller(s)} should have supplied same or similar Category Products for 30% of bid quantity, in at least one of the last three Financial years before the bid opening date to any Central / State Govt Organization / PSU. Copies of relevant contracts (proving supply of cumulative order quantity in any one financial year) to be submitted along with bid in support of quantity supplied in the relevant Financial year. In case of bunch bids, the category related to primary product having highest bid value should meet this criterion.

एक्सेल में अपलोड किए जाने की आवश्यकता /Excel Upload Required :

Cost of disposables, consumables, accessories, etc - [1769767527.xlsx](#)

Pre Bid Detail(s)

मूल्य भिन्नता खंड दस्तावेज़/Pre-Bid Date and Time	प्री-बिड स्थान/Pre-Bid Venue
10-02-2026 16:00:00	eProcurement Cell /through VC

मूल्यांकन विधि(मदवार मूल्यांकन विधि) / Evaluation Method (Item Wise Evaluation Method)

Contract will be awarded schedulewise and the determination of L1 will be done separately for each schedule. The

details of item-consignee combination covered under each schedule are as under:

मूल्यांकन अनुसूचियां / Evaluation Schedules	अनुमानित मूल्य / Estimated Value	वस्तु/श्रेणी / Item/Category	मात्रा / Quantity
Schedule 1	1200000	Microneurography (msna, Ssna) , Ssr Add-on Powerlab And Labchart Software	1
Schedule 2	5600000	Portable Hemodynamic Recorder For Cardiovascular Lab(Portable Continous Non Invasive Haemodynamic M	1
Schedule 3	3600000	Functional Near-infrared Spectroscopy (fnirs) System For Human Brain Imaging	1

Microneurography (MSNA, SSNA) , SSR Add-On Powerlab And Labchart Software (1 Unit)

(क्रमशः श्रेणी 1 और श्रेणी 2 के स्थानीय आपूर्तिकर्ता के रूप में अर्हता प्राप्त करने के लिए आवश्यक/Minimum 50% and 20% Local Content required for qualifying as Class 1 and Class 2 Local Supplier respectively)

तकनीकी विशिष्टियाँ /Technical Specifications

क्रेता विशिष्टि दस्तावेज़ /Buyer Specification Document	Download
---	--------------------------

व्यापक रखरखाव / Comprehensive Maintenance	
Warranty of required product	2 Year
Comprehensive Maintenance Duration (Post Warranty)	3 Year

*Warranty displayed under the AMC/CMC Details section will supersede the warranty displayed under the catalog specification

परेषिती/रिपोर्टिंग अधिकारी तथा मात्रा/Consignees/Reporting Officer and Quantity

क्र.सं./S.N o.	परेषिती/रिपोर्टिंग अधिकारी /Consignee Reporting/Officer	पता/Address	मात्रा /Quantity	डिलीवरी के दिन/Delivery Days
1	Khrawkupar Jithod Katrai	793018,P.O. NEIGRIHMS, Mawdiangdiang, Shillong	1	90

Portable Hemodynamic Recorder For Cardiovascular Lab(Portable Continous Non Invasive Haemodynamic M (1 Unit)

(क्रमशः श्रेणी 1 और श्रेणी 2 के स्थानीय आपूर्तिकर्ता के रूप में अर्हता प्राप्त करने के लिए आवश्यक/Minimum 50% and 20% Local Content required for qualifying as Class 1 and Class 2 Local Supplier respectively)

तकनीकी विशिष्टियाँ /Technical Specifications

क्रेता विशिष्टि दस्तावेज़ /Buyer Specification Document	Download
--	--------------------------

व्यापक रखरखाव / Comprehensive Maintenance	
Warranty of required product	2 Year
Comprehensive Maintenance Duration (Post Warranty)	8 Year

*Warranty displayed under the AMC/CMC Details section will supersede the warranty displayed under the catalog specification

परिषी/रिपोर्टिंग अधिकारी तथा मात्रा/Consignees/Reporting Officer and Quantity

क्र.सं./S.N o.	परिषी/रिपोर्टिंग अधिकारी /Consignee Reporting/Officer	पता/Address	मात्रा /Quantity	डिलीवरी के दिन/Delivery Days
1	Khrawkupar Jithod Katrai	793018,P.O. NEIGRIHMS, Mawdiangdiang, Shillong	1	90

Functional Near-Infrared Spectroscopy (FNIRS) System For Human Brain Imaging (1 Unit)

(क्रमशः श्रेणी 1 और श्रेणी 2 के स्थानीय आपूर्तिकर्ता के रूप में अर्हता प्राप्त करने के लिए आवश्यक/Minimum 50% and 20% Local Content required for qualifying as Class 1 and Class 2 Local Supplier respectively)

तकनीकी विशिष्टियाँ /Technical Specifications

क्रेता विशिष्टि दस्तावेज़ /Buyer Specification Document	Download
--	--------------------------

व्यापक रखरखाव / Comprehensive Maintenance	
Warranty of required product	2 Year
Comprehensive Maintenance Duration (Post Warranty)	8 Year

*Warranty displayed under the AMC/CMC Details section will supersede the warranty displayed under the catalog specification

परिषी/रिपोर्टिंग अधिकारी तथा मात्रा/Consignees/Reporting Officer and Quantity

क्र.सं./S.N o.	प्रेषित/रिपोर्टिंग अधिकारी /Consignee Reporting/Officer	पता/Address	मात्रा /Quantity	डिलीवरी के दिन/Delivery Days
1	Khrawkumar Jithod Katrai	793018,P.O. NEIGRIHMS, Mawdiangdiang, Shillong	1	90

क्रेता द्वारा जोड़ी गई बिड की विशेष शर्तें/Buyer Added Bid Specific Terms and Conditions

1. Authorised Service Centre within the state of Odisha, along with a dedicated contact person with telephone number for technical solution in a fast track basis at this institution as and when required basis.
2. Experience Certificate for the supply of the same to any Govt/ PSU/ any renowned private organisation along with Supply/ Purchase Order.
3. If the agency is registered under MSME or NSIC, then EMD exemption certificate needs to be enclosed.
4. Make in india specific authorisation certificate needs to be enclosed.
5. **Generic**

Bidder financial standing: The bidder should not be under liquidation, court receivership or similar proceedings, should not be bankrupt. Bidder to upload undertaking to this effect with bid.

6. **Generic**

Bidders are advised to check applicable GST on their own before quoting. Buyer will not take any responsibility in this regards. GST reimbursement will be as per actuals or as per applicable rates (whichever is lower), subject to the maximum of quoted GST %.

7. **Generic**

Data Sheet of the product(s) offered in the bid, are to be uploaded along with the bid documents. Buyers can match and verify the Data Sheet with the product specifications offered. In case of any unexplained mismatch of technical parameters, the bid is liable for rejection.

8. **Generic**

Buyer Organization specific Integrity Pact shall have to be complied by all bidders. Bidders shall have to upload scanned copy of signed integrity pact as per Buyer organizations policy along with bid. [Click here to view the file](#)

9. **Service & Support**

Availability of Service Centres: Bidder/OEM must have a Functional Service Centre in the State of each Consignee's Location in case of carry-in warranty. (Not applicable in case of goods having on-site warranty). If service center is not already there at the time of bidding, successful bidder / OEM shall have to establish one within 30 days of award of contract. Payment shall be released only after submission of documentary evidence of having Functional Service Centre.

10. **Service & Support**

Dedicated /toll Free Telephone No. for Service Support : BIDDER/OEM must have Dedicated/toll Free Telephone No. for Service Support.

11. **Service & Support**

Escalation Matrix For Service Support : Bidder/OEM must provide Escalation Matrix of Telephone Numbers for Service Support.

12. **Buyer Added Bid Specific ATC**

पूर्वोत्तरइंदिरागांधीक्षेत्रीयस्वास्थ्यएवंआयुर्विज्ञानसंस्थान

North Eastern Indira Gandhi Regional Institute of Health and Medical Sciences

(भारतसरकार, स्वास्थ्यएवंपरिवारकल्याणमंत्रालय, स्वायत्तसंस्थान)

(An Autonomous Institute, under Ministry of Health and Family Welfare, Government of India)

निदेशकब्लॉक, मावडियांगडंग, शिलांग -793 018 (मेघालय) /Director's Block, Mawdiangdiang, Shillong -793 018 (Meghalaya)

Store & Procurement Section; Email: storeneigrihms@gmail.com; Tele Fax: (0364) 2538032; Website: neigrihms.gov.in

F. No: STOPRO -PHYS/2/2025-Store (PHYSIOLOGY)

-

Reference:

ü IJPC No. 1/2025 -26 /Agenda No. 76

ü IJPC No. 2/2025 -26 /Agenda No. 9

-

A. **Description of Items:** Processing of Functional Near-Infrared Spectroscopy (fNIRS) System for Human Brain Imaging, with 2 years warranty and thereafter 8 years CMC, for department of Physiology.

1. Should be able to measure hemodynamic neuro-activation via HbDeoxy, HbOxy, Hb total changes in the cerebral cortex, capable of measuring anywhere on the Head.
2. The whole system should be housed in one chassis. Stacking of units are not acceptable.
3. Should offer sampling rate upto 160 hz.
4. Should have a high resolution of 7.5mm.
5. Should support both types of Sources - LED or Laser.
6. Computer free recording option should be available.
7. System should provide 255 TTL trigger inputs. Should have an automated logging of input triggers from 3rd party devices for syncing data with other modalities like EEG (preferably using TTL).
8. Should provide topography software licensing free and unlimited.
9. Should have an option to use manual button press triggers during mobile data recording scenarios. Trigger button should connect directly to the fNIRS device using cable.
10. System should be modular and expandable in the same device. Should be Upgradable at least up to 80 Sources and 80 Detectors. It should be expandable to use 5 units to give maximum no of sources and detectors.
11. System should be offered with 8 High Powered individual Sources and 8 detector optodes with programmable/automated gain adjustment to maximize dynamic range of measurement. Each source optode should have a dual tip. Device should offer upto 60-65 channels.

12. Freely configurable source & detectors montages to measure anywhere/ wholehead, concurrently with (nearly) any other modality such as F-MRI, TMS, EEG and tDCS within a single cap.
13. Should be supplied with 2 pcs caps, each with 128 positions for detectors and sources and an Overcap to cover the probes mounted on the caps.
14. Should be able to record the data for up to 6 hrs with the integrated battery. The battery life should be extended by using power bank or AC mains adapter.
15. System should be able to transmit the data via both USB and Wi-fi to the recording laptop. It should be able to store data onboard the device standalone mode, when not connected to the USB or Wi-fi.
16. System should be expandable to record physiological data - such as PPG, heart rate variability (HRV), SPO2, respiration, temperature, galvanic skin response (GSR), and bipolar signals such as EMG and ECG measurement in addition to fNIRS data.
17. Should have two wavelengths - 760 nm & 850 nm.
18. Should have Digital Trigger (I/O).
19. Should have Options of standardized probe positioning as well as end-user-configurable setups.
20. System should have an option for short channel detectors at min. at 0.8 cm distance without sacrificing any sources.
21. Availability for raw data in open format (for example - ascii) for independence
22. Programmable source firing sequence to support multi-site brain studies (e.g., left and right hemispheres).
23. Safety certification (CE Certified) for human research applications.
24. Softwares for presentation, Data Acquisition, Topography & Tomography from the same manufacturer for full compatibility
25. Software should automatically suggest based on selected ROIs.
26. Software should do a contact quality check for each source-detector pair individually before acquisition
27. Software should allow the real time data collection, real time visualisation of data streams and activation on 2D/3D standard head models. Should have Digital Trigger (I/O) to synchronize experiments (for example from E-Prime, BCI & Neurofeedback)
28. Data analysis software should offer Topographic and tomographic data analysis and image reconstruction software with 3D MR overlay capabilities, Image reconstruction, Finite Element Mesh, Statistical Parametric modeling.
29. Should be offered with Static phantom for system calibration and testing.
30. Should offer features for real time data streams, brain activation, and block averages - data exportable for analysis in any environment.
31. Should offer Real-time updates of hemodynamic response to event related protocols. Should offer Real-time hyperscanning topographic displays (Hbdeoxy, Hboxy, Hbttotal).
32. Should have Programmable display capability to highlight pre-selected sensor locations. Should have On-line frequency filtering of displayed results. Should offer 3D Depth-discrimination, time multiplexing, dynamic gain state switching. Should be able to Quick automated calibration, automated hardware diagnostics; signal impedance quality checks.
33. Should also be supplied with software for experiment presentation and hardware devices for trigger marking.
34. System should be supplied with spring loaded holders.
35. Software should provide all pre-processing steps (such as filtering, artifact reduction, data trimming, conversion from light to hemodynamic data, statistical analysis using SPM - Level 1 for one subject as

well as Level 2 for group analysis.

36. Data can be exported for processing in NIRS-SPM, MATLAB toolbox or homer, as per the researchers desire.
37. System should have more than 500 high impact formal publications.
38. Weight: Should be less than 1kg
39. System should be offered with stimulation software such as E-Prime which allows the designing of more complex psychological experiments. Proper event locking with the hardware (e.g. response key) of third party software should be there.
40. Designing of complex psychological experiments requiring multiple levels of responses should be possible.
41. Compatible with visual, auditory and speech tasks. All the hardware related to stimulus presentation (for visual, auditory, speech related experiments) such as response pads, microphone, headphone, keyboards etc. which are typically required in ERP experiments must be provided.
42. Hardware related to third party stimulus presentation software must be provided.
43. Milliseconds time accuracy of the EEG/ERP systems must be guaranteed.
44. The hardware should allow the customization of trigger pulses so that the events are marked perfectly on EEG.
45. Should be supplied with stimulus presentation software with a laptop like E Prime, etc. & hardware for experimental control and data acquisition such as response pads or response boxes, microphones, that are needed for ERP experiments must be included.
46. Should be supplied with a desktop having the configuration: Windows: 10/11 (64-bit), 8GB RAM, i7 Processor, 512 GB SSD.

–

- B. **Description of Items:** Processing of **Microneurography (MSNA, SSNA), SSR add-on Power lab and Labchart Software**, with 2 years warranty and thereafter 3 years CMC, for department of Physiology.

–

1. The FE185 Neuro Amp EX is designed to perform isolated measurements of single-unit or multi-unit extracellular action potentials from single neurons or axons, from humans or animals.
2. The Neuro Amp EX is certified safe for Human Use, BF rated, when used with the supplied MLT185 headstage.
3. The Neuro Amp EX can only be used with metal electrodes and is supplied with six Amphenol connectors (microelectrodes are not supplied)
4. These research microelectrode neurography needles are designed for peripheral nerve recording and stimulation, Epoxylite Insulated microelectrode; special tip shape Sold in pack of 12. The PTK12 Electrodermal Activity Kit is suitable for measuring the galvanic skin response and skin temperature in humans and includes the FE116 GSR Amp with MLT118F Finger Electrodes, ML309 Thermistor Pod and MLT422/A Skin Temperature Probe (2 m).

–

- C. **Description of Items:** Processing of Portable Hemodynamic Recorder for Cardiovascular Lab (Portable Continuous Non Invasive Haemodynamic Monitoring), with 2 years warranty and thereafter 8 years CMC, for department of Physiology.

1. Should be able to record continuous beat-to-beat blood pressure non-invasively from the human subject for long term recording.

2. Should have a recording range of at least 60 to 250 mm Hg for systolic Bp and 40 to 200 mm Hg for diastolic BP.
3. Should be working on the principle of volume clamp (penaz technique) using reusable finger cuffs
4. Should have provision to switch between 2 finger cuffs for pressure recording to reduce subject discomfort of long-term recordings.
5. One of each of finger cuffs small, medium and large sizes.
6. Reusable finger cuffs with unlimited usage without damage and self-life of at-least 3 years.
7. Should have the provision to manually switch off the calibration process to record continuous
8. uninterpreted blood pressure signal at least for a duration of 5 minutes or for a user defined customizable duration.
9. Should have facility for Height correction unit to account for the level difference between recording site and heart.
10. Accuracy 1% for cuff Pressure (within 6 mmHg) and Heart Rate Accuracy [HR (bpm)/ 60] % Should have facility for online and offline recording & Analysis.
11. Online streaming of data to Matlab and Excel.
12. The analyzed data should be exported to text, IGoR and other file formats online and offline.
13. The Blood pressure data should be integrated & analyzed with HRV data (time domain and Frequency domain) online and offline.
14. The Blood pressure data should be integrated with handgrip dynamometer for handgrip test.
15. The system should provide cardiac output (CO), Stroke volume (SV), Total peripheral Resistance (TPR), Baroreflex sensitivity (BRS).
16. The software should perform heart rate variability online and offline with real time cardiac and vector analysis.
17. Necessary software for recording and analysis of blood pressure signal. The software should provide hemodynamic parameters from pressure data, wave forms etc.
18. The operating temperature must be within 10- 40 deg C (50 - 104 deg F)
19. The operating humidity (non - condensing) should be within 5 -90%.
20. Software update and upgrades should be provided for at least next 5 years.
21. The complete system should be portable and easy to carry less than 5 kgs including the carry case.
22. Onsite training from the company experts should be provided.
23. Should be compatible and seamlessly interface to software for recording and analysis without any additional hardware.
24. Should be compatible with Lab Chart Software.
25. Must have necessary certification like ISO, IEC, CE marked or as per DPIIT and Make in India policies.
26. The complete system should be portable and easy to carry less than 5 kgs including the carry case for field studies.

27. Onsite training from the company experts should be provided.

-
-

Buyer Added Bid Specific Terms and Conditions:-

-

A. Following mandatory documents must be attached in the bid document as specified, failing which bid will be treated as " Non-Responsive":

-

1. Cost of spares, consumables and accessories not covered under warranty and CMC period shall be offered as percentage value of the system/unit in the Technical Bid ATC1/2. However, warranty includes all the spares including probes mentioned in the technical specification.
2. Documents with regard to details compliance statement to be attached at "**Additional Doc 2 (Requested in ATC)**"
3. Documents with regard to Original Literature, Product catalogue, technical datasheet from the firm/O.E.M with Highlighting as per the technical specification must attach At "**Additional Doc 3 (Requested in ATC)**"
4. Documents with regard to list of Offering/Quoted items mentioning make, model & quantity of each store must be "**Additional Doc 4(Requested in ATC)**"
5. Component wise pricing of all equipment/turnkey/electrical/accessories/HVAC etc, must be submitted in the "Financial Document" and not in technical Bid. Any detail price bid/Component wise pricing should not be attached in the technical bid ,failing which bid will be consider as "Techno Commercially Non Responsive"
6. Turnkey scope includes vinyl flooring, wall partitions and alterations as per drawings /as finalised by user, and metal ceiling installation as per bid document; electrical wiring, lighting as per bid document; other fittings as required for installation of the equipment across the entire area specified.

-

B. Warranty and Maintenance:

-

1. Warranty for ___ years followed by CMC for ___ years including spares & service for all the items supplied in this particular tender including third-party items and turnkey works.
2. Mandatory 2 PMs /Year with unlimited breakdown calls have to be attended by the Bidder/manufacturer throughout the warranty & CMC period at site.i.e. NEIGRIHMS, SHILONG
3. Duly signed Mandatory PM reports must be submitted periodically failing which necessary action will be initiated as per term& condition of the tender.

C. Scope of work

1. The bidder has to mandatorily quote for all the items /stores mentioned in the Technical Specification & Scope of Work ,failing which bid will be treated as" Non-Responsive" Specification & Scope of Work ,failing which bid will be treated as" Non-Responsive"

2. Prices of all accessories (reagents, consumables, controls closed/ open etc) need to be clearly indicated as % of the total price in the technical bid Technical Bid ATC1/2 and will be considered for evaluation as in BOQ or else bid will be treated as "non responsive". The CMC cost offered to be indicated in the ATC and shall be quoted in percentage (inclusive of GST) and shall not exceed 5% of the system cost.
3. E-bidder must adhere to Government of India, Ministry of Finance, PPD division Public procurement order OM F.No.6/18/2019-PPD dated 23rd July, 2020 inserting Rule 144(Xi) in GFR 2017, No 1 dated: 23/7/2020 and subsequent Orders No 2 & 3 or as amended from time to time, failing which the bids shall be treated as non-responsive.

D. Buyer Added Bid Specific Terms and Conditions :-

I. Generic

1. End User Certificate: Wherever Bidders are insisting for End User Certificate from the Buyer, same shall be provided in Buyer's standard format only.
2. Experience Criteria: The Bidder or its OEM {themselves or through reseller(s)} should have regularly, manufactured and supplied same or similar Category Products to any Central / State Govt Organization / PSU / Public Listed Company for 3 years before the bid opening date. Copies of relevant contracts to be submitted along with bid in support of having supplied some quantity during each of the year. In case of bunch bids, the primary product having highest value should meet this criterion.
3. IT equipment shall be IPv6 ready from day one.
4. Installation, Commissioning, Testing, Configuration, Training (As applicable as per scope of supply) is to be carried out by OEM / OEM Certified resource or OEM authorized Reseller.
5. Upload Manufacturer authorization: Wherever Authorized Distributors are submitting the bid, Manufacturers Authorization Form (MAF) /Certificate with OEM details such as name, designation, address, e-mail Id and Phone No. required to be furnished along with the bid.
6. The successful bidder has to supply all essential accessories required for the successful installation and commissioning of the goods supplied. Besides standard accessories as per normal industry practice, following accessories must be part of supply and cost should be included in bid price: All the items and accessories as per Technical Specification.
7. The Buyer has an existing set up / inventory of similar products. The offered / supplied product must be compatible with existing system. The bidder has to ensure Compatibility of the supplied items or shall have to include in the supply the necessary hardware / software to make them compatible at no extra cost to the buyer. The details of items with which compatibility is required are as under: all the spares including UPS, PC, battery, Printer, Probes & upgradation of System Software & third party Software

II. Scope of Supply

Scope of supply (Bid price to include all cost components): Supply Installation Testing Commissioning of Goods, Training of operators and providing Statutory Clearances required (if any)

III. **Turnover**

-

1. Bidder Turnover Criteria: The minimum average annual financial turnover of the bidder during the last three years, ending on 31st March of the previous financial year, should be as indicated in the bid document. Documentary evidence in the form of certified Audited Balance Sheets of relevant periods or a certificate from the Chartered Accountant / Cost Accountant indicating the turnover details for the relevant period shall be uploaded with the bid. In case the date of constitution / incorporation of the bidder is less than 3 year old, the average turnover in respect of the completed financial years after the date of constitution shall be taken into account for this criteria.
2. OEM Turnover Criteria: The minimum average annual financial turnover of the OEM of the offered product during the last three years, ending on 31st March of the previous financial year, should be as indicated in the bid document. Documentary evidence in the form of certified Audited Balance Sheets of relevant periods or a certificate from the Chartered Accountant / Cost Accountant indicating the turnover details for the relevant period shall be uploaded with the bid. In case the date of constitution / incorporation of the OEM is less than 3 year old, the average turnover in respect of the completed financial years after the date of constitution shall be taken into account for this criteria. In case of bunch bids, the OEM of CATEGORY RELATED TO primary product having highest bid value should meet this criterion.

IV. **OEM**

IMPORTED PRODUCTS: In case of imported products, OEM or Authorized Seller of OEM should have a registered office in India to provide after sales service support in India. The certificate to this effect should be submitted.

V. **Purchase Preference (Centre)**

As per DPIIT notification at the time of e-tender , bidding or solicitation the bids shall be required to indicate percentage of local content and provide self-certification (by Director/ Company Secretary) and also give details of the location/s at which value addition is made". Since the bidder here is not the local supplier, the same was required to be obtained from the "Class-I local supplier /Class II local supplier"

Further the details of Calculations of local content areas under:

Question 1. How to calculate Local Content?

Answer: Para 2 of the PPP-MII Order, 2017 (as amended on 16.09.2020) defines local content as the amount of value added in India which shall, unless otherwise prescribed by the Nodal Ministry, be the total value of the item procured (excluding net domestic indirect taxes) minus the value of imported content in the item (including all customs duties) as a proportion of the total value, in percent.

Mathematically,

Local content = (Sale price - Value of imported content) * 100/ Sale price

Where, "Sale price" means price excluding net domestic indirect taxes and "Value of imported content" means price of imported content inclusive of all customs duties

Question2. How to calculate Local Content in bids involving supply of multiple items from single bidder?

Answer: In case of bids requiring supply of multiple items (say "X1", "X2" and "X3") by a single bidder, the local content in the bid shall be

Local content = ((Sale price of "X1" - Value of imported content in "X1") + (Sale price of "X2" - Value of imported content in "X2") + (Sale price of "X3" - Value of imported content in "X3")) * 100/ (Sale price of "X1" + Sale price of "X2" + Sale price of "X3")

VI. **Service & Support**

1. Availability of Service Centres: Bidder/OEM must have a Functional Service Centre in the State of each Consignee's Location in case of carry-in warranty. (Not applicable in case of goods having on-site warranty). If service center is not already there at the time of bidding, successful bidder / OEM shall have to establish one within 30 days of award of contract. Payment shall be released only after submission of documentary evidence of having Functional Service Centre.
2. Dedicated /toll Free Telephone No. for Service Support : BIDDER/OEM must have Dedicated/toll Free Telephone No. for Service Support.
3. Escalation Matrix For Service Support : Bidder/OEM must provide Escalation Matrix of Telephone Numbers for Service Support.

VII. **Certificates**

1. Bidder's offer is liable to be rejected if they don't upload any of the certificates/ documents sought in the Bid document, ATC and Corrigendum if any.
2. The bidder or the OEM of the offered products must have BIS/WHO-GMP/ CDSCO Indian certification or alternate certification as recognized by Government of India
3. Material Test Certificate Should Be Sent Along with The Supply. The Material Will Be Checked by Buyer's Lab & the Results of the Lab will be the Sole Criteria for Acceptance of the Item.
4. The bidder is required to upload, along with the bid, all relevant certificates such as BIS licence, type test certificate, approval certificates and other certificates as prescribed in the Product Specification given in the bid document.
5. To be eligible for award of contract, Bidder / OEM must possess following Certificates / Test Reports on the date of bid opening (to be uploaded with bid): All the quality & electrical safety certificates.

VIII. **Warranty**

1. Bidder /OEM has to give an undertaking that after expiry of warranty period, it will provide Comprehensive Maintenance Service for next 8 years for the offered products at the rate not more than 5% of contract price per annum. Buyer reserves the right to enter into a CMC agreement with the Successful Bidder / OEM after expiry of the Warranty period at above mentioned rate and the payment for the CMC charges would be made biannually after rendering of the CMC Services of the relevant CMC period. Performance Security of the successful bidder shall be forfeited if it fails to accept the CMC contract when called upon by the buyer. CMC would include cost of all the spares including UPS, PC, battery, Printer, Probes & upgradation of System Software & third party Software (Upload the undertaking). The original Performance Security of contract will be returned only after submission and verification of CMC Performance Security for 5% of total CMC value valid up to CMC period plus 2 months (if there is no other claim).
2. Warranty period of the supplied products shall be ___ years (as indicated in the technical specification /GeM condition) years from the date of final acceptance of goods or after completion of installation, commissioning & testing of goods (if included in the scope of supply), at consignee location. OEM Warranty certificates must be submitted by Successful Bidder at the time of delivery of Goods. The seller should guarantee the rectification of goods in case of any breakdown during the guarantee period. Seller should have well established Installation, Commissioning, Training, Troubleshooting and Maintenance Service group in INDIA for attending the after sales service. Details of Service Centres near consignee destinations are to be uploaded along with the bid.
3. Over and above the normal Warranty terms as per GeM GTC, the successful bidder / OEM shall have to provide Comprehensive Warranty during the entire Standard warranty period as per contract. : The comprehensive warranty shall be covering the following scope all the spares including UPS, PC, battery ,Printer ,Probes & upgradation of System Software & third party Software (Upload an undertaking with the bid confirming compliance by the bidder if Bidder is taking onus of this compliance. In case OEM is taking onus of this compliance, OEM undertaking is to be uploaded along with Bidder undertaking)
4. Successful bidder will have to ensure that adequate number of dedicated technical service personals / engineers are designated / deployed for attending to the Service Request in a time bound manner and for ensuring Timely Servicing / rectification of defects during warranty period, as per Service level agreement indicated in the relevant clause of the bid.
5. Timely Servicing / rectification of defects during warranty period: After having been notified of the defects / service requirement during warranty period, Seller has to complete the required Service / Rectification within 3 days' time limit. If the Seller fails to complete service / rectification with defined time limit, a penalty of 0.5% of Unit Price of the product shall be charged as penalty for each week of delay from the seller. Seller can deposit the penalty with the Buyer directly else the Buyer shall have a right to recover all such penalty amount from the Performance Security (PBG). Cumulative Penalty cannot exceed more than 10% of the total contract value after which the Buyer shall have the right to get the service / rectification done from alternate sources at the risk and cost of the Seller besides forfeiture of PBG. Seller shall be liable to re-imburse the cost of such service / rectification to the Buyer.

IX. **Past Project Experience**

1. For fulfilling the experience criteria any one of the following documents may be considered as valid proof for meeting the experience criteria:

- (a) Purchase Order copy along with Invoice(s) with self-certification by the bidder that supplies against the invoices have been executed.
 - (b) Execution certificate by client with order value.
 - (c) Any other document in support of order execution like Third Party Inspection release note, etc.
2. The Bidder /OEM (themselves or through reseller(s)), should have executed project for supply and installation /commissioning of same or similar Category Products during preceding 3 financial years (i.e. current year and three previous financial years) as on opening of bid, as per following criteria:

Single order of at least 35% of estimated bid value;

or

Two orders of at least 20% each of estimated bid value;

or

Three orders of at least 15% each of estimated bid value.

Satisfactory Performance certificate issued by respective Buyer Organization for the above Orders should be uploaded with bid. In case of bunch bids, the Category related to primary product having highest bid value should meet this criterion

X. **Forms of EMD and PBG**

1. Bidders can also submit the EMD with Account Payee Demand Draft in favour of NEIGRIHMS EMD SECURITY DEPOSITS payable at MAWDIANGDIANG, SHILLONG -793018, MEGHALAYA. Bidder has to upload scanned copy / proof of the DD along with bid and has to ensure delivery of hardcopy to the Buyer within Bid End date & time / Bid Opening date & time.
2. Bidders can also submit the EMD with Fixed Deposit Receipt made out or pledged in the name of A/C (Name of the Buyer). The bank should certify on it that the deposit can be withdrawn only on the demand or with the sanction of the pledge. For release of EMD, the FDR will be released in the favour of the bidder by the Buyer after making endorsement on the back of the FDR duly signed and stamped along with covering letter. Bidder has to upload scanned copy/ proof of the FDR along with bid and has to ensure delivery of hardcopy to the Buyer within Bid End date & time / Bid Opening date & time.
3. Bidders can also submit the EMD with Banker's Cheque in favour of NEIGRIHMS EMD SECURITY DEPOSITS payable at MAWDIANGDIANG, SHILLONG-793018, MEGHALAYA . Bidder has to upload scanned copy / proof of the BC along with bid and has to ensure delivery of hardcopy to the Buyer within Bid End date & time / Bid Opening date & time.
4. Bidders can also submit the EMD with Payment online through RTGS / internet banking in Beneficiary name NEIGRIHMS EMD SECURITY DEPOSITS Account No. 30270200000027 IFSC Code BARB0MAWDIA Bank Name BANK OF BARODA Branch address MAWDIANGDIANG, SHILLONG-793018, MEGHALAYA. Bidder to indicate bid number and name of bidding entity in the transaction details field at the time of on- line transfer. Bidder has to upload scanned copy / proof of

the Online Payment Transfer along with bid.

5. Successful Bidder can submit the Performance Security in the form of Account Payee Demand Draft also (besides PBG which is allowed as per GeM GTC). DD should be made in favour of NEIGRIHMS EMD SECURITY DEPOSITS payable at MAWDIANGDIANG, SHILLONG -793018, MEGHALAYA. After award of contract, Successful Bidder can upload scanned copy of the DD in place of PBG and has to ensure delivery of hard copy to the original DD to the Buyer within 15 days of award of contract.
6. Successful Bidder can submit the Performance Security in the form of Fixed Deposit Receipt also (besides PBG which is allowed as per GeM GTC). FDR should be made out or pledged in the name of NEIGRIHMS EMD SECURITY DEPOSITS A/C (Name of the Seller). The bank should certify on it that the deposit can be withdrawn only on the demand or with the sanction of the pledgee. For release of Security Deposit, the FDR will be released in favour of bidder by the Buyer after making endorsement on the back of the FDR duly signed and stamped along with covering letter. Successful Bidder has to upload scanned copy of the FDR document in place of PBG and has to ensure delivery of hard copy of Original FDR to the Buyer within 15 days of award of contract.
7. Successful Bidder can submit the Performance Security in the form of Payment online through RTGS / internet banking also (besides PBG which is allowed as per GeM GTC). On-line payment shall be in Beneficiary name NEIGRIHMS EMD SECURITY DEPOSITS Account No. 3027020000027 IFSC Code BARB0MAWDIA Bank Name BANK OF BARODA Branch address MAWDIANGDIANG, SHILLONG-793018, MEGHALAYA. Successful Bidder to indicate Contract number and name of Seller entity in the transaction details field at the time of on-line transfer. Bidder has to upload scanned copy / proof of the Online Payment Transfer in place of PBG within 15 days of award of contract.

(E) Additional Terms and conditions & Scope of Work for CMC:-

Tenderer/Vendors/contractor should note that the following terms and conditions will apply specifically in addition to the Rules and the Regulation as applicable to such provide services in the Government of India.

1. Comprehensive Annual Maintenance Contract must include Labour, spares & Preventive Maintenance of all the excluding of battery, Accessories/Consumables
2. The terms and conditions of the tender and the agreement executed will be binding on the vendor/contractor. This offer is being issued in accordance with the terms & conditions of NEIGRIHMS /Government of India and in the manner specified herein shall operate to create a specific contract between the vendor/contractor (with whom the contract referred to) on one part and NEIGRIHMS, Shillong, on the other part.
3. The required spares to be replaced must be genuine and certified from the OEM.
4. **Repairs to be undertaken should be within specified configuration and maintaining the integration on internal circuit of equipment, any deviation on configuration/ specification the repair will not be acceptable. After repairs, a certificate to the effect that the equipment is in working order and safe for patient care and non-hazardous for the handler shall be submitted by the CMC holder.**
5. Tenderer/Vendors/contractor is responsible to provide electrical and patient safety certificate after major repair of equipment which are used for direct patient care.
6. The system must be checked & calibrated after every spare changes and detail service report must be submitted to the user & BME.
7. 2 nos of Periodic preventive maintenance is mandatory irrespective of unlimited service /breakdown calls.
8. Same /Similar Standby system must be provided by the bidder if the system needs to send to workshop for any major repair.

9. Receipt of this offer may be acknowledged and a copy duly signed/stamped by the authorized signatory should be submitted before finalization of the agreement.
10. The Performance security shall be denominated in any one of the forms namely Account Payee Demand Draft or Fixed Deposit Receipt drawn from any nationalized bank in India or Bank Guarantee issued by a nationalized bank in India, pledge in favor of Deputy Director, NEIGRIHMS, and Shillong-793018 for an amount equivalent to 3% of the total cost of annual CMC. The validity of the Fixed Deposit receipt or Bank Guarantee will be upto 2 months beyond CMC period.
- 11. It may also be noted that there should be no negligence in providing services of any type, if any, complaint is received the contract will be terminated with immediate effect.**
12. There will be 98% uptime warranty during CMC period on 24 (hrs) X 7 (days) X 365 (days) basis, with penalty, to extend CMC period by double the downtime period. The vendor shall ensure optimum uptime of the system during CMC period, failing which the initiator shall initiate stern action, as deemed fit.
13. During Comprehensive Maintenance Contract period, the supplier shall visit at each consignee's site for preventive maintenance including testing and calibration as per the manufacturer's service/ technical/ operational manual. The supplier shall visit each consignee site as recommended in the manufacturer's manual, but at least once in 6 months commencing from the date of the successful completion of warranty period for preventive maintenance of the goods.
14. Processing of bill may be considered on yearly basis with satisfactory report from the user department. The AMC/CMC bills should be certified by the concerned Head of the Department/ In- Charge, BME and the respective DMS/MS.
15. Software updates should be provided free of cost during CMC. The first service call by the team of service engineers should be within 7 days of issue of this order.
16. Settlement of disputes – Director, NEIGRIHMS or his authorized representative shall be the final authority in all disputes and decision will be binding on all concerned.
17. All other terms & conditions are as per award of contract mentioned in pre-page.
18. Bidders are required to sign the CMC contract agreement within 15 (fifteen) days from the issue of the letter of award/supply order, failing which EMD/security deposit may be forfeited or Contract declared null and void.
19. The bidder must abide by all statutory requirements of the region (Meghalaya) and trading license issued from the KHADC should be produced within 30 days after award of contract.
20. Pre bid:- E-Procurement Cell , Ground Floor, Director block, NEIGRIHMS Shillong-793018 Tel:0364-2538032/2539458 or through VC link
21. Technical Specification

The specifications in Additional Terms and Conditions – Buyer Specific Clauses (ATC) shall supersede the “GeM category specification”. The technical evaluation will be as per these specifications only, and the bidders shall quote the equipment and submit the technical compliance accordingly.

-

(F) Special Terms and Conditions-

1. All Provisions of Drugs and Cosmetics Act, 1940 and Rules (including Medical Device Rule 2017) made there under as amended till date will always be applicable. This will include all notifications issued by Central Drugs Standard Control Organization (CDSCO), Ministry of Health & Family Welfare (MoHFW) and Department of Pharmaceuticals (DOP), Ministry of Chemicals & Fertilizers time to time in this regard.
1. The sellers are registered on GeM based on self declaration of valid Medical Device License, product certification, test reports etc. However, buyers must check and validate the details at their end for all applicable licenses and certifications e.g., validity and authenticity/genuineness of Medical Device license, product certification, manufacturer certification/licenses, test reports etc.

2. In case of authorized resellers/distributors, it will be the legal & regulatory liability of the manufacturer to ensure that their resellers/distributors are operating in compliance with all relevant laws and regulations and are properly licensed to sell the manufacturer's products, including verifying the validity and authenticity of Medical Device license held by them.
3. The price offered by the seller/bidder shall not, in any case exceed the 8 / 23 DPCO/NPPA controlled price or price fixed by State Government, if any. The seller must reduce the prices if there is any reduction in DPCO/NPPA ceiling price or price fixed by State Government, if any.
4. Any other Terms and Conditions which is not included or at variance with the conditions specified in STC/GTC, may be added by the buyer through Additional Terms and Conditions (ATC) in the bid to ensure items are procured from authentic/validated source with appropriate and applicable quality. The above terms and conditions are in reverse order of precedence i.e. ATC shall supersede specific STC which shall supersede General Terms and Conditions (GTC), whenever there are any conflicting provisions.
5. Comprehensive warranty: Comprehensive warranty shall include preventive maintenance including calibration as per technical/ service /operational manual of the manufacturer, service charges and spares. During the warranty period commencing from date of the successful completion of warranty period, Service personnel shall visit each consignee site as recommended in the manufacturer's technical/ service /operational manual, at least once in six months. warranty shall not be including the consumables. Further there will be 98% uptime warranty during warranty period on 24 (hrs) X 7 (days) X 365 (days) basis, with penalty, to extend warranty period by double the downtime period.
6. Service centres: Details of Service outlets in India to render services for equipment to be furnished to buyer/consignees with complete address, telephone numbers, e mails etc at time of making the supplies. It shall be the responsibility of seller to ensure that authorized service centres are available to cater to the areas where supplies are made within reasonable distance from where the service calls can be handled. Details of toll-free numbers for service call and online registration of service requests also to be provided buyer/consignee at the time of supplies.
7. Source of supply: It shall be responsibility of seller to provide Documents regarding source of equipments such as copy of Performa invoice or any other documents to establish that the products supplied are manufactured by OEM indicated and sourced from them.
8. Packing and Marking: Medical equipments being very delicate and sensitive packing for the goods should be strong and durable enough to withstand transit including transshipment (if any), rough handling, open storage etc. without any damage, deterioration etc. .The size, weights and volumes of the packing cases, remoteness of the final destination of the goods, availability or otherwise of transport and handling facilities at all points during transit up to final destination,. Quality of packing, the manner of marking within & outside the packages and provision of accompanying documentation shall take into consideration the type of medical equipments being supplied. The accessories shall be suitably labelled and packed. Each of the package shall be marked on three sides with indelible paint of proper quality: indicating contract number and date, brief description of goods including quantity, Packing list reference number, country of origin of goods and any other relevant details.
9. Spare Parts: Seller shall provide materials, information etc. pertaining to spare parts manufactured and supplied by the OEM. It shall be ensured that the required spares are available for purchase at least for 10 years from date of supplies. In case due to any reasons the production of the spare parts is discontinued sufficient advance notice should be given to the 9 / 23 buyer/consignee before such discontinuation to provide adequate time to purchase the required spare parts etc. Further, OEM and their service centres/dealers shall carry sufficient inventories to assure ex-stock supply of consumables and spares for the equipments so that the same are available. OEM or reseller shall always accord most favoured client status to the buyer/consignee and shall give the most competitive price for spares and consumables of its machines/equipments supplied.

10. Installation, Training, Manuals: Seller shall be responsible to carry out Installation & commissioning, Supervision and Demonstration of the goods. They shall provide required jigs and tools for assembly, minor civil works for the completion of the installation and Training of Consignee's representatives for operating and maintaining the equipment and supplying required number of operation & maintenance manual for the goods. In case the category parameters are specifying any requirements regarding the installations, training and manuals the same shall also be applicable.
11. Electrical safety checking: Sellers are required to make sure that they furnish the list of equipments for carrying out routine and preventive maintenance to buyer/consignee. They should make sure to periodically check the electrical safety aspects as per BIS Safety Standards or equivalent. In case they do not have required equipment for such testing should ensure that the equipments checked for electrical safety compliance through labs with facilities for such checking during every preventive maintenance call.
12. Software: All software updates should be provided free of cost during warranty period.

(G) Integrity Pact (Pre -Agreement)-

Bidders has to sign and submit the Integrity Pact (Pre -Agreement), as per the format attached.

Bidders may raise disputes /complaints, if any, with the Independent External Monitor(s) through email followed by written document or telephone and written document.

-

Independent External Monitor (IEM)

Sl. No.	Name of Independent External Monitor (IEM)	Address and Contact details
1.	Shri. R. Govindarajan, IRS (Retd.)	Flat F4, Capella, 29/10, Arcot Road, Saligramam, Chennai -600093; Mobile No. 9444908194 /8985970019; Email id: rgvig@gmail.com
2.	Shri. Ganganna Satish, IFoS (Retd.)	108, Devikripa, Vinayaka Layout, 2 nd main, Bhopasandra, near Sunrise School, Bangalore -560094; Mobile No. 9845256115; Email id: satishifs@gmail.com

-

अस्वीकरण/Disclaimer

The additional terms and conditions have been incorporated by the Buyer after approval of the Competent

Authority in Buyer Organization, whereby Buyer organization is solely responsible for the impact of these clauses on the bidding process, its outcome, and consequences thereof including any eccentricity / restriction arising in the bidding process due to these ATCs and due to modification of technical specifications and / or terms and conditions governing the bid. If any clause(s) is / are incorporated by the Buyer regarding following, the bid and resultant contracts shall be treated as null and void and such bids may be cancelled by GeM at any stage of bidding process without any notice:-

1. Definition of Class I and Class II suppliers in the bid not in line with the extant Order / Office Memorandum issued by DPIIT in this regard.
2. Seeking EMD submission from bidder(s), including via Additional Terms & Conditions, in contravention to exemption provided to such sellers under GeM GTC.
3. Publishing Custom / BOQ bids for items for which regular GeM categories are available without any Category item bunched with it.
4. Creating BoQ bid for single item.
5. Mentioning specific Brand or Make or Model or Manufacturer or Dealer name.
6. Mandating submission of documents in physical form as a pre-requisite to qualify bidders.
7. Floating / creation of work contracts as Custom Bids in Services.
8. Seeking sample with bid or approval of samples during bid evaluation process. (However, in bids for [attached categories](#), trials are allowed as per approved procurement policy of the buyer nodal Ministries)
9. Mandating foreign / international certifications even in case of existence of Indian Standards without specifying equivalent Indian Certification / standards.
10. Seeking experience from specific organization / department / institute only or from foreign / export experience.
11. Creating bid for items from irrelevant categories.
12. Incorporating any clause against the MSME policy and Preference to Make in India Policy.
13. Reference of conditions published on any external site or reference to external documents/clauses.
14. Asking for any Tender fee / Bid Participation fee / Auction fee in case of Bids / Forward Auction, as the case may be.
15. Buyer added ATC Clauses which are in contravention of clauses defined by buyer in system generated bid template as indicated above in the Bid Details section, EMD Detail, ePBG Detail and MII and MSE Purchase Preference sections of the bid, unless otherwise allowed by GeM GTC.
16. In a category based bid, adding additional items, through buyer added additional scope of work/ additional terms and conditions/or any other document. If buyer needs more items along with the main item, the same must be added through bunching category based items or by bunching custom catalogs or bunching a BoQ with the main category based item, the same must not be done through ATC or Scope of Work.

Further, if any seller has any objection/grievance against these additional clauses or otherwise on any aspect of this bid, they can raise their representation against the same by using the Representation window provided in the bid details field in Seller dashboard after logging in as a seller within 4 days of bid publication on GeM. Buyer is duty bound to reply to all such representations and would not be allowed to open bids if he fails to reply to such representations.

Additional Clause For Comprehensive Maintenance Charges

- 1.CMC shall include preventive maintenance including calibration as per technical/ service /operational manual of the manufacturer, service charges and spares, after satisfactory completion of Warranty. During the CMC period commencing from date of the successful completion of warranty period, Service personnel shall visit each consignee site as recommended in the manufacturer's technical/ service /operational manual, at least once in six months or as per user requirement. Cost of consumables shall not be included in CMC.Further there will be 98% uptime warranty during CMC period on 24 (hrs) X 7 (days) X 365 (days) basis, with penalty, to extend CMC period by double the downtime period.
- 2.CMC charges to be indicated as percentage of cost of equipment quoted for each year after the warranty period.
- 3.GST shall be included in the CMC Charges quoted.
- 4.Cost of CMC will be added for Ranking/Evaluation purpose with depreciation formula.A 10% discounting rate per year shall be applied on CMC Charges for price evaluation on Net Present Value.
- 5.The payment of CMC will be made on quarterly basis after satisfactory completion of said period, duly certified by end user.
- 6.While creating a bid or RA, buyers shall indicate whether CMC is required against Yes/No" options. If CMC Charges

are included, an option for number of years for CMC required after the warranty period shall be available. Under this option up to 10 years can be chosen for CMC charges beyond warranty period.

7. In case the bid has a provision for CMC, the warranty of the product will also be deemed to have been converted into Comprehensive warranty including preventive maintenance and calibration as per technical/ service /operational manual of the manufacturer, service charges and spares, during the Warranty Period also. Sellers are therefore advised to include the cost of Comprehensive Warranty including spares (excluding consumables) also in product Cost.

8. The CMC functionality shall be available in bid only and no direct RA shall be applicable. In case of bid to R/A decrement rules shall be applicable on total price inclusive of CMC charges. Bunching of products shall not be available while creating bids with CMC charges.

8.1. Buyer shall indicate number of years of warranty by selecting different options available in the field depending on warranty parameter applicable in category parameters for the equipment. No. of years of warranty indicated here shall supersede the warranty period indicated elsewhere in bid or product specifications. The Seller while participating in Bid/RA will get fields to indicate CMC charges as percentage depending on number of years of CMC selected by Buyer. The following shall be applicable, if 5 year CMC selected:

CMC charges for 1st year after warranty period- Percentage to be indicated- A1

CMC charges for 2nd year after warranty period- Percentage to be indicated- A2

CMC charges for 3rd year after warranty period - Percentage to be indicated- A3

CMC charges for 4th year after warranty period - Percentage to be indicated- A4

CMC charges for 5th year after warranty period - Percentage to be indicated- A5

Similarly, A6 to A10 are to be indicated for 6th to 10th year of CMC if applicable.

8.2. The calculation of CMC Charges shall take into account the number of years of warranty and duration of CMC as specified while creating bid.

8.3. In the price evaluation, the system shall provide function to calculate the cost of each equipment by formula indicated below including CMC and then show the inter-se-ranking of the bidders. The following are the variables

(i) Number of years for which CMC required.

(ii) Number of years of product warranty

The formula for calculating total cost including CMC charges shall be as under:

Total Cost for evaluation=

$C + C \times \{ (A1/100)/(1.10^n) + (A2/100)/(1.10^{n+1}) + (A3/100)/(1.10^{n+2}) + (A4/100)/(1.10^{n+3}) + (A5/100)/(1.10^{n+4}) \}$ and so on

C - Cost for equipment quoted and n shall be number of years of product warranty specified.

If 2 year warranty specified, n shall be 2 and if 5 year warranty specified, n shall be 5. A1, A2, A3, A4 & A5 shall depend on how many years CMC selected. For 3 year CMC, only A1, A2 and A3 factors are to be taken into account and A4 and A5 will not be applicable.

8.4. CMC charges offered for each subsequent year should be same or higher than preceding year.

8.5. The CMC charges shall be offered within range of 3 to 25% of cost of equipment.

9. Since CMC charges are to be paid only later for each year during CMC period, applicable performance guarantee amount after placement of contract shall be based on the cost of equipment excluding the cost of CMC Charges.

10. Performance bank guarantee applicable for CMC is to be submitted at start of the CMC and shall be applicable between 2.5% to 10% as specified in bid on total CMC Charges. The PBG submitted after award of contract shall be released only after new PBG for the CMC period is submitted and accepted by buyer/consignee after due verification. Bank guarantee for CMC is to remain valid till completion of CMC period plus one year. The bank guarantee for CMC shall be submitted to buyer directly. In case, seller fails to submit the PBG or does not provide services for the CMC contract after expiry of warranty period then PBG of equipment shall be forfeited.

11. In case of splitting of order quantity, equipment cost and CMC charges offered by L1 bidder shall be matched by higher quoting eligible bidders on one-to-one basis. The equipment cost and CMC charges (year to year) shall be matched individually.

12. The CMC Contract shall be an offline contract to be handled by buyer. The payment of CMC will be made on quarterly basis after satisfactory completion of said period, duly certified by end user and scope of CMC will be as per para 1 above.

13. CMC Charges are inclusive of all the charges for Transportation, Lodging, Boarding, all insurances including third party insurance and all other incidental charges. The same shall include GST. The prices also include cost of spares and damaged parts. Purchaser does not have any liability, whatsoever, over and above the cost of CMC. It also includes for arranging hand tools & tackles, special tools etc. required to carry out the work.

All GeM Sellers/Service Providers shall ensure full compliance with all applicable labour laws, including the provisions, rules, schemes and guidelines under the four Labour Codes i.e. the Code on Wages, 2019; the Industrial Relations Code, 2020; the Occupational Safety, Health and Working Conditions Code, 2020; and the Code on Social Security, 2020 as and when notified and brought into

force by the Government of India.

For all provisions of the Labour Codes that are pending operationalisation through rules, schemes or notifications, the corresponding provisions of the pre-existing labour enactments (such as The Minimum Wages Act, 1948, The Payment of Wages Act, 1936, The Payment of Bonus Act, 1965, The Equal Remuneration Act, 1976, The Payment of Gratuity Act, 1972, etc. and relevant State Rules) shall continue to remain applicable.

The Seller/ Service Providers shall, therefore, be responsible for ensuring compliance under:

- **All notified and enforceable provisions of the new Labour Codes as mentioned hereinabove; and**
- **All operative provisions of the erstwhile Labour Laws until their complete substitution.**

All obligations relating to wages, social security, safety, working conditions, industrial relations etc. and any other statutory requirements shall be strictly met by the Seller/ Service Provider. Any non-compliance shall constitute a breach of the contract and shall entitle the Buyer to take appropriate action in accordance with the contract and applicable law.

[यह बिड सामान्य शर्तों के अंतर्गत भी शासित है /This Bid is also governed by the General Terms and Conditions](#)

जेम की सामान्य शर्तों के खंड 26 के संदर्भ में भारत के साथ भूमि सीमा साझा करने वाले देश के बिडर से खरीद पर प्रतिबंध के संबंध में भारत के साथ भूमि सीमा साझा करने वाले देश का कोई भी बिडर इस निविदा में बिड देने के लिए तभी पात्र होगा जब वह बिड देने वाला सक्षम प्राधिकारी के पास पंजीकृत हो। बिड में भाग लेते समय बिडर को इसका अनुपालन करना होगा और कोई भी गलत घोषणा किए जाने व इसका अनुपालन न करने पर अनुबंध को तत्काल समाप्त करने और कानून के अनुसार आगे की कानूनी कार्रवाई का आधार होगा।/In terms of GeM GTC clause 26 regarding Restrictions on procurement from a bidder of a country which shares a land border with India, any bidder from a country which shares a land border with India will be eligible to bid in this tender only if the bidder is registered with the Competent Authority. While participating in bid, Bidder has to undertake compliance of this and any false declaration and non-compliance of this would be a ground for immediate termination of the contract and further legal action in accordance with the laws.

---धन्यवाद/Thank You---