

**1. BID No.: GEM/2026/B/7520287: Uroflowmetry System**

<b>Sl. No.</b>	<b>Tendered specification</b>	<b>May be Read As</b>	<b>Justification</b>
9c	Accuracy: +/- 10% of reading, +/- 20ml	Accuracy: +/- 10% of reading, +/- 20ml with volume rendering resolution of 1ml.	Resolution of 1ml ensures better precision.
9e (to be added)		Probe frequency of 2.5 Hz	Probe frequency not mentioned in the specification.
9f (to be added)		LCD screen must be >3.5 inches.	LCD screen size not mentioned in the specification.
9g (to be added)		Working time with Lithium rechargeable battery, minimum 4 hours.	Duration of working of lithium battery not mentioned in the specification.
9h (to be added)		Must be supplied with sage aluminium storage case.	Protects the equipment during storage and transportation.