

Bid/RA



(https://gem.gov.in/)

PRODUCTS SERVICES CONTENT

Note:

1. Response to all representation is mandatory.
2. Response to all Representations should be published at least 72 hours before bid end date and time. Buyer will not be allowed to open bid without responding to all representation.
3. Buyer can create responses and publish the same only after closure of representation window.

Bid Listing

Contains 4045

Bid Number: GEM/2026/B/7394045 **Published**
 Only up to 5 mins to react to any modification in published bid in search results.



P

B

By B

A

P

P

D

B

C

R

G

By B

F

Bid I

Bid I
Sho

1 - 1

Bid/RA Section	Seller Query/Representation	Buyer Response	Submitted On
Product/Service specification	The requirement of a mobile U-shaped wraparound system with instant conversion to a 90-degree L-shape appears to be a design feature specific to a limited number of manufacturers, rather than a universally adopted standard.	The equipment must guarantee the best coverage radiation protection for the whole body of operators and staffs	28-03-2026 11:18 PM
	The combination of electrically controlled height adjustment at all positions along with docking-type mobility and unrestricted operator movement narrows down the eligibility to very few proprietary systems.	The equipment must guarantee the best coverage radiation protection for the whole body of operators and staffs	28-03-2026 11:18 PM
	The clause stating 99% protection from feet to waist in LAO view without defining standardized test methodology (kVp, geometry, scatter conditions) may lead to ambiguity and selective qualification.	The system should provide attenuation at least 95-99% at 90KVp when scattering radiation occurs in LAO and RAO geometry angulations in the laboratory and live clinical all type of	28-03-2026 11:18 PM

Reset



records

of 1 Sort by: Bid End Date: Latest First

Bid/RA

records

BID NO: GEM/2026/B/7394045 (/showbidDocument/9165916)

Items: Note: Movable wraparound modular and...

Specification(S) (<https://bidplus.gem.gov.in/bidding/product-specifications/9165916>)

Remaining Days: View
1. Response to all Representations is mandatory.
2. Response to all Representations should be published at least 72 hours before bid end date and time. Buyer will not be allowed to open bid without responding to all representation.

Notification(S) (https://bidplus.gem.gov.in/bidding/show_bid_notifications/9165916)

Bid Document (/showbidDocument/9165916)

View Representations

Invite Bidder

Start Date: 26-03-2026 1:50 PM

End Date: 15-05-2026 2:00 PM

Bid Status: Active

Bid Doc Hash: View

Bid Number: GEM/2026/B/7394045

Published

Bid/RA Section	Seller Query/Representation	Buyer Response	Submitted On
Product/Service specification	The requirement of a mobile U-shaped wraparound system with instant conversion to a 90-degree L-shape appears to be a design feature specific to a limited number of manufacturers, rather than a universally adopted standard.	The equipment must guarantee the best coverage radiation protection for the whole body of operators and staffs	28-03-2026 11:18 PM
	The combination of electrically controlled height adjustment at all positions along with docking-type mobility and unrestricted operator movement narrows down the eligibility to very few proprietary systems.	The equipment must guarantee the best coverage radiation protection for the whole body of operators and staffs	28-03-2026 11:18 PM
	The clause stating 99% protection from feet to waist in LAO view without defining standardized test methodology (kVp, geometry, scatter conditions) may lead to ambiguity and selective qualification.	The system should provide attenuation at least 95-99% at 90KVp when scattering radiation occurs in LAO and RAO geometry angulations in the laboratory and live clinical all type of	28-03-2026 11:18 PM

Pre

Bid/RA ([HTTPS://GEM.GOV.IN/TESTIMONIAL BUYERS](https://gem.gov.in/testimonial_buyers))

Miscellaneous
(<https://gem.gov.in/support/miscellaneous>)

MouUs (<https://gem.gov.in/mou>)

Note:

1. Response to all representation is mandatory.
2. Response to all Representations should be published at least 72 hours before bid end date and time. Buyer will not be allowed to open bid without responding to all representation.

LMS (<https://lms.gem.gov.in/>)

FAQs (<https://gem.gov.in/userFaqs>)

Training Calendar (<https://gem.gov.in/training>)

Vendor Guidelines (<https://gem.gov.in/training/vendor/buyers>)

Training Module
(https://gem.gov.in/training/training_module)

Raise-a Ticket (<https://gem.gov.in/gemtickets/create>)

Contact Us (<https://gem.gov.in/contactUs>)

Facilitators (<https://gem.gov.in/training/facilitators>)

Careers (<https://gem.gov.in/landing/index/careers>)

Bid Number: GEM/2026/B/7394045

Published

Download Logo
(<https://gem.gov.in/resources/images/download-gem-logs.png>)

Bid/RA Section	Seller Query/Representation	Buyer Response	Submitted On
Product/Service specification	The requirement of a mobile U-shaped wraparound system with instant conversion to a 90-degree L-shape appears to be a design feature specific to a limited number of manufacturers, rather than a universally adopted standard.	The equipment must guarantee the best coverage radiation protection for the whole body of operators and staffs	28-03-2026 11:18 PM
	The combination of electrically controlled height adjustment at all positions along with docking-type mobility and unrestricted operator movement narrows down the eligibility to very few proprietary systems.	The equipment must guarantee the best coverage radiation protection for the whole body of operators and staffs	28-03-2026 11:18 PM
	The clause stating 99% protection from feet to waist in LAO view without defining standardized test methodology (kVp, geometry, scatter conditions) may lead to ambiguity and selective qualification.	The system should provide attenuation at least 95-99% at 90KVp when scattering radiation occurs in LAO and RAO geometry angulations in the laboratory and live clinical all type of	28-03-2026 11:18 PM